

This is a preview of "ISO 11608-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Third edition
2022-04

Needle-based injection systems for medical use — Requirements and test methods —

Part 2: Double-ended pen needles

*Systèmes d'injection à aiguille pour usage médical — Exigences et
méthodes d'essai —*

Partie 2: Aiguilles à deux extrémités pour stylos-injecteurs



Reference number
ISO 11608-2:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 11608-2:2022". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	v
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols	3
5 Requirements	3
5.1 Needle tube requirements.....	3
5.1.1 General.....	3
5.1.2 Needle tubing materials.....	4
5.1.3 Tubing characteristics.....	4
5.2 Double-ended pen needle requirements.....	4
5.2.1 General.....	4
5.2.2 Biocompatibility.....	4
5.2.3 Dimensions for double-ended pen needle assembly.....	4
5.2.4 Needle points.....	5
5.2.5 Freedom from defects.....	5
5.2.6 Flow rate through the needle.....	5
5.2.7 Bond between hub and needle tube.....	6
5.2.8 Dislocation of measuring point at patient end.....	7
5.2.9 Ease of assembly.....	7
5.2.10 Sterility.....	7
5.2.11 Pyrogenicity.....	8
5.3 Functional compatibility with NISs.....	8
5.3.1 General.....	8
5.3.2 Dose delivery.....	8
5.3.3 Needle removal torque.....	9
6 Sampling	10
7 Preconditioning of needles	13
7.1 Preconditioning in a dry-heat, cold storage and damp heat atmosphere.....	13
7.2 Preconditioning in a cyclical atmosphere.....	13
8 Standard atmosphere and test apparatus	13
8.1 Standard test atmosphere.....	13
8.2 Test gauge.....	13
8.3 Test apparatus.....	14
9 Test methods	14
9.1 Bond between hub and needle tube.....	14
9.2 Determination of dislocation of measuring point at patient end.....	14
9.3 Ease of assembly.....	15
9.4 Functional compatibility with NISs.....	15
9.4.1 Sample quantity requirements.....	15
9.4.2 Test procedures for testing dose delivery.....	16
9.4.3 Procedure for testing needle hub removal torque.....	17
10 Packaging	17
11 Information supplied with the needle(s)	17
11.1 General.....	17
11.2 Marking.....	18
11.2.1 Marking on the unit packaging.....	18
11.2.2 Marking on the user packaging.....	19
11.3 Instructions for use.....	20

This is a preview of "ISO 11608-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Annex A (normative) Determination of flow rate through needle	21
Annex B (informative) Needle bonding strength test method	24
Annex C (informative) Additional background for requirements	25
Bibliography	30

This is a preview of "ISO 11608-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11608-2:2012), which has been technically revised.

The main changes are as follows:

- terminology for 'needles' was updated throughout the document and in the title of the document to 'double-ended pen needles' in order to more precisely describe the subject of the document;
- where possible, references to other parts of ISO 11608 (all parts) have been made instead of repeating the content in this document (for example, conditions for preconditioning). Additionally, changes have been made to align with ISO 11608-1 (e.g. sample sizes and test case matrix);
- content within this document has been reorganized to create a separate clause for symbols and abbreviated terms, to delineate needle tube requirements, double-ended pen needle requirements and requirements for functional compatibility with needle-based injection systems (NISs), to group the test method sections into defined sections;
- new figures have been added to more precisely illustrate the dimensions of the double-ended pen needle and test gauge referred to in this document;
- requirements for biocompatibility and pyrogenicity introduced and acceptance criteria for flow rate through the needle have been added;
- testing requirements necessary to establish functional compatibility between a specific NIS and a specific double-ended pen needle were revised to include dose delivery and needle hub removal force testing;

This is a preview of "ISO 11608-2:2022". [Click here to purchase the full version from the ANSI store.](#)

- new annexes have been introduced which provide an example method for testing needle bond force ([Annex B](#)) and additional background for the requirements in this document ([Annex C](#)).

A list of all parts in the ISO 11608 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO 11608-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Introduction

This document covers sterile double-ended needles intended for single use in conjunction with needle-based injection systems (e.g. pen injectors). These needles are often referred to as pen needles.

The devices described in this document are designed to be used with the devices described in ISO 11608-1 and ISO 11608-3. This document is intended to be used in conjunction with ISO 11608-1.

The first edition of this document, i.e. ISO 11608-2:2000, introduced the concept of interchangeability and the labelling designations "Type A" (i.e., interchangeable) and "non-Type A" for needles and container closure systems. Since its promulgation, experience has shown that the complexity of these systems makes it very difficult to ensure functional compatibility as defined in the different clauses of this document, particularly when products are made by different manufacturers and the design is not verified as a system. Based on this experience, it is believed that the Type A designation does not represent adequate guidance to the user in making decisions on the compatibility of needles and container closures with specific needle-based injection systems (NIS). As such, the labelling designation "Type A" has been removed.

The second edition of this document, i.e. ISO 11608-2:2012, addressed functional compatibility of the system through testing in accordance with [Clause 11](#) (functional compatibility testing is in [Clause 9](#) in this document). Flow rate was introduced as a new parameter. The sampling plans for inspection selected for this document and outlined in ISO 11608-1 are intended to verify the design, at a high confidence level. The sampling plan does not replace the more general manufacturing quality systems, including lot release, which appear in quality management systems, e.g. ISO 9001 or ISO 13485.

This document refers to ISO 11608-1 for dose accuracy requirements for functional compatibility and includes requirements for double-ended pen needle device adapted from ISO 7864:2016 and new requirements for cleanliness (as part of the freedom from defects requirement), biocompatibility and pyrogenicity testing. After the experience gained through testing for functional compatibility with the second edition of this document, the testing approach was reassessed to ensure that the appropriate tests were included. Following a rigorous review of anonymized dose accuracy data generated by manufacturers over several years, assessing the relationship of dose accuracy and the flow rate through the pen needle, it was determined that there is no relationship between the pen needle flow rate and dose accuracy. This further supports the understanding that the primary contributors to dose accuracy in the NIS system are the NIS and the cartridge.

The needle provides the fluid path from the cartridge to the subcutaneous tissue and its influence on dose accuracy is mainly affected by the correct position of the needle bevel inside the cartridge, a possible leakage and indirectly over the injection force and injection time by the flow rate. The inner diameter has an influence on the time that is needed to deliver the entire volume but not directly to the precision of the dosage. Additionally, a prescribed holding time may be specified in the NIS IFU, which provides for completion of dose delivery during expansion and or relaxation of soft parts within the NIS and cartridge. Therefore, the testing requirements necessary to establish functional compatibility between a specific NIS and a specific pen needle were revised in this edition to include dose delivery and needle hub removal force testing. Dose delivery may be demonstrated either through dose accuracy testing or through confirmation of dose delivery, which includes demonstrating that the needle bevel is positioned inside the cartridge, visual verification that the dose is expelled and inspection for absence of leakage at the base of the non-patient end of the cannula.