



ISO 11929-1

Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application —

Part 1: Elementary applications

Détermination des limites caractéristiques (seuil de décision, limite de détection et limites de l'intervalle élargi) pour mesurages de rayonnements ionisants — Principes fondamentaux et applications —

Partie 1: Applications élémentaires

Third edition
2025-12

This is a preview of ISO 11929-1:2025. Click [here](#) to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of ISO 11929-1:2025. [Click here to purchase the full version from the ANSI store.](#)

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Quantities and symbols	6
5 Summary of procedures for evaluating a measurement and calculating the characteristic limits	8
5.1 General aspects.....	8
5.2 Modelling the measurement.....	9
5.3 General considerations about the applicability of the ISO/IEC Guide 98-3.....	9
5.4 Evaluating the input quantities, standard uncertainties and covariances, and the primary result and its associated standard uncertainty.....	10
5.5 Evaluating the standard uncertainty as a function of an assumed true value of the measurand.....	10
5.6 Calculating the decision threshold and decisions to be made.....	11
5.7 Calculating the detection limit and assessment of the measurement procedure.....	12
5.8 Calculating a coverage interval for the measurand.....	12
5.9 Calculating the best estimate of the measurand and its associated standard uncertainty.....	13
5.10 Reporting the results.....	14
6 Fundamentals and evaluation of the measurement	16
6.1 General aspects concerning the measurand.....	16
6.2 Model of evaluation.....	16
6.2.1 General model.....	16
6.2.2 Model in ionizing radiation counting measurements.....	17
6.3 Evaluation of the primary measurement result.....	18
6.4 Standard uncertainty associated with the primary measurement result.....	18
7 Standard uncertainty as a function of an assumed value of the measurand	19
8 Decision threshold, detection limit and assessments	21
8.1 Specifications.....	21
8.2 Decision threshold.....	21
8.3 Detection limit.....	21
8.4 Assessments.....	22
9 Limits of the coverage interval	23
9.1 General aspects.....	23
9.2 The probabilistically symmetric coverage interval.....	23
9.3 The shortest coverage interval.....	24
10 The best estimate and its associated standard uncertainty	24
11 Documentation	25
Annex A (normative) Repeated counting measurements with random influences	27
Annex B (normative) Measurements with ratemeters	32
Annex C (informative) Calculation of the detection limit by iteration	34
Annex D (informative) Calculations with a generalized net count rate	36
Annex E (informative) Distribution function of the standardized normal distribution	39
Bibliography	44

This is a preview of ISO 11929-1:2025. [Click here to purchase the full version from the ANSI store.](#)

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, *Nuclear energy, nuclear technologies, and radiological protection*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11929-1:2019), of which it constitutes a minor revision.

The main changes are as follows:

- correction of the internal references within the text;
- correction of [Formulae \(3\)](#), [\(A.5\)](#) and [\(B.3\)](#);
- correction of the definitions of decision threshold [\(3.12\)](#) and the detection limit [\(3.13\)](#);
- introduction of NOTE 3 in [5.8](#), for clarity;
- complement of the NOTE in [9.1](#);
- correction of the 1st paragraph and NOTE 2 in [Clause 10](#);
- correction of the 8th paragraph in [Annex B](#).

A list of all the parts in the ISO 11929 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of ISO 11929-1:2025. [Click here to purchase the full version from the ANSI store.](#)

Measurement uncertainties and characteristic values, such as the decision threshold, the detection limit and limits of the coverage interval for measurements, as well as the best estimate and its associated standard measurement uncertainty, are of importance in metrology in general and for radiological protection in particular. The quantification of the uncertainty associated with a measurement result provides a basis for the trust an individual can have in a measurement result. Conformity with regulatory limits, constraints or reference values can only be demonstrated by taking into account and quantifying all sources of uncertainty. Characteristic limits provide, at the end, the basis for deciding if uncertainties have to be taken into account.

This standard provides characteristic values of a non-negative measurand of ionizing radiation. It is also applicable for a wide range of measuring methods extending beyond measurements of ionizing radiation.

The limits to be provided according to ISO 11929 series for specified probabilities of wrong decisions allow detection possibilities to be assessed for a measurand and for the physical effect quantified by this measurand as follows:

- the “decision threshold” allows a decision to be made on whether or not the physical effect quantified by the measurand is present;
- the “detection limit” indicates the smallest true quantity value of the measurand that can still be detected with the applied measurement procedure; this gives a decision on whether or not the measurement procedure satisfies the requirements and is therefore suitable for the intended measurement purpose;
- the “limits of the coverage interval” enclose, in the case of the physical effect recognized as present, a coverage interval containing the true quantity value of the measurand with a specified probability.

Hereinafter, the limits mentioned are jointly called the “characteristic limits”.

NOTE According to ISO/IEC Guide 99:2007 updated by JCGM 200:2012 the term “coverage interval” is used here instead of “confidence interval” in order to distinguish the wording of Bayesian terminology from that of conventional statistics.

All the characteristic values are based on Bayesian statistics and on the ISO/IEC 98-3 as well as on the ISO/IEC Guide 98-3:2008/Suppl 1:2008 and ISO/IEC Guide 98-3:2008/Suppl 2:2011. As explained in detail in ISO 11929-2, the characteristic values are mathematically defined by means of moments and quantiles of probability distributions of the possible measurand values.

Since measurement uncertainty plays an important part in ISO 11929, the evaluation of measurements and the treatment of measurement uncertainties are carried out by means of the general procedures according to the ISO/IEC Guide 98-3 and to the ISO/IEC Guide 98-3:2008/Suppl 1:2008; see also References [13] to [17]. This enables the strict separation of the evaluation of the measurements, on the one hand, and the provision and calculation of the characteristic values, on the other hand. The ISO 11929 series makes use of a theory of uncertainty in measurement^{[18] to [20]} based on Bayesian statistics (e.g. see References [21] to [26]) in order to take into account those uncertainties that cannot be derived from repeated or counting measurements. The latter uncertainties cannot be handled by frequentist statistics.

Because of developments in metrology concerning measurement uncertainty laid down in the ISO/IEC Guide 98-3, ISO 11929:2010 was drawn up on the basis of the ISO/IEC Guide 98-3, but using Bayesian statistics and the Bayesian theory of measurement uncertainty. This theory provides a Bayesian foundation for the ISO/IEC Guide 98-3. Moreover, ISO 11929:2010 was based on the definitions of the characteristic values,^[13] the standard proposal,^[14] and the explanatory article.^[15] It unified and replaced all earlier parts of ISO 11929 and was applicable not only to a large variety of particular measurements of ionizing radiation but also, in analogy, to other measurement procedures.

In 2008 the ISO/IEC Guide 98-3:2008/Suppl 1:2008 has been published, dealing comprehensively with a more general treatment of measurement uncertainty using the Monte Carlo method in complex measurement evaluations. This provided an incentive for writing a corresponding Monte Carlo supplement^[16] to ISO 11929:2010 and to revise ISO 11929:2010. The revised ISO 11929 is also essentially founded on Bayesian statistics and can serve as a bridge between ISO 11929:2010 and the ISO/IEC Guide 98-3:2008/Suppl 1:2008. Moreover, more general definitions of the characteristic values (ISO 11929-2) and the Monte

This is a preview of ISO 11929-1:2025. [Click here to purchase the full version from the ANSI store.](#)

be propagated. It is thus more comprehensive and extending the range of applications.

The revised ISO 11929, moreover, is more explicit on the calculation of the characteristic values. It corrects also a problem in ISO 11929:2010 regarding uncertain quantities and influences, which do not behave randomly in measurements repeated several times. Reference [17] gives a survey on the basis of the revision. Furthermore, in ISO 11929-3, it gives detailed advice how to calculate characteristic values in the case of multivariate measurements using unfolding methods. For such measurements, the ISO/IEC Guide 98-3:2008/Suppl 2:2011 provides the basis of the uncertainty evaluation.

Formulae are provided for the calculation of the characteristic values of an ionizing radiation measurand via the “standard measurement uncertainty” of the measurand (hereinafter the “standard uncertainty”) derived according to the ISO/IEC Guide 98-3 as well as via probability density functions (PDFs) of the measurand derived in accordance with the ISO/IEC Guide 98-3:2008/Suppl 1:2008. The standard uncertainties or probability density functions take into account the uncertainties of the actual measurement as well as those of sample treatment, calibration of the measuring system and other influences. The latter uncertainties are assumed to be known from previous investigations.