First edition 2001-03-01

## Determination of particle size distribution by centrifugal liquid sedimentation methods —

# Part 1: General principles and guidelines

Détermination de la distribution granulométrique par les méthodes de sédimentation centrifuge dans un liquide —

Partie 1: Principes généraux et lignes directrices



Reference number ISO 13318-1:2001(E)

#### **PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.ch Web www.iso.ch

Printed in Switzerland

## Contents

Forewordiv	
Introduction	/
1 Scope	1
2 Normative references	1
3 Terms definitions and symbols	2
4 Principles	4
5 Particle size, shape and porosity limitations	9
6 Test conditions	D
7 Sampling1	1
8 Preparation for a sedimentation analysis1	1
9 Tests in duplicate and validation12	2
10 Reporting of results	2
Annex A (informative) The effect of measurement zone depth14	4
Annex B (informative) Accuracy of Stokes law as a function of Reynolds number	5
Bibliography10	6

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 13318 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 13318-1 was prepared by Technical Committee ISO/TC 24, Sieves, sieving and other sizing methods, Subcommittee SC 4, Sizing by methods other than sieving.

ISO 13318 consists of the following parts, under the general title *Determination of particle size distribution by centrifugal liquid sedimentation methods*:

- Part 1: General principles and guidelines
- Part 2: Centrifugal photosedimentation method
- Part 3: Centrifugal X-ray method

Annexes A and B of this part of ISO 13318 are for information only.

### Introduction

Centrifugal sedimentation particle size analysis methods are among those in current use for determining size distribution of many powders. Typically, centrifugal methods apply to samples in the 0,1  $\mu$ m to 5  $\mu$ m size range and where the sedimentation condition for a Reynolds number < 0,25 is satisfied.

No single method of size analysis can be specified to cover the many different types of material encountered, but it is possible to recommend procedures that may be applied to the majority of cases. The purpose of this part of ISO 13318 is to obtain uniformity in procedure of centrifugal methods in order to facilitate comparisons of size analysis made in different laboratories.

Centrifugal sedimentation methods may be undertaken:

- as part of a research project involving an investigation of the particle size distribution of a material;
- as part of a control procedure for the production of a material where the particle size distribution is important;
- as the basis of a contract for the supply of material specified to be within stated specification limits.

Gravitational sedimentation methods are discussed in ISO 13317-1, ISO 13317-2 and ISO 13317-3.