

This is a preview of "ISO 13794:1999". [Click here to purchase the full version from the ANSI store.](#)

First edition  
1999-07-15

---

---

## **Ambient air — Determination of asbestos fibres — Indirect-transfer transmission electron microscopy method**

*Air ambiant — Dosage des fibres d'amiante — Méthode par microscopie électronique à transmission par transfert indirect*



Reference number  
ISO 13794:1999(E)

## Contents

	Page
<b>1 Scope</b> .....	<b>1</b>
<b>1.1 Substance determined</b> .....	<b>1</b>
<b>1.2 Type of sample</b> .....	<b>1</b>
<b>1.3 Range</b> .....	<b>1</b>
<b>1.4 Limit of detection</b> .....	<b>1</b>
<b>2 Terms and definitions</b> .....	<b>2</b>
<b>3 Abbreviated terms</b> .....	<b>6</b>
<b>4 Principle</b> .....	<b>6</b>
<b>5 Apparatus</b> .....	<b>7</b>
<b>5.1 Air sampling</b> .....	<b>7</b>
<b>5.2 Specimen preparation laboratory</b> .....	<b>8</b>
<b>5.3 Equipment for analysis</b> .....	<b>8</b>
<b>5.4 Consumable supplies</b> .....	<b>13</b>
<b>6 Reagents</b> .....	<b>14</b>
<b>7 Air sample collection</b> .....	<b>14</b>
<b>7.1 Calculation of analytical sensitivity</b> .....	<b>14</b>
<b>7.2 Sample collection procedure</b> .....	<b>15</b>
<b>8 Procedure for analysis</b> .....	<b>15</b>
<b>8.1 General</b> .....	<b>15</b>
<b>8.2 Cleaning of sample cassettes</b> .....	<b>16</b>
<b>8.3 Preparation of analytical filters</b> .....	<b>16</b>
<b>8.4 Preparation of TEM specimens from PC analytical filters</b> .....	<b>17</b>

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet iso@iso.ch

Printed in Switzerland

This is a preview of "ISO 13794:1999". [Click here to purchase the full version from the ANSI store.](#)

<b>8.5 Preparation of TEM specimens from cellulose ester analytical filters</b> .....	<b>19</b>
<b>8.6 Criteria for acceptable TEM specimen grids</b> .....	<b>20</b>
<b>8.7 Procedure for structure counting by TEM</b> .....	<b>21</b>
<b>8.8 Blank and quality control determinations</b> .....	<b>24</b>
<b>8.9 Calculation of results</b> .....	<b>25</b>
<b>9 Performance characteristics</b> .....	<b>25</b>
<b>9.1 General</b> .....	<b>25</b>
<b>9.2 Interferences and limitations of fibre identification</b> .....	<b>25</b>
<b>9.3 Precision and accuracy</b> .....	<b>26</b>
<b>9.4 Limit of detection</b> .....	<b>26</b>
<b>10 Test report</b> .....	<b>27</b>
<b>Annex A (normative) Determination of operating conditions for plasma asher</b> .....	<b>30</b>
<b>Annex B (normative) Determination and standardization of operating conditions for ultrasonic bath</b> .....	<b>31</b>
<b>Annex C (normative) Calibration procedures</b> .....	<b>33</b>
<b>Annex D (normative) Structure-counting criteria</b> .....	<b>36</b>
<b>Annex E (normative) Fibre identification procedure</b> .....	<b>44</b>
<b>Annex F (normative) Determination of concentrations of asbestos fibres and bundles longer than 5 µm, and of PCM-equivalent asbestos fibres</b> .....	<b>53</b>
<b>Annex G (normative) Calculation of results</b> .....	<b>54</b>
<b>Annex H (normative) Test procedure to determine suitability of cellulose ester sample collection filters</b> .....	<b>60</b>
<b>Annex I (informative) Strategies for collection of air samples</b> .....	<b>61</b>
<b>Bibliography</b> .....	<b>62</b>

This is a preview of "ISO 13794:1999". [Click here to purchase the full version from the ANSI store.](#)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 13794 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 3, *Ambient atmospheres*.

Annexes A to H form a normative part of this International Standard. Annex I is for information only.

This is a preview of "ISO 13794:1999". [Click here to purchase the full version from the ANSI store.](#)

## Introduction

This International Standard is applicable to the measurement of airborne asbestos in a wide range of ambient air situations, including the interior atmospheres of buildings, and for detailed evaluation of any atmosphere in which asbestos fibres are likely to be present. Because the best available medical evidence indicates that the numerical fibre concentration and the fibre size and type are the relevant parameters for evaluation of the inhalation hazards, a fibre counting and measuring technique is the only logical approach. Most fibres in ambient atmospheres are not asbestos, and therefore there is a requirement for fibres to be identified. Many airborne asbestos fibres in ambient atmospheres have diameters below the resolution limit of the optical microscope. This International Standard is based on transmission electron microscopy, which has adequate resolution to allow detection of small fibres and is currently the only technique capable of unequivocal identification of the majority of individual fibres of asbestos. The fibres found suspended in an ambient atmosphere can often be identified unequivocally, if sufficient measurement effort is expended. However, if each fibre were to be identified in this way, the analysis becomes prohibitively expensive. Because of instrumental deficiencies or because of the nature of the particulate, some fibres cannot be positively identified as asbestos, even though the measurements all indicate that they could be asbestos. Subjective and instrumental factors therefore contribute to this measurement, and consequently a very precise definition of the procedure for identification and enumeration of asbestos fibres is required.

In addition to single fibres and bundles, asbestos is often found in air samples as very complex, aggregated structures which may or may not be also aggregated with other particles. The number of asbestos fibres and bundles incorporated in these complex structures often exceeds the number of isolated fibres and bundles observed, and many of them may be completely obscured in direct-transfer TEM preparations. The method defined in this International Standard incorporates specimen preparation procedures that result in selective concentration of asbestos fibres, and removal of organic and water-soluble materials. These procedures have the effect of dispersing the majority of the complex clusters and aggregates of fibres into their component fibres and bundles so that the asbestos in the air sample can be more accurately quantified. All of the feasible specimen preparation techniques result in some modification of the airborne particulate. Even the collection of particles from a three-dimensional airborne dispersion on to a two-dimensional filter surface can be considered a modification of the particulate, and some of the particles in most samples are modified by the specimen preparation procedures. Although this method results in dispersal of complex clusters and aggregates, it minimizes other effects on the size distribution of fibres and fibre bundles.

This International Standard is necessarily complex, because the instrumental techniques used are complex, and also because a very detailed and logical procedure must be specified to reduce the subjective aspects of the measurement. The method of data recording specified in this International Standard is designed to allow re-evaluation of the fibre counting data as new medical evidence becomes available.

This International Standard describes the method of analysis for a single air filter. However, one of the largest potential errors in characterizing asbestos in ambient atmospheres is associated with the variability between filter samples. For this reason, it is necessary to design a replicate sampling scheme in order to determine the standard's accuracy and precision.

Comparison of results obtained using this indirect-transfer procedure with those from the direct-transfer procedure may not be done *a priori*. A site-specific intercomparison study must be done which takes into account the fibre size and type of asbestos, and also the nature of the source of the airborne asbestos.