

This is a preview of "ISO 15901-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2022-01

Pore size distribution and porosity of solid materials by mercury porosimetry and gas adsorption —

Part 2: Analysis of nanopores by gas adsorption

*Distribution des dimensions des pores et porosité des matériaux
solides par porosimétrie au mercure et par adsorption de gaz —*

Partie 2: Analyse des nanopores par adsorption de gaz



Reference number
ISO 15901-2:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 15901-2:2022". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Symbols	3
5 Principles	4
5.1 General.....	4
5.2 Methods of measurement.....	8
5.3 Choice of adsorptive.....	9
6 Measurement procedure	10
6.1 Sampling.....	10
6.2 Sample pretreatment.....	10
6.3 Measurement.....	10
7 Verification of apparatus performance	10
8 Calibration	11
9 Pore size analysis	11
9.1 General.....	11
9.2 Classical, macroscopic, thermodynamic methods for pore size analysis.....	12
9.2.1 Assessment of microporosity.....	12
9.2.2 Assessment of meso/macroporosity.....	19
9.3 Advanced, microscopic approaches based on density functional theory and molecular simulation.....	20
9.3.1 General.....	20
9.3.2 Application for pore size analysis: Kernel and integral adsorption equation.....	20
10 Reporting	21
Annex A (informative) Horvath-Kawazoe and Saito-Foley method	22
Annex B (informative) NLDFT method	25
Bibliography	28

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 24, *Particle characterization including sieving*, Subcommittee SC 4, *Particle characterization*.

This second edition cancels and replaces ISO 15901-2:2006 and ISO 15901-3:2007, which have been technically revised. It also incorporates the Technical Corrigendum ISO 15901-2:2006/Cor.1:2007.

The main changes compared to the previous edition are as follows:

- the analysis of nanopores by gas adsorption which combines the characterization of both micro- and mesopores is now addressed;
- the classification of adsorption isotherms and hysteresis loops has been updated.

A list of all parts in the ISO 15901 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO 15901-2:2022". [Click here to purchase the full version from the ANSI store.](#)

Introduction

In general, different types of pores may be pictured as apertures, channels, or cavities within a solid body or as the space (i.e. interstices or voids) between solid particles in a bed, compact or aggregate. Porosity is a term which is often used to indicate the porous nature of solid material and is more precisely defined as the ratio of the volume of accessible pores and voids to the total volume occupied by a given amount of the solid. According to the 2015 IUPAC recommendations^[1], nanopores are defined as pores with internal widths of equal or less than 100 nm and are divided into several subgroups dependent on their pore width:

- pores with width greater than about 50 nm are called macropores;
- pores of widths between 2 nm and 50 nm are called mesopores;
- pores with width of about 2 nm and less are called micropores;

Further, IUPAC suggested a subclassification of micropores into supermicropores (pore width 0,7 nm to 2 nm), and ultramicropores (pore width < 0,7 nm). In addition to the accessible pores, a solid may contain closed pores which are isolated from the external surface and into which fluids are not able to penetrate. The characterization of closed pores, i.e. cavities with no access to an external surface, is not covered in this document.

Porous materials may take the form of fine or coarse powders, compacts, extrudates, sheets or monoliths. Their characterization usually involves the determination of the pore size distribution as well as the total pore volume or porosity. For some purposes it is also necessary to study the pore shape and interconnectivity, and to determine the internal and external surface area.

Porous materials have great technological importance, e.g. in the context of the following:

- a) controlled drug release;
- b) catalysis;
- c) gas separation;
- d) filtration including sterilization;
- e) materials technology;
- f) environmental protection and pollution control;
- g) natural reservoir rocks;
- h) building material properties;
- i) polymer and ceramic industries.

It is well established that the performance of a porous solid (e.g. its strength, reactivity, permeability or adsorbent power) is dependent on its pore structure. Many different methods have been developed for the characterization of pore structure. The choice of the most appropriate method depends on the application of the porous solid, its chemical and physical nature and the range of pore size.

Different methods for the characterization of nanopores are available, including spectroscopy, electron and tunnel microscopy and sorption methods. In view of the complexity of most porous solids, it is not surprising that the results obtained are not always in agreement and that no single technique can be relied upon to provide a complete picture of the pore structure. Among these, mercury porosimetry (see ISO 15901-1) and gas adsorption are popular ones because by combining both it is possible to assess a wide range of pore sizes from below 0,5 nm up to 400 μm . While mercury porosimetry is the standard technique for macropore analysis, gas adsorption techniques allow to assess pores up to approximately 100 nm. In this case, physical adsorption can be conveniently used, is not destructive, and is not that cost intensive as compared to some of the above-mentioned methods. Particularly, with regard to the

This is a preview of "ISO 15901-2:2022". [Click here to purchase the full version from the ANSI store.](#)

application of microporous material as specific sorbents, molecular sieves and carriers for catalysts and biological active material, the field-proven methods of gas sorption are of special value.

The measuring techniques of the method described in this document are similar to those described in ISO 9277 for the measurement of gas adsorption at low temperature. However, in order to assess the full range of pore sizes including microporosity, adsorption experiments have to be performed over a wide range of pressures from the ultralow pressure range (e.g. turbomolecular pump vacuum) up to atmospheric pressure (0,1 MPa).