

This is a preview of "ISO 16063-13:2001". [Click here to purchase the full version from the ANSI store.](#)

First edition
2001-12-01

Methods for the calibration of vibration and shock transducers —

Part 13: Primary shock calibration using laser interferometry

*Méthodes pour l'étalonnage des transducteurs de vibrations et de chocs —
Partie 13: Étalonnage primaire de chocs par interférométrie laser*



Reference number
ISO 16063-13:2001(E)

© ISO 2001

This is a preview of "ISO 16063-13:2001". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

This is a preview of "ISO 16063-13:2001". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
1 Scope	1
2 Normative references	1
3 Uncertainty of measurement	1
4 Requirements for apparatus	2
4.1 General	2
4.2 Shock machine based on rigid body motion of an anvil	2
4.3 Shock machine based on wave propagation inside a long thin bar	2
4.4 Seismic block(s) for shock machine and laser interferometer	4
4.5 Laser	4
4.6 Interferometer	4
4.7 Oscilloscope	7
4.8 Waveform recorder with computer interface	7
4.9 Computer with data-processing program	7
4.10 Filters	7
4.11 Other requirements	7
5 Ambient conditions	8
6 Preferred accelerations and pulse durations	8
7 Method	8
7.1 Test procedure	8
7.2 Data acquisition	9
7.3 Data processing	9
8 Reporting the calibration results	13
 Annexes	
A Expression of uncertainty of measurement in calibration	14
B Explanation of the procedures	16
C Alternative method of calculation of magnitude and phase shift of the complex sensitivity	20
Bibliography.....	22

This is a preview of "ISO 16063-13:2001". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 16063 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 16063-13 was prepared by Technical Committee ISO/TC 108, *Mechanical vibration and shock*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

ISO 16063 consists of the following parts, under the general title *Methods for the calibration of vibration and shock transducers*:

- *Part 1: Basic concepts*
- *Part 11: Primary vibration calibration by laser interferometry*
- *Part 12: Primary vibration calibration by the reciprocity method*
- *Part 13: Primary shock calibration using laser interferometry*
- *Part 21: Secondary vibration calibration by comparison*

Annex A forms a normative part of this part of ISO 16063. Annexes B and C are for information only.

This is a preview of "ISO 16063-13:2001". [Click here to purchase the full version from the ANSI store.](#)

Introduction

The shock sensitivity S_{sh} is determined, according to definition, as the relationship between the peak values of the accelerometer output quantity and the acceleration. S_{sh} is not a unique quantity but may vary depending on the duration and shape of the shock pulse and the bandwidth over which the sensitivity of the transducer under test and the frequency response of the optional conditioning amplifier are sufficiently uniform.

A unique quantity applicable for linearity tests of accelerometers is the complex sensitivity at a frequency f_n , calculated in the frequency domain. This part of ISO 16063 makes use of data-processing procedures which allow the magnitude S_n and phase shift $\Delta\varphi_n$ of the complex sensitivity to be calculated, in addition or alternatively to the shock sensitivity S_{sh} (cf. informative annex C).

The method specified in this part of ISO 16063 is based on the absolute measurement of the time history of the motion. This method fundamentally deviates from another shock calibration method which is based on the principle of the change in velocity, described in ISO 16063-1. The shock sensitivity therefore differs fundamentally from the shock calibration factor obtained by the latter method, but is in compliance with the calibration factor stated in ISO 5347-4¹⁾.

1) To be revised as ISO 16063-22.