

Second edition
2022-10

In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) —

Part 3: In vitro diagnostic instruments for professional use

*Dispositifs médicaux de diagnostic in vitro — Informations fournies
par le fabricant (étiquetage) —*

Partie 3: Instruments de diagnostic in vitro à usage professionnel



Reference number
ISO 18113-3:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 18113-3:2022". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Essential requirements	2
5 Labels and marking	2
5.1 General.....	2
5.2 Identification of the IVD instrument.....	2
5.2.1 IVD instrument name.....	2
5.2.2 Serial number.....	2
5.2.3 In vitro diagnostic use.....	2
5.2.4 Unique device identifier (UDI).....	3
6 Elements of the instructions for use	3
7 Content of the instructions for use	4
7.1 Manufacturer.....	4
7.2 Identification of the IVD instrument.....	4
7.2.1 IVD instrument name.....	4
7.2.2 Module and software identification.....	4
7.3 Intended use/Intended purpose.....	5
7.4 Storage and handling.....	5
7.5 Warnings and precautions and/or measures to be taken and limitations of use regarding the instrument.....	5
7.6 Instrument installation.....	6
7.6.1 General.....	6
7.6.2 Action upon delivery.....	6
7.6.3 Site preparation prior to installation.....	6
7.6.4 Bringing into operation.....	6
7.7 Theory of operation.....	7
7.8 Functions.....	7
7.9 Limitations.....	7
7.10 Preparation prior to operation.....	7
7.11 Operating procedure.....	7
7.12 Control procedure.....	8
7.13 Calculation of examination results.....	8
7.14 Special functions.....	8
7.15 Emergency samples.....	8
7.16 Shut-down procedure.....	8
7.17 Disposal information.....	8
7.18 Maintenance.....	9
7.19 Troubleshooting.....	9
7.20 Document control.....	9
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 18113-3:2009), which has been technically revised.

The main changes are as follows:

- Updated text to reflect changes in regulations and provide examples for clarity;
- Added information pertaining to unique device identifier-device identifier (UDI);
- Updated the Bibliography.

A list of all parts in the ISO 18113 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO 18113-3:2022". [Click here to purchase the full version from the ANSI store.](#)

Introduction

Manufacturers of in vitro diagnostic (IVD) instruments for professional use, supply users with information to enable the safe use and expected performance of their devices. The type and level of detail varies according to the intended uses and country-specific regulations.

The International Medical Device Regulators Forum (IMDRF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions can allow patients earlier access to new technologies and treatments. This document provides a basis for harmonization of labelling requirements for IVD instruments for professional use.

This document is concerned solely with information supplied with IVD instruments and equipment intended for professional use. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of general labelling concepts.

This document is intended to support the essential labelling requirements of all the IMDRF^[5] partners, as well as other countries that have or plan to enact labelling regulations for IVD medical devices.

For IVD instruments for professional use that are intended to be used as a system with reagents provided by the same manufacturer, this document is also intended to be used together with ISO 18113-1 and ISO 18113-2.