

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

First edition
2010-08-15

Sterilization of health care products — Dry heat — Requirements for the development, validation and routine control of a sterilization process for medical devices

*Stérilisation des produits de santé — Chaleur sèche — Exigences pour
l'élaboration, la validation et le contrôle de routine d'un processus de
stérilisation pour dispositifs médicaux*



Reference number
ISO 20857:2010(E)

© ISO 2010

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword	v
Introduction.....	vi
1 Scope	1
1.1 Inclusions	1
1.2 Exclusions.....	1
2 Normative references	2
3 Terms and definitions	2
4 Quality management system elements	10
4.1 Documentation	10
4.2 Management responsibility	10
4.3 Product realization	10
4.4 Measurement, analysis and improvement — Control of nonconforming product.....	10
5 Sterilizing agent characterization	11
5.1 Sterilizing agent.....	11
5.2 Microbicidal effectiveness.....	11
5.3 Material effects	11
5.4 Environmental considerations.....	11
6 Process and equipment characterization	11
6.1 Process characterization	11
6.2 Equipment characterization	11
7 Product definition	13
7.1 General	13
7.2 Product safety and performance	13
7.3 Packaging considerations	14
7.4 Microbiological quality.....	14
7.5 Product family.....	14
7.6 Biological safety	14
8 Process definition	15
9 Validation	16
9.1 General	16
9.2 Installation qualification	16
9.3 Operational qualification	16
9.4 Performance qualification	16
9.5 Additional sterilization systems	18
9.6 Review and approval of validation	18
10 Routine monitoring and control.....	19
10.1 Routine control	19
10.2 Routine monitoring	19
10.3 Process monitoring locations.....	20
11 Product release from sterilization/depyrogenation	21
12 Maintaining process effectiveness.....	21
12.1 General	21
12.2 Recalibration.....	21
12.3 Maintenance of equipment	21
12.4 Requalification.....	21
12.5 Assessment of change	22

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

Annex A (informative) Guidance on the application of this International Standard	23
Annex B (informative) Process definition based on inactivation of the microbial population in its natural state (bioburden-based approach)	46
Annex C (informative) Process definition based on the inactivation of reference microorganisms and knowledge of bioburden (combined bioburden/biological indicator approach)	48
Annex D (informative) Conservative process definition based on inactivation of reference microorganisms (overkill method)	51
Annex E (informative) Process development	54
Bibliography	57

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20857 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

Introduction

A sterile medical device is one that is free of viable microorganisms. International Standards that specify requirements for development, validation and routine control of sterilization processes, require, when it is necessary to supply a sterile medical device, that adventitious microbiological contamination of a medical device prior to sterilization be minimized. Even so, medical devices produced under standard manufacturing conditions in accordance with the requirements for quality management systems (see, for example, ISO 13485) may, prior to sterilization, have microorganisms on them, albeit in low numbers. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize medical devices can generally best be described by an exponential relationship between the numbers of microorganisms surviving and the extent of treatment with the sterilizing agent; inevitably this means that there is always a finite probability that a microorganism may survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the organisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization processing cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product.

This International Standard describes requirements that, if met, will provide a dry heat sterilization process capable of sterilizing medical devices through appropriate microbicidal activity. This International Standard also describes requirements that, if met, will provide a dry heat depyrogenation process through an appropriate denaturation activity. Furthermore, such compliance permits prediction, with reasonable confidence, that there is a low probability of there being a viable microorganism present on the product after processing. Specification of this probability is a matter for regulatory authorities and may vary from country to country (see for example EN 556-1 and ANSI/AAMI ST67). Additionally, there will be a low probability of pyrogenic material of bacterial origin being present on the product after the application of a depyrogenation process.

Generic requirements of the quality management systems for design/development, production, installation and servicing are given in ISO 9001 and particular requirements for quality management systems for medical device production in ISO 13485. The standards for quality management systems recognise that, for certain processes used in manufacturing or reprocessing, the effectiveness of the process cannot be fully verified by subsequent inspection and testing of the product. Sterilization and depyrogenation are examples of such processes. For this reason, sterilization and depyrogenation processes are validated for use, the performance of the processes is monitored routinely, and the equipment is maintained.

Exposure to a properly validated, accurately controlled sterilization process is not the only factor associated with the provision of reliable assurance that the product is sterile and, in this regard, suitable for its intended use. Attention is therefore given to a number of factors including:

- a) the microbiological status of incoming raw materials and/or components;
- b) the validation and routine control of any cleaning and disinfection procedures used on the product;
- c) the control of the environment in which the product is manufactured, assembled and packaged;
- d) the control of equipment and processes;
- e) the control of personnel and their hygiene;
- f) the manner and materials in which the product is packaged;
- g) the conditions under which product is stored.

This is a preview of "ISO 20857:2010". [Click here to purchase the full version from the ANSI store.](#)

These factors also need consideration for the provision of reliable assurance of depyrogenation.

The type of contamination on the product to be sterilized varies and this variation influences the effectiveness of a sterilization and depyrogenation process. Product that has been used in a health care setting and is being presented for resterilization in accordance with the manufacturer's instructions (see ISO 17664) should be regarded as a special case. There is potential for such product to possess a wide range of contaminating microorganisms and residual inorganic and/or organic contamination in spite of the application of a cleaning process. Hence, particular attention has to be given to the validation and control of the cleaning and disinfection processes used during reprocessing.

The requirements are the normative parts of this International Standard with which compliance is claimed. The guidance given in the informative annexes is not normative and is not provided as a check list for auditors. The guidance provides explanations as well as methods that are accepted as being suitable means for complying with the requirements. Approaches other than those given in the guidance may be used if they are effective in achieving compliance with the requirements of this International Standard.

The development, validation and routine control of a sterilization process and/or a depyrogenation process comprise a number of discrete but interrelated activities, for example calibration, maintenance, product definition, process definition, installation qualification, operational qualification and performance qualification. While the activities required by this International Standard have been grouped together and are presented in a particular order, this International Standard does not require that the activities be performed in the order that they are presented. The activities required are not necessarily sequential, as the programmes of development and validation might be iterative. It is possible that performing these different activities will involve a number of separate individuals and/or organizations, each of whom undertake one or more of these activities. This International Standard does not specify the particular individuals or organizations to carry out the activities.