



ISO 21018-1

**Hydraulic fluid power —
Monitoring the level of particulate
contamination of the fluid —**

**Part 1:
General principles**

*Transmissions hydrauliques — Surveillance du niveau de
pollution particulaire des fluides —*

Partie 1: Principes généraux

**Second edition
2024-10**

This is a preview of ISO 21018-1:2024. [Click here to purchase the full version from the ANSI store.](#)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of ISO 21018-1:2024. [Click here to purchase the full version from the ANSI store.](#)

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Health and safety	4
4.1 General.....	4
4.2 Electric power.....	4
4.3 Mechanical fluid power.....	4
4.4 Process liquids.....	4
4.4.1 Flammable or combustible liquids.....	4
4.4.2 Chemical compatibility.....	4
4.5 Electrical earthing/grounding.....	4
4.6 Environmental.....	4
5 Selection of monitoring technique	5
5.1 General.....	5
5.2 Selection.....	5
6 Procedures and precautions	5
6.1 General.....	5
6.2 Sampling.....	5
6.2.1 Obtaining representative samples.....	5
6.2.2 Off-line sampling.....	6
6.2.3 Off-line analysis.....	6
6.3 Analysis.....	6
6.3.1 On-line analysis.....	6
6.3.2 In-line analysis.....	6
6.3.3 Off-line analysis.....	7
6.4 Calibration procedures.....	7
6.5 Checking data repeatability.....	7
6.6 Training.....	8
6.7 Controlling the precision of the technique.....	8
7 Test report	8
Annex A (informative) Summary of various technique attributes	9
Annex B (informative) Description and relative merits of different contaminant monitoring techniques	15
Bibliography	23

This is a preview of ISO 21018-1:2024. [Click here to purchase the full version from the ANSI store.](#)

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 131, *Fluid power systems*, Subcommittee SC 6, *Contamination control*.

This second edition cancels and replaces the first edition (ISO 21018-1:2008), which has been technically revised. The main changes are as follows:

- [3.1](#) contains an updated definition for automated particle counter;
- [3.2](#) contains an updated definition for particle contamination model;
- [3.10](#) contains an updated definition for mesh;
- [3.11](#) now contains a note for the particle size definition;
- [B.8.1](#) has been updated to accurately describe the capabilities of image analysis.

A list of all parts in the ISO 21018 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of ISO 21018-1:2024. [Click here to purchase the full version from the ANSI store.](#)

In hydraulic fluid power systems, power is transmitted through a liquid under pressure within a closed circuit. The liquid is both a lubricant and a power-transmitting medium. The presence of solid particulate contamination in the liquid interferes with the ability of the hydraulic liquid to lubricate and causes wear to the components. The extent of this form of contamination in the liquid has a direct bearing on the performance and reliability of the system and it is necessary to control this to levels that are considered appropriate for the system concerned. Hydraulic oil filters are used to control the amount of particulate contamination to a level that is suitable for both the contaminant sensitivity of the system and the level of reliability required by the user.

Operators of hydraulic equipment are gradually defining maximum particle concentration levels for components, systems and processes. These are often referred to as the required cleanliness level (RCL). This cleanliness level is obtained by sampling the hydraulic liquid and measuring the particulate contamination level. If the contamination level is above the RCL, then corrective actions are necessary to reduce the contamination level. To avoid taking unnecessary actions, which can often prove costly, precision in sampling and measuring the particulate contamination level is required.

A comprehensive range of measurement equipment is available, but the instruments used are usually laboratory-based. This often requires that the equipment is operated in a special environment by specialist laboratories and this delays delivery of the test result to the user. To overcome this disadvantage, instruments are being continuously developed to determine the particulate contamination level, either using equipment that can be operated in or near the workplace or directly using on-line or in-line techniques. For equipment operated in the workplace, direct traceability to national measurement standards can be inappropriate, or irrelevant, as the instruments are used to monitor the general level of particulate contamination or to inform the user of a significant change in the level. When a significant change in the particulate contamination level is detected, the actual level is then usually qualified by using an approved particle-counting method. Also, these monitors can have simplified circuitry compared to similar laboratory units and this means that they can be less accurate and precise.

In addition, some instruments are designed to work on the “go/no-go” principle and their ability to rapidly evaluate the cleanliness level has resulted in an increase in their usage both in the fluid power industry and other markets. Unfortunately, the lack of a standardized method for their use, recalibration (if applicable) and means of checking the output validity means that the variability in the measurement data is at a level higher than is desirable.

This document has been developed to provide uniform and consistent procedures for instruments that are used for monitoring the contamination levels in hydraulic systems, especially those where direct traceability to national measurement standards is not possible or is not applicable.