

This is a preview of "ISO 21530:2004". [Click here to purchase the full version from the ANSI store.](#)

First edition
2004-06-15

Dentistry — Materials used for dental equipment surfaces — Determination of resistance to chemical disinfectants

Art dentaire — Matériaux utilisés pour les surfaces du matériel dentaire — Détermination de la résistance aux désinfectants chimiques



Reference number
ISO 21530:2004(E)

© ISO 2004

This is a preview of "ISO 21530:2004". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 21530:2004". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Sampling	1
5 Test methods	2
5.1 General	2
5.2 Preconditioning of test specimens	2
5.3 Preparation of test agent	2
5.4 Immersion test	2
5.5 Spray test	5
5.6 Contact test	7
6 Inspection and evaluation criteria	9
6.1 General	9
6.2 Visual inspection	9
6.3 Tactile inspection	9
7 Test report	9
Annex A (informative) Overview of chemical disinfectant solutions	11
Bibliography	12

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21530 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

This is a preview of "ISO 21530:2004". [Click here to purchase the full version from the ANSI store.](#)

Introduction

All materials used for external and touchable surfaces of dental equipment which can be contaminated by aerosols, splatters and droplets in normal use should be capable of undergoing disinfection without deterioration or discoloration when using the relevant disinfectant agent recommended by the manufacturer of the dental equipment.

This is a preview of "ISO 21530:2004". [Click here to purchase the full version from the ANSI store.](#)