

This is a preview of "ISO 21881:2019". [Click here to purchase the full version from the ANSI store.](#)

First edition
2019-10

Sterile packaged ready for filling glass cartridges

Cartouches en verre préremplissables sous emballage stérile



Reference number
ISO 21881:2019(E)

© ISO 2019



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 21881:2019". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Quality system	3
4.1 General.....	3
4.2 Testing.....	3
5 Process description and requirements	3
5.1 Washing.....	3
5.2 Drying.....	3
5.3 Lubrication.....	3
5.4 Capping and crimping.....	4
5.5 Plunger insertion.....	4
5.6 Packaging.....	4
5.7 Sterilization.....	4
6 Requirements for glassware	4
6.1 General.....	4
6.2 Material.....	4
6.3 Dimensions.....	5
6.4 Particles.....	5
6.4.1 Visible particles.....	5
6.4.2 Sub-visible particles.....	5
6.5 Bacterial endotoxin level.....	5
7 Requirements for packaging system	6
7.1 General.....	6
7.2 Nest and tub configuration.....	7
7.3 Tray configuration.....	7
7.4 Nest.....	7
7.5 Tub and tray.....	8
7.6 Insert liner.....	8
7.7 Sealing lid.....	8
7.8 Protective bag.....	8
7.9 Information to be provided by the manufacturer.....	9
8 Marking of the tub or tray	9
9 Labelling	10
Annex A (informative) Design of tub	11
Annex B (informative) Design of nest	12
Annex C (informative) Design of tray	13
Annex D (normative) Glide force test method to evaluate cartridge lubrication	14
Annex E (informative) Schematic illustrations of examples for the orientation of tubs or trays within the protective bag	16
Annex F (normative) Closure systems liquid leakage	19
Annex G (informative) Sample preparation for endotoxin and particulate determination	21
Annex H (informative) Product and packaging configuration	25
Bibliography	28

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO 21881:2019". [Click here to purchase the full version from the ANSI store.](#)

Introduction

In the last few years, following the more and more urgent request for ready for filling containers, packaging manufacturers managed to offer to the pharmaceutical industry containers already washed and sterilized. This category of products was born about 30 years ago with the appearance on the market of ready for filling syringes.

Only recently, the sterilized sub-assembled ready for filling syringes have been standardized by ISO 11040-4 and ISO 11040-7, including the corresponding packaging system. These two International Standards define the performance requirements of the glass syringes and the related test methods, as well as the ready for filling packaging system for these syringes, also including the test methods.

ISO 13926-1 specifies the design, dimensions, materials, performance and test methods for glass cylinders used with pen-injectors for medical use.

Due to the increasing market presence of syringes ready for filling and the associated advantages of this product for the pharmaceutical industry, the suppliers of packaging materials have started to develop such systems of this type for cartridges.

The availability of two packaging configurations makes ready for filling glass cartridges suitable for use both in clinical trials and in mass production. Nest and tub configuration has been conceived to be used usually with automated filling machines, while tray configuration is usually suitable for small batches filled manually or by means of semi-automated filling machines.

This duality of packaging configurations calls for a standardization of the production processes, materials quality and analytical methods when launching these products on the market, in order to avoid conceiving too highly customized processes.