

This is a preview of "ISO 22413:2010". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2010-06-15

Transfer sets for pharmaceutical preparations — Requirements and test methods

*Ensemble de transfert pour préparations pharmaceutiques —
Exigences et méthodes d'essai*



Reference number
ISO 22413:2010(E)

© ISO 2010

This is a preview of "ISO 22413:2010". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 22413:2010". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword	v
Introduction.....	vi
1 Scope	1
2 Normative references	1
3 Design and designation	1
3.1 Design.....	1
3.2 Design for a transfer set with housing.....	4
3.3 Designation	4
4 Material	4
5 Physical requirements	5
5.1 Particulate contamination.....	5
5.2 Tensile strength.....	5
5.3 Tightness.....	5
5.4 Free flow	5
5.5 Piercing device	5
5.6 Penetration force	5
5.7 Fragmentation.....	6
5.8 Air inlet and air outlet.....	6
5.9 Protective caps	6
5.10 Transfer sets with a housing.....	6
5.11 Luer connector	6
5.12 Filter for particles	6
6 Chemical requirements.....	6
7 Biological requirements.....	6
8 Testing of physical requirements	6
8.1 Particulate contamination.....	6
8.2 Tensile strength.....	7
8.3 Tightness of transfer set.....	7
8.4 Free flow	7
8.5 Piercing device	7
8.6 Penetration force	7
8.7 Testing on fragmentation	7
8.8 Effectiveness of air inlet and air outlet with air filter.....	7
8.9 Efficiency of protective caps.....	7
8.10 Luer connector	7
8.11 Filter for particles	7
9 Testing of chemical requirements	8
10 Testing of biological requirements.....	8
11 Packaging.....	8
12 Storage	8
13 Labelling	8
13.1 Unit container	8
13.2 Shelf or multi-unit container	8
Annex A (normative) Testing of fragmentation of transfer sets with plastic piercing devices	9

This is a preview of "ISO 22413:2010". [Click here to purchase the full version from the ANSI store.](#)

Annex B (normative) Testing of fragmentation of transfer sets with metal piercing devices	11
Bibliography	13

This is a preview of "ISO 22413:2010". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22413 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection equipment for medical and pharmaceutical use*.

This second edition cancels and replaces the first edition (ISO 22413:2007), of which the scope was enhanced by introducing further product groups like transfer sets with integrated Luer connectors and particle filters. In that framework the following major changes were introduced:

- the Introduction was amended by h) and j);
- the Normative references were updated;
- the Figures in 3.1 were updated;
- 5.11 and 8.10 on the physical requirements and testing for Luer connector were added;
- 5.12 and 8.11 on the physical requirements and testing for filter for particles were added.

Introduction

Transfer sets for pharmaceutical preparations transmit fluids from one container to another. The transfer sets mix fluids or dissolve dry substances and are used in combination with infusion and injection containers.

The transfer sets consist either of two piercing devices or of a piercing device in combination with a Luer connector, which may be connected with each other in different ways. Transfer sets may have a housing.

Examples of different designs:

- a) two piercing devices connected to each other (similar to piercing devices of infusion containers);
- b) a metal cannula, bevelled on both sides or a combination of a) and b);
- c) metal cannulae mostly having a hub or a grip plate in the middle to be fixed to the plastic part;
- d) plastic piercing devices directly connected to a grip plate, or held by a tube at a distance to allow a higher hydrostatic pressure;
- e) piercing devices with an additional ventilation channel that may end in the other tip or outside;
- f) piercing devices also with an air filter;
- g) piercing devices with housings serving, among other things, as a guide and a fixation on the connected containers for a secure, injury-free and contactless application;
- h) piercing device in combination with a Luer connector;
- i) piercing device in combination with a Luer connector and a particle filter.