

This is a preview of "ISO 23409:2011". [Click here to purchase the full version from the ANSI store.](#)

First edition  
2011-02-15

---

---

## Male condoms — Requirements and test methods for condoms made from synthetic materials

*Préservatifs masculins — Exigences et méthodes d'essai pour les préservatifs fabriqués en matières synthétiques*



Reference number  
ISO 23409:2011(E)

© ISO 2011

This is a preview of "ISO 23409:2011". [Click here to purchase the full version from the ANSI store.](#)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

This is a preview of "ISO 23409:2011". [Click here to purchase the full version from the ANSI store.](#)

## Contents

Page

Foreword .....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions .....	1
4 Quality verification .....	3
5 Lot size .....	4
6 Biocompatibility.....	4
7 Product claims.....	4
8 Design.....	4
9 Preclinical evaluation.....	5
10 Clinical (human use) investigations .....	6
11 Bursting volume and pressure .....	6
12 Freedom from holes .....	7
13 Stability and shelf-life .....	7
14 Visible defects .....	8
15 Package integrity for individual container.....	8
16 Packaging and labelling.....	8
17 Data sheet and test report .....	11
Annex A (normative) Sampling plans intended for assessing compliance of a continuing series of lots with sufficient number to allow the switching rules to be applied.....	12
Annex B (informative) Sampling plans that are intended for assessing the compliance of isolated lots .....	13
Annex C (normative) Determination of total lubricant for condoms in individual containers.....	14
Annex D (normative) Determination of length.....	16
Annex E (normative) Determination of width.....	18
Annex F (normative) Determination of thickness.....	19
Annex G (informative) Determination of barrier properties using the bacteriophage method.....	20
Annex H (normative) Determination of the bursting volume and pressure.....	24
Annex I (informative) Calibration of air inflation equipment for determination of burst volume and pressure .....	27
Annex J (normative) Testing for holes .....	31
Annex K (normative) Determination of shelf-life by real-time stability tests.....	36
Annex L (informative) Guidance on conducting and analysing ageing studies .....	38
Annex M (normative) Tests for individual container integrity.....	40
Annex N (informative) Oven treatment for condoms made from synthetic materials .....	42
Bibliography.....	43

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 23409 was prepared by Technical Committee ISO/TC 157, *Non-systemic contraceptives and STI barrier prophylactics*.

This is a preview of "ISO 23409:2011". [Click here to purchase the full version from the ANSI store.](#)

## Introduction

Synthetic condoms can be made from 100 % synthetic materials or a blend of synthetic materials and natural rubber latex. The material(s) used in synthetic condoms should be validated as constituting a barrier to human immunodeficiency virus (HIV), to other infectious agents responsible for the transmission of sexually transmitted infections (STIs), and to spermatozoa. It is essential that the condoms fit the penis properly, remain on the penis during use, are free from holes and have adequate physical strength so as not to break or tear during use so that the condoms can be deemed to be effective for contraceptive purposes and in order to help prevent the transmission of STIs. It is also important that they be correctly packaged so that they are protected during storage and suitably labelled. All of these issues are addressed in this International Standard.

To be safe, it is essential that the condom and any lubricant, additive, marking materials, dressing, individual packaging material or powder applied to it neither contain nor liberate substances in amounts that are toxic, likely to produce allergies (sensitization), locally irritating or otherwise harmful under normal conditions of storage and use.

Condoms are medical devices. To ensure high quality product, it is essential that condoms be produced under a quality management system using design controls. Reference can be made, for example, to ISO 9001<sup>[4]</sup>, to ISO 14971, and to ISO 13485<sup>[8]</sup>. Additional guidance can be found in ISO 16038<sup>[9]</sup>.

Condoms are non-sterile medical devices; however, a clean environment is essential to minimize microbiological contamination of the product during manufacturing and packaging.

Condoms can be of the designs given in the following terms, which are not intended to be exhaustive: smooth, textured, parallel-sided, non-parallel-sided, plain-ended, reservoir-ended, dry, lubricated, transparent, translucent, opaque, coloured, preshaped, welded or non-welded.

This International Standard specifies preclinical, clinical, and lot-by-lot physical requirement testing for condoms made from synthetic materials, including condoms made from a blend of synthetic materials and natural rubber latex. Application of lot-by-lot testing requirements becomes relevant only after the preclinical and clinical requirements of this International Standard have been met.