

This is a preview of "ISO 23747:2007". [Click here to purchase the full version from the ANSI store.](#)

First edition
2007-07-15

Anaesthetic and respiratory equipment — Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans

*Matériel d'anesthésie et de réanimation respiratoire — Débitmètres à
débit de pointe expiratoire pour l'évaluation de la fonction pulmonaire
chez les êtres humains respirant spontanément*



Reference number
ISO 23747:2007(E)

© ISO 2007

This is a preview of "ISO 23747:2007". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 23747:2007". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 General requirements.....	2
4.1 Safety for PEFMs that utilize electricity.....	2
4.2 Mechanical safety for all PEFMs	2
5 Identification, marking and documents.....	2
5.1 Marking of the scale or display	2
5.2 Marking of PEFM or packaging	3
5.3 Instructions for use	3
5.4 Technical description	4
6 PEFM measurement range.....	4
7 Performance requirements	4
7.1 Error of measurement	4
7.2 Linearity	4
7.3 Resistance to flow	5
7.4 Frequency response.....	5
8 Dismantling and reassembly	5
9 Effects of mechanical ageing	5
10 Effects of dropping a hand-held PEFM.....	5
11 Cleaning, sterilization and disinfection.....	5
11.1 Re-usable PEFM and parts	5
11.2 PEFM and parts delivered sterile	6
12 Compatibility with substances	6
13 Biocompatibility	6
Annex A (informative) Rationale for tests and examples of test apparatus.....	7
Annex B (normative) Determination of error, repeatability and resistance to PEFM output	10
Annex C (normative) Determination of frequency response	13
Annex D (normative) Test methods for determination of the effects of dismantling, ageing and dropping	15
Annex E (informative) Environmental aspects.....	17
Annex F (informative) Reference to the essential principals.....	19
Bibliography	22

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 23747 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*.

This is a preview of "ISO 23747:2007". [Click here to purchase the full version from the ANSI store.](#)

Introduction

The development of a standard for peak expiratory flowrate (PEF) measurement is considered important for the enhancement of the ability of clinicians to diagnose and monitor lung conditions by ensuring that all devices for such purposes meet minimum levels for safety and performance. An agreed standard means that peak expiratory flow meters (PEFM) can be tested to meet the same requirements with the latest accepted methods. Clinicians and patients can then be confident that these PEFM are fit for the purposes for which they are intended.

The American Thoracic Society has been foremost in proposing initial standards for testing PEFM [14]. They have proposed 26 waveforms for testing PEF, which are deemed suitable for checking that these PEFMs can correctly measure PEF.

The work of Miller et al. [16] first showed the problem of PEFM inaccuracy and they have recently defined the population characteristics of the PEF profile [18] and demonstrated limitations of pump systems for testing PEFM [17]. The European Respiratory Society has published a comprehensive statement on PEF [18].

This International Standard is based on the best currently available evidence concerning the methods and waveforms suited for testing PEFM [15].

Throughout this International Standard, text for which a rationale is provided in Annex A, is indicated by an asterisk (*).