

This is a preview of "ISO 5167-5:2022". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2022-10

Measurement of fluid flow by means of pressure differential devices inserted in circular cross-section conduits running full —

Part 5: Cone meters

*Mesurage de débit des fluides au moyen d'appareils déprimogènes
insérés dans des conduites en charge de section circulaire —*

Partie 5: Cônes de mesure



Reference number
ISO 5167-5:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 5167-5:2022". Click here to purchase the full version from the ANSI store.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principles of the method of measurement and computation	2
5 Cone meters	3
5.1 Field of application.....	3
5.2 General shape.....	3
5.3 Material and manufacture.....	7
5.4 Pressure tappings.....	8
5.5 Discharge coefficient, C	8
5.5.1 Limits of use.....	8
5.5.2 Discharge coefficient of the cone meter.....	8
5.6 Expansibility (expansion) factor, ε	9
5.7 Uncertainty of the discharge coefficient, C	9
5.8 Uncertainty of the expansibility (expansion) factor, ε	9
5.9 Pressure loss.....	9
6 Installation requirements	10
6.1 General.....	10
6.2 Minimum upstream and downstream straight lengths for installations between various fittings and the cone meter.....	10
6.2.1 General.....	10
6.2.2 Single 90° bend.....	11
6.2.3 Two 90° bends in perpendicular planes.....	11
6.2.4 Concentric expander.....	11
6.2.5 Partially closed valves.....	11
6.3 Additional specific installation requirements for cone meters.....	11
6.3.1 Circularity and cylindricality of the pipe.....	11
6.3.2 Roughness of the upstream and downstream pipe.....	11
6.3.3 Positioning of a thermowell.....	11
7 Flow calibration of cone meters	12
7.1 General.....	12
7.2 Test facility.....	12
7.3 Meter installation.....	12
7.4 Design of the test programme.....	12
7.5 Reporting the calibration results.....	13
7.6 Uncertainty analysis of the calibration.....	13
7.6.1 General.....	13
7.6.2 Uncertainty of the test facility.....	13
7.6.3 Uncertainty of the discharge coefficient of the cone meter.....	13
Annex A (informative) Table of expansibility (expansion) factor	14
Bibliography	15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 2, *Pressure differential devices*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/SS F05, *Measuring instruments*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition of ISO 5167-5 cancels and replaces the first edition (ISO 5167-5:2016), which has been technically revised.

The main changes are as follows:

- this document is consistent with ISO/IEC Guide 98-3;
- errors in [Figure 2](#) and in [5.2.7](#) have been corrected;
- the expansibility uncertainty is given as a relative uncertainty for ease of use with ISO 5167-1 (the calculated flow rate uncertainty is unchanged).

A list of all parts in the ISO 5167 series can be found on the ISO website.

This is a preview of "ISO 5167-5:2022". [Click here to purchase the full version from the ANSI store.](#)

Introduction

ISO 5167, consisting of six parts, covers the geometry and method of use (installation and operating conditions) of orifice plates, nozzles, Venturi tubes, cone meters and wedge meters when they are inserted in a conduit running full to determine the flow rate of the fluid in the conduit. It also gives necessary information for calculating the flow rate and its associated uncertainty. ISO 5167 (all parts) also provides methodology for bespoke calibration of differential pressure meters.

ISO 5167 (all parts) is applicable only to pressure differential devices in which the flow remains subsonic throughout the measuring section and where the fluid can be considered as single-phase, but is not applicable to the measurement of pulsating flow. Furthermore, each of these devices can only be used within specified limits of pipe size and Reynolds number, or alternatively they can be used across their calibrated range.

ISO 5167 (all parts) deals with devices for which direct calibration experiments have been made sufficient in number, spread, and quality to enable coherent systems of application to be based on their results and coefficients to be given with certain predictable limits of uncertainty.

The devices introduced into the pipe are called primary devices. The term primary device also includes the pressure tappings. All other instruments or devices required to facilitate the instrument readings are known as secondary devices, and the flow computer that receives these readings and performs the algorithms is known as a tertiary device. ISO 5167 (all parts) covers primary devices; secondary devices (see ISO 2186) and tertiary devices will be mentioned only occasionally.

Aspects of safety are not dealt with in ISO 5167 (all parts). It is the responsibility of the user to ensure that the system meets applicable safety regulations.