

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

Third edition
2008-06-15

Low-pressure hose assemblies for use with medical gases

*Flexibles de raccordement à basse pression pour utilisation avec les
gaz médicaux*



Reference number
ISO 5359:2008(E)

© ISO 2008

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword.....	v
Introduction	vi
0.1 General.....	vi
0.2 Standardization of screw-threaded connectors for use in hose assemblies	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions.....	2
4 General requirements.....	5
4.1 Safety	5
4.2 * Alternative construction	6
4.3 Materials	6
4.4 Design requirements	6
4.4.1 Hose internal diameter	6
4.4.2 Mechanical strength	6
4.4.3 Deformation under pressure	7
4.4.4 Resistance to occlusion.....	7
4.4.5 Adhesion strength	7
4.4.6 Flexibility	7
4.4.7 Gas-specificity	7
4.4.8 End connectors.....	8
4.4.9 Design of NIST connectors.....	8
4.4.10 Design of DISS connectors.....	8
4.4.11 Design of SIS connectors	16
4.4.12 Joining hoses to hose inserts	16
4.4.13 Leakage.....	17
4.4.14 * Pressure drop	17
4.4.15 Expulsion of nipple.....	17
4.5 Constructional requirements.....	17
4.5.1 * Cleaning	17
4.5.2 * Lubricants	18
5 Test methods.....	18
5.1 General.....	18
5.1.1 Ambient conditions	18
5.1.2 Test gas	18
5.1.3 Reference conditions	18
5.2 Test method for pressure drop	18
5.3 Test method for leakage	18
5.3.1 For all hose assemblies	18
5.3.2 For hose assemblies fitted with a hose assembly check valve	18
5.4 Test method for gas-specificity.....	18
5.5 Test method for mechanical strength.....	19
5.6 Test method for deformation under pressure	19
5.7 Test method for resistance to occlusion.....	19
5.8 Test method for durability of markings and colour coding.....	21
6 Marking, colour coding and packaging	21
6.1 Marking	21
6.2 Colour coding.....	22
6.3 Packaging	22

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

7	Information to be supplied by the manufacturer	22
Annex A (informative)	Rationale	24
Annex B (informative)	Environmental aspects	25
Annex C (informative)	Reported regional and national deviations of colour coding and nomenclature for medical gases	26
Bibliography		28

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5359 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

This third edition cancels and replaces the second edition (ISO 5359:2000), which has been technically revised.

Introduction

0.1 General

This International Standard has been prepared in response to the need for a safe method of connecting medical equipment to a fixed medical gas pipeline system or other medical gas supply system such that hose assemblies carrying different gases, or the same gas at different pressures, cannot be interchanged. Fixed medical gas pipelines, once installed, are rarely disturbed and are subjected to commissioning procedures to avoid the possibility of cross-connections or contamination of the medical gas conveyed. However, hose assemblies are subjected to physical wear and tear, misuse and abuse throughout their relatively short service life and are frequently connected to, and disconnected from, the medical equipment and the fixed pipeline.

While recognising that no system is absolutely safe, this International Standard includes those requirements considered necessary to prevent foreseeable hazards arising from the use of hose assemblies. Operators should be continually alert to the possibility of damage being caused by external factors, and therefore regular inspection and repair should be undertaken to ensure that hose assemblies continue to meet the requirements of this International Standard.

This International Standard pays particular attention to:

- suitability of materials;
- gas-specificity;
- cleanliness;
- testing;
- identification;
- information supplied.

Rationales for some of the requirements of this International Standard are given in Annex A. Such requirements are indicated by the asterisk (*) after the clause number in the main text.

0.2 Standardization of screw-threaded connectors for use in hose assemblies

Whilst the desirability of achieving agreement on a single International Standard for screw-threaded connectors has never been in doubt, the present pattern of usage has made such agreement impossible. Nevertheless, fears that proliferation of individual national standards or practices will eventually result in potentially dangerous cross-connection between components for different gases have led to the choice of three screw-threaded connector systems for inclusion in this International Standard.

The three systems of connectors, which are non-interchangeable, are diameter-index safety system (DISS), non-interchangeable screw-threaded (NIST) and sleeve indexed system (SIS). Tables 1 and 5 detail those gases and gas mixtures for which DISS, NIST and SIS connectors have been allocated. Dimensions of NIST connectors are given in Tables 2, 3 and 4 and Figures 2, 3, 4 and 5. Dimensions of DISS connectors can be obtained from the Compressed Gas Association Inc., 1725 Jefferson Davis Highway, Arlington, VA 22202, USA. Dimensions of SIS connectors can be obtained from Standards Australia, GPO Box 476 Sydney, New South Wales, 2001, Australia.

This is a preview of "ISO 5359:2008". [Click here to purchase the full version from the ANSI store.](#)

As an alternative to the screw-threaded connector, a "quick connector" which is gas-specific can be used at the inlet (outlet for vacuum) of the hose assembly, i.e. to connect the hose assembly to the fixed pipeline. Quick-connector systems of differing design should be non-interchangeable with each other in any one health-care facility.