

First edition
2022-09

Dentistry — Mobile dental units and dental patient chairs —

Part 2: Air, water, suction and wastewater systems

*Médecine bucco-dentaire — Units dentaires et fauteuils dentaires
patient mobiles —*

*Partie 2: Systèmes d'alimentation en air et en eau, d'aspiration et
d'évacuation des eaux usées*



Reference number
ISO 5467-2:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of ISO 5467-2:2022. Click [here](#) to purchase the full version from the ANSI store.

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Classification	2
4.1 Classification of dental air supply.....	2
4.2 Classification of suction system and suction air flow rate.....	2
5 Requirements	2
5.1 Connections from the mobile dental unit to dental handpiece.....	2
5.2 Supply connections to the mobile dental unit.....	3
5.3 Water and wastewater systems.....	3
5.3.1 Incoming water and bottled water.....	3
5.3.2 Materials used for construction of procedural water systems within the mobile dental unit.....	3
5.3.3 Backflow prevention device for mobile dental units connected to the external drinking water supply.....	3
5.3.4 Cuspidors.....	4
5.3.5 Water venturi.....	4
5.3.6 Particle filter.....	4
5.3.7 Bacterial filter.....	4
5.3.8 Bottled water system supplying procedural water or solution.....	4
5.3.9 Retraction.....	4
5.3.10 Treatment method for biofilm.....	4
5.3.11 Water sampling connection point.....	4
5.3.12 Wastewater drain connection.....	5
5.3.13 Wastewater container.....	5
5.4 Air system.....	5
5.4.1 Incoming dental air.....	5
5.4.2 Purity class of dental air supplied by an integrated compressed air source equipment.....	5
5.4.3 Particle filters.....	5
5.4.4 Bacterial filters.....	5
5.4.5 Condensate drain.....	5
5.4.6 Compressed air filter.....	5
5.4.7 Sound level of compressed air source equipment.....	6
5.5 Mobile dental unit suction systems.....	6
5.5.1 Maximum suction pressure.....	6
5.5.2 Suction pressure head loss.....	6
5.5.3 Configuration of cannula connectors and cannula.....	6
5.5.4 Operating hoses with cannula connectors.....	6
5.5.5 Solids filter.....	6
5.5.6 Air separator.....	6
5.5.7 Mobile dental unit suction source connection point.....	6
5.5.8 Suction flow rate.....	7
5.5.9 Bacterial filter.....	7
5.6 Test report.....	7
6 Sampling	7
7 Measurement and test methods	7
7.1 Visual inspection.....	7
7.1.1 Visual inspection of equipment.....	7
7.1.2 Visual inspection of documentation or test reports.....	7

This is a preview of ISO 5467-2:2022. [Click here to purchase the full version from the ANSI store.](#)

7.2	Particle filters test.....	8
7.3	Suction flow rate test.....	8
8	Instructions for use.....	8
9	Technical description.....	8
	Annex A (informative) Test report.....	9
	Bibliography.....	13

This is a preview of ISO 5467-2:2022. [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 5467 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document was developed in consideration of ISO 7494-2:2022, so as to be applicable to mobile dental units. The requirements in this document focus on certain technical aspects to be appropriate for international standardization. The requirements for microbiological aspects of the fluids transported by mobile dental units are acknowledged to be also worthy of standardization and requirements pertaining to the prevention, inhibition, and removal of mobile dental unit waterline biofilm are being developed.