

First edition
2003-04-01

Water quality — Determination of biochemical oxygen demand after n days (BOD _{n}) —

Part 1: Dilution and seeding method with allylthiourea addition

*Qualité de l'eau — Détermination de la demande biochimique en
oxygène après n jours (DBO _{n}) —*

*Partie 1: Méthode par dilution etensemencement avec apport
d'allylthiourée*



Reference number
ISO 5815-1:2003(E)

© ISO 2003

This is a preview of "ISO 5815-1:2003". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 5815-1:2003". Click here to purchase the full version from the ANSI store.

Contents		Page
Foreword		iv
Introduction		v
1 Scope		1
2 Normative references		1
3 Terms and definitions		2
4 Principle		2
5 Reagents		2
6 Apparatus		4
7 Storage of the sample		4
8 Procedure		5
8.1 Pretreatment		5
8.2 Preparation of test solutions		6
8.3 Blank test		7
8.4 Procedure		7
8.5 Control analysis		7
9 Calculation and expression of results		8
9.1 Examination for valid oxygen consumption during test		8
9.2 Calculation of biochemical oxygen demand after n days (BOD_n)		8
10 Special cases		8
11 Test report		9
Annex A (informative) Alternative incubation methods		10
Annex B (informative) Multitesting		12
Annex C (informative) Trueness and precision		13
Bibliography		15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5815-1 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

This first edition of ISO 5815-1, together with ISO 5815-2, cancels and replaces ISO 5815:1989, which has been technically revised.

ISO 5815 consists of the following parts, under the general title *Water quality — Determination of biochemical oxygen demand after n days (BOD_n)*:

- *Part 1: Dilution and seeding method with allylthiourea addition*
- *Part 2: Method for undiluted samples*

ISO 5815-1 is the equivalent of European Standard EN 1899-1.

This is a preview of "ISO 5815-1:2003". [Click here to purchase the full version from the ANSI store.](#)

Introduction

This part of ISO 5815 is a modified version of ISO 5815:1989, *Water quality — Determination of biochemical oxygen demand after 5 days (BOD₅) — Dilution and seeding method*.

The times of incubation specified in this part of ISO 5815 are 5 days, as in ISO 5815:1989 and as applied in many European countries, or 7 days, as applied in several Nordic countries for many years. The 7-day incubation typically gives higher BOD results than 5 days incubation time.

With an incubation period of 5 days, weekend work can only be avoided if samples are collected Wednesdays, Thursdays or Fridays. With an incubation period of 7 days, samples collected on the first five weekdays can be analysed without implying weekend work. For this reason, a 7-day incubation period can be considered more convenient than the conventional 5-day incubation.

A new, modified 7-day incubation period is described in Annex A. Early investigations indicate that BOD results obtained by this modified method are identical to results obtained by the 5-day method described in the main text of this part of ISO 5815. It is hoped that more comparative data on these two incubation methods will be obtained during the coming years, so that the modified 7-day incubation method can be included fully at the time of review and revision of this part of ISO 5815.

For the determination of BOD_n of water samples, the respirometric method described in ISO 9408 may also be used.

In this part of ISO 5815, the limit of determination, D_L , is defined as

$$D_L = t_{0,95(f)} \cdot 2 \cdot s_B \cdot \sqrt{1 + \frac{1}{n}} \quad (1)$$

where

s_B is the within-series standard deviation;

$t_{0,95(f)}$ is the Student t -value;

f is the degrees of freedom for the determination of s_B ;

n is the number of analyses for determination of the blank in an analytical series;

s_B is calculated from determinations of real samples with a BOD content near the estimated D_L .

In cases where the analytical method does not require any blank correction, the term

$$\sqrt{1 + \frac{1}{n}} \quad (2)$$

is omitted.