

This is a preview of "ISO 6975:1997". [Click here to purchase the full version from the ANSI store.](#)

STANDARD

6975

Second edition
1997-04-01

**Natural gas — Extended analysis —
Gas-chromatographic method**

*Gaz naturel — Analyse étendue — Méthode par chromatographie
en phase gazeuse*



Reference number
ISO 6975:1997(E)

Contents	Page
1 Scope	1
2 Normative references	2
3 Definitions	3
4 Principle	4
5 Analysis and analytical requirements	5
6 Procedures	7
7 Calibration and control charts	8
8 Calculations	11
9 Precision	13
10 Test report	14
 Annexes	
A Determination of the response curves for the main components in a sample	15
B Further details on the matrices used in annex A	21
C List of retention indices	22
D Bibliography	24

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet central@iso.ch
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

This is a preview of "ISO 6975:1997". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 6975 was prepared by Technical Committee ISO/TC 193, *Natural gas*, Subcommittee SC 1, *Analysis of natural gas*.

This second edition cancels and replaces the first edition (ISO 6975:1986), which has been technically revised.

Annexes A to D of this International Standard are for information only.

This is a preview of "ISO 6975:1997". [Click here to purchase the full version from the ANSI store.](#)

Introduction

This International Standard describes the specifications that an accurate method for the complete and extended analysis of natural gas shall fulfil.

The analytical methods involve injection of natural gas on to packed or open tubular (capillary) columns in one or more gas chromatographs. The components coming off the columns are detected by thermal-conductivity detector (TCD) or flame ionization detector (FID).

Compared to other analytical methods, the extended gas-chromatographic analysis supplies considerable knowledge of the individual components in the natural-gas mixture. This allows quantitative measurement of particular components in the C_6+ fraction and calculation of physical properties. Although the higher hydrocarbons influence physical properties such as calorific values and density by less than 0,3 % and 0,004 %, respectively, knowledge of them is required for evaluation of the pseudo-values for the C_6+ fraction.

In addition, vapour phase equilibrium calculations often require detailed composition data, especially of hydrocarbons with carbon numbers higher than 6.

The simultaneous determination of benzene and other aromatics is needed for gas consumers using natural gas as a chemical feedstock since these components can interfere in some processes.