

This is a preview of "ISO 7405:2008". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2008-12-15

Dentistry — Evaluation of biocompatibility of medical devices used in dentistry

*Art dentaire — Évaluation de la biocompatibilité des dispositifs
médicaux utilisés en art dentaire*



Reference number
ISO 7405:2008(E)

© ISO 2008

This is a preview of "ISO 7405:2008". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 7405:2008". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Categorization of medical devices	3
4.1 Categorization by nature of contact	3
4.2 Categorization by duration of contact.....	3
5 Biological evaluation process.....	4
5.1 General	4
5.2 Selection of tests and overall assessment.....	4
5.3 Selection of test methods.....	4
5.4 Types of test	5
5.5 Re-evaluation of biocompatibility.....	6
6 Test procedures specific to dental materials	6
6.1 Recommendations for sample preparation	6
6.2 Agar diffusion test.....	8
6.3 Filter diffusion test	10
6.4 Pulp and dentine usage test.....	13
6.5 Pulp capping test.....	17
6.6 Endodontic usage test	19
Annex A (informative) Types of test to be considered for evaluation of biocompatibility of medical devices used in dentistry.....	23
Annex B (informative) Dentine barrier cytotoxicity test.....	25
Annex C (informative) Acute toxicity testing	32
Bibliography.....	33

This is a preview of "ISO 7405:2008". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7405 was prepared by Technical Committee ISO/TC 106, *Dentistry*.

This second edition cancels and replaces the first edition (ISO 7405:1997) which has been technically revised. The following changes have been made:

- a) addition of dentine barrier cytotoxicity test to Annex B;
- b) improved description of test methods;
- c) updated cross-references to ISO 10993 series.

This is a preview of "ISO 7405:2008". [Click here to purchase the full version from the ANSI store.](#)

Introduction

This International Standard concerns the evaluation of the biocompatibility of medical devices used in dentistry. It is to be used in conjunction with the ISO 10993 series of standards. This International Standard contains special tests, for which ample experience exists in dentistry and which acknowledge the special needs of dentistry.

Only test methods for which the members of the committee considered there was sufficient published data have been included. In recommending test methods, the need to minimize the use of animals was given a high priority. It is essential that the decision to undertake tests involving animals be reached only after a full and careful review of the evidence indicating that a similar outcome cannot be achieved by other types of test. In order to keep the number of animals required for tests to an absolute minimum, consistent with achieving the objective indicated, it can be appropriate to conduct more than one type of test on the same animal at the same time, e.g. pulp and dentin usage test and pulp capping test. However, in accordance with ISO 10993-2 these tests are performed both in an efficient and humane way. On all occasions when animal testing is undertaken, such tests are conducted empathetically and according to standardized procedures as described for each test.

This International Standard does not explicitly describe test methods for occupationally related risks.

Annexes B and C are included to encourage the development of *in vitro* and *ex vivo* test methods which will further reduce the use of animals in the evaluation of the biocompatibility of medical devices used in dentistry.