

First edition
2011-12-15

Medical electrical equipment —

Part 2-55:

Particular requirements for the basic safety and essential performance of respiratory gas monitors

Appareils électromédicaux —

Partie 2-55: Exigences particulières relatives à la sécurité de base et aux performances essentielles des moniteurs de gaz respiratoires

Reference number
ISO 80601-2-55:2011(E)



This is a preview of "ISO 80601-2-55:2011". [Click here to purchase the full version from the ANSI store.](#)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 80601-2-55:2011". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	vi
Introduction.....	vii
1 Scope.....	1
201.1 Scope, object and related standards.....	1
201.1. 1 * Scope	1
201.1. 2 Object.....	2
201.1. 3 Collateral standards	2
201.1. 4 Particular standards	2
201.2 Normative references.....	3
201.3 Terms and definitions	4
201.4 General requirements	6
201.4. 3 ESSENTIAL PERFORMANCE	6
201.4. 3.101 * Additional requirements for ESSENTIAL PERFORMANCE	6
201.4. 3.102 Additional requirements for acceptance criteria	6
201.4. 6 * ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT	6
201.4.10.2.101 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	6
201.5 General requirements for testing ME EQUIPMENT	7
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	7
201.7 ME EQUIPMENT identification, marking and documents	7
201.7. 2.3 * Consult ACCOMPANYING DOCUMENTS.....	7
201.7. 2.101 * Additional requirements for marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts	7
201.7. 2.4.101 Additional requirements for ACCESSORIES	8
201.7. 2.13.101 * Additional requirements for physiological effects (safety signs and warning statements)	8
201.7. 2.17.101 Additional requirements for protective packaging	8
201.7. 4.3 Unit of measure.....	8
201.7. 9.1 General requirements.....	9
201.7. 9.2.1.101 * Additional general requirements.....	9
201.7. 9.2.2.101 * Additional requirements for warnings and safety notices	9
201.7. 9.2.5.101 Additional requirements for ME EQUIPMENT description	10
201.7. 9.2.8.101 * Additional requirements for start-up procedure.....	10
201.7. 9.2.9.101* Additional requirements for operating instructions	10
201.7. 9.2.13.101 * Additional requirements for maintenance.....	11
201.7. 9.2.14.101 * Additional requirements for ACCESSORIES, supplementary equipment, used material	11
201.7. 9.2.15.101* Additional requirements for environmental protection	11
201.7. 9.3.101 * Additional requirements for technical description.....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	12
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	12
201.10 Protection against unwanted and excessive radiation HAZARDS	12
201.11 Protection against excessive temperatures and other HAZARDS	12
201.11. 6.4 Leakage	12
201.11. 6.5 * Ingress of water or particulate matter into ME EQUIPMENT or ME SYSTEMS	13
201.11. 6.6 * Cleaning and disinfection of ME EQUIPMENT or ME SYSTEMS	13
201.11. 6.7 Sterilization of ME EQUIPMENT or ME SYSTEMS	13

This is a preview of "ISO 80601-2-55:2011". [Click here to purchase the full version from the ANSI store.](#)

201.11. 6.8	Compatibility with substances used with ME EQUIPMENT.....	13
201.11. 8.101	Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT.....	14
201.11. 8.101.1	* Supply failure TECHNICAL ALARM CONDITION.....	14
201.11. 8.101.2	* Settings and data storage following short interruptions or automatic switchover	14
201.11. 8.101.3	* Operation following long interruptions.....	14
201.11. 8.101.4	* RESERVE ELECTRICAL POWER SOURCE	14
201.11. 8.101.5	* RESERVE ELECTRICAL POWER SOURCE for transport outside a healthcare facility	15
201.12	Accuracy of controls and instruments and protection against hazardous outputs.....	15
201.12. 1	Accuracy of controls and instruments	15
201.12. 1.101	* Measurement accuracy.....	15
201.12. 1.101.1	General	15
201.12. 1.101.2	* DRIFT of MEASUREMENT ACCURACY	16
201.12. 1.101.3	* MEASUREMENT ACCURACY of GAS READINGS for gas mixtures	17
201.12. 1.102	* TOTAL SYSTEM RESPONSE TIME and rise time	17
201.12. 1.103	* Indication of units of measure for GAS READINGS	18
201.12. 1.104	* Indication of operating mode.....	19
201.13	HAZARDOUS SITUATIONS and fault conditions	19
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	19
201.15	Construction of ME EQUIPMENT.....	19
201.15. 3.5.101	* Additional requirements for rough handling.....	19
201.15. 3.5.101.1	* Shock and vibration	19
201.15. 3.5.101.2	* Shock and vibration for professional transportation	20
201.15. 101	* Mode of operation	21
201.16	ME SYSTEMS	21
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	21
201.101	* Interfering gas and vapour effects	22
201.102	* Gas leakage.....	22
201.103	* Port connector for DIVERTING RGM.....	22
201.104	* Minimum sampling flowrate	23
201.105	* Contamination of breathing systems.....	23
201.105. 1	Sampling tube	23
201.105. 2	Exhaust tube	23
202	Electromagnetic compatibility — Requirements and tests.....	23
202.6.2.1.7	* PATIENT simulation.....	23
202.6.2.1.10	Compliance criteria	23
202.6.2.3.1	* Requirements.....	24
206	Usability	24
206.6.2.2.2	Primary operating functions.....	24
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	24
208.6.1.2	* ALARM CONDITION priority.....	24
208.6.5.1	* General requirements.....	26
208.6.6.2.101	* Additional requirements for adjustable ALARM LIMIT	26
208.6.8.5.101	* Additional requirements for ALARM SIGNAL deactivation states, indication and access 26	
209	Requirements for environmentally conscious design.....	26

This is a preview of "ISO 80601-2-55:2011". [Click here to purchase the full version from the ANSI store.](#)

210	Requirements for the development of physiologic closed-loop controllers	26
211	Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the home healthcare environment.....	27
	Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	28
	201.C.1 Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	28
	201.C.4 ACCOMPANYING DOCUMENTS, general	28
	201.C.5 ACCOMPANYING DOCUMENTS, instructions for use.....	29
	201.C.6 ACCOMPANYING DOCUMENTS, technical description	30
	Annex D (informative) Symbols on marking	31
	Annex AA (informative) Particular guidance and rationale	33
	Annex BB (informative) Environmental aspects.....	43
	Annex CC (informative) Test gas mixtures for calibration	45
	Annex DD (informative) Reference to the essential principles	46
	Bibliography.....	48

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 80601-2-55 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines* and IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D, *Electromedical equipment*.

This first edition cancels and replaces ISO 21647:2004 and ISO 21647:2004/Cor.1:2005. This edition constitutes a minor technical revision and alignment with the third edition of IEC 60601-1.

ISO 80601 consists of the following parts, under the general title *Medical electrical equipment*:

- *Part 2-12: Particular requirements for basic safety and essential performance of critical care ventilators*
- *Part 2-13: Particular requirements for basic safety and essential performance of an anaesthetic workstation*
- *Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors*
- *Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement*
- *Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment*

IEC 80601-2-30: *Particular requirements for basic safety and essential performance of automated non-invasive sphygmomanometers*, IEC 80601-2-35: *Particular requirements for basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use*, IEC 80601-2-58: *Particular requirements for basic safety and essential performance of lens removal devices and vitrectomy devices for ophthalmic surgery*, IEC 80601-2-59: *Particular requirements for basic safety and essential performance of screening thermographs for human febrile temperature screening* and IEC 80601-2-60: *Particular requirements for basic safety and essential performance of dental equipment* are published by IEC.

This is a preview of "ISO 80601-2-55:2011". [Click here to purchase the full version from the ANSI store.](#)

Introduction

In this International Standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this International Standard, the term

- “clause” means one of the 17 numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 201.7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

References to clauses within this International Standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this International Standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this International Standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or a rationale related to that item in Annex AA.

The attention of Member Bodies and National Committees is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.