

Second edition
2018-02

Medical electrical equipment —

Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors

Appareils électromédicaux —

Partie 2-55: Exigences particulières relatives à la sécurité de base et aux performances essentielles des moniteurs de gaz respiratoires



Reference number
ISO 80601-2-55:2018(E)

© ISO 2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Published in Switzerland

This is a preview of "ISO 80601-2-55:2018". Click here to purchase the full version from the ANSI store.

Contents	Page
Foreword	iv
Introduction	vi
201.1 Scope, object, and related standards	1
201.2 Normative references	3
201.3 Terms and definitions	4
201.4 General requirements	6
201.5 General requirements for testing of ME EQUIPMENT	7
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	7
201.7 ME EQUIPMENT identification, marking, and documents	7
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14
201.10 Protection against unwanted and excessive radiation HAZARDS	14
201.11 Protection against excessive temperatures and other HAZARDS	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs	16
201.13 HAZARDOUS SITUATIONS and fault conditions	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	22
201.15 Construction of ME EQUIPMENT	22
201.16 ME SYSTEMS	24
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	24
201.101 *Interfering gas and vapour effects	24
201.102 *Gas leakage	25
201.103 *Port connectors for DIVERTING RGMS	25
201.104 *Sampling flowrate	25
201.105 *Contamination of breathing systems	25
201.106 FUNCTIONAL CONNECTION	25
202 Electromagnetic disturbances — Requirements and tests	27
206 USABILITY	27
208 General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	28
211 General requirements, tests and guidance for medical electrical equipment and medical electrical systems used in the home healthcare environment	30
212 General requirements, tests and guidance for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS intended for use in the EMERGENCY MEDICAL SERVICES ENVIRONMENT	30
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	31
Annex D (informative) Symbols on marking	35
Annex AA (informative) Particular guidance and rationale	38
Annex BB (informative) Test gas mixtures for calibration	48
Annex CC (informative) Data interface requirements	49
Annex DD (informative) Alphabetized index of defined terms used in this document	54
Bibliography	61

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*.

This second edition cancels and replaces the first edition (ISO 80601-2-55:2011), which has been technically revised.

The main changes compared to the previous edition are as follows:

- additional requirements on respiratory gas monitors for use during professional transport of a patient outside a healthcare facility have been deleted because these are now covered by IEC 60601-1-12;
- requirements on marking, warning and safety notices, as well as accompanying documents have been updated;
- 201.11.6.5 and 201.15.3.5 have been revised to distinguish between requirements for stand-alone respiratory gas monitors and requirements for respiratory gas monitors that are incorporated into another medical electrical equipment;
- requirements on port connectors for diverting respiratory gas monitors have been revised;
- a new subclause on functional connection has been added (see 201.106) accompanied by the related rationale and informative annex on data interface requirements;

This is a preview of "ISO 80601-2-55:2018". [Click here to purchase the full version from the ANSI store.](#)

- Clause 202 has been updated to align with IEC 60601-1-2:2014;
- Clause 208 has been updated to align with IEC 60601-1-8:2006/Amd 1:2012;
- IEC 60601-1-9 has been excluded;
- Annex BB has been deleted;
- requirements on calibration/zeroing have been added.

A list of all the parts of ISO 80601 can be found on the ISO website.

Introduction

In this document, the following print types are used:

- requirements and definitions: roman type.
- compliance checks: *italic type*.
- informative material appearing outside of tables such as notes, examples and references: smaller type. Normative text of tables is also in a smaller type.
- terms defined in Clause 3 of the general standard, in this document or as noted: SMALL CAPITALS.

In referring to the structure of this document,

- “clause” means one of the 17 numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes 7.1, 7.2, etc.), and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 201.7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or”, so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in ISO/IEC Directives, Part 2, Clause 7. For the purposes of this document, the auxiliary verb

- “shall” means that compliance with a requirement or a test is mandatory for compliance with document,
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document, and
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.