

First edition
2011-04-01

Medical electrical equipment —
Part 2-61:
Particular requirements for basic safety
and essential performance of pulse
oximeter equipment

Appareils électromédicaux —

*Partie 2-61: Exigences particulières pour la sécurité de base
et les performances essentielles pour les oxymètres de pouls*

Reference number
ISO 80601-2-61:2011(E)



This is a preview of "ISO 80601-2-61:2011". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO 80601-2-61:2011". Click here to purchase the full version from the ANSI store.

Contents

Page

Foreword	vi
Introduction.....	viii
201.1 Scope, object and related standards.....	1
201.1. 1 * Scope.....	1
201.1. 2 Object.....	1
201.1. 3 Collateral standards	2
201.1. 4 Particular standards	2
201.2 Normative references.....	3
201.3 Terms and definitions	4
201.4 General requirements	8
201.4. 3 ESSENTIAL PERFORMANCE	8
201.4. 101 * Additional requirements for ESSENTIAL PERFORMANCE	8
201.4. 102 Additional requirements for acceptance criteria	9
201.4. 103 Additional requirements for PULSE OXIMETER EQUIPMENT, parts and ACCESSORIES	9
201.5 General requirements for testing of ME EQUIPMENT.....	9
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.7 ME EQUIPMENT identification, marking and documents	10
201.7. 2.3 Consult ACCOMPANYING DOCUMENTS	10
201.7. 2.9 IP classification	10
201.7. 2.101 Additional requirements for marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts.....	10
201.7. 2.4.101 Additional requirements for ACCESSORIES.....	11
201.7. 2.13.101 Additional requirements for physiological effects.....	11
201.7. 2.17.101 Additional requirements for protective packaging.....	11
201.7. 4.3 Unit of measure.....	12
201.7. 9.1 Additional general requirements.....	12
201.7. 9.2.1.101 Additional general requirements.....	12
201.7. 9.2.2.101 Additional requirements for warnings and safety notices	13
201.7. 9.2.8.101 Additional requirements for start-up PROCEDURE.....	13
201.7. 9.2.9.101 Additional requirements for operating instructions	13
201.7. 9.2.14.101 Additional requirements for ACCESSORIES, supplementary equipment, used material	14
201.7. 9.3.1.101 * Additional general requirements	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	15
201.8. 3.101 Additional requirements for classification of APPLIED PARTS.....	15
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	15
201.10 Protection against unwanted and excessive radiation HAZARDS	15
201.11 Protection against excessive temperatures and other HAZARDS	15
201.11. 6.5.101 * Additional requirements for ingress of water or particulate matter into ME EQUIPMENT or ME SYSTEM	16
201.11. 8.101 Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT	16
201.11. 8.101.1 Supply failure TECHNICAL ALARM CONDITION	16
201.11. 8.101.2 Settings and data storage following short interruptions or automatic switchover	16
201.11. 8.101.3 Operation following long interruptions	17
201.12 ACCURACY of controls and instruments and protection against hazardous outputs.....	17

This is a preview of "ISO 80601-2-61:2011". Click here to purchase the full version from the ANSI store.

201.12. 1	ACCURACY of controls and instruments	17
201.12. 1.101	* SpO_2 accuracy of pulse oximeter equipment	17
201.12. 1.101.1	* Specification.....	17
201.12. 1.101.2	Determination of SpO_2 ACCURACY	18
201.12. 1.101.2.1	* Data collection	18
201.12. 1.101.2.2	* Data analysis	18
201.12. 1.101.2.3	Characteristics of the clinical study population	19
201.12. 1.102	Accuracy under conditions of motion	19
201.12. 1.103	ACCURACY under conditions of low perfusion	19
201.12. 1.104	Pulse rate ACCURACY.....	19
201.12. 4	Protection against hazardous output.....	19
201.12. 4.101	* DATA UPDATE PERIOD.....	19
201.12. 4.102	* Signal inadequacy	20
201.13	HAZARDOUS SITUATIONS and fault conditions	20
201.13. 101	Detection of PULSE OXIMETER PROBE faults and PROBE CABLE EXTENDER faults	20
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	20
201.15	Construction of ME EQUIPMENT.....	21
201.15. 3.5.101	* Additional requirements for rough handling	21
201.15. 3.5.101.1	* Shock and vibration.....	21
201.15. 3.5.101.2	* Shock and vibration for professional transport	22
201.15. 101	Mode of operation.....	23
201.16	ME SYSTEMS.....	23
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	23
201.101	* PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS	23
201.101. 1	General.....	23
201.101. 2	Labelling	24
201.102	Saturation pulse INFORMATION SIGNAL	24
201.103	SIGNAL INPUT/OUTPUT PART	24
201.103. 1	General.....	24
201.103. 2	Connection to electronic health record.....	24
201.103. 3	Connection to a distributed alarm system.....	25
201.103. 4	Connection for remote control.....	25
202	Medical electrical equipment — Part 1-2: General requirements for safety — Collateral standard: Electromagnetic compatibility — Requirements and tests	25
202.6.2.1.1	IMMUNITY TEST LEVELS	25
202.6.2.1.7	PATIENT simulation	25
202.6.2.1.10	* Requirements.....	26
202.6.2.3	* Radiated RF electromagnetic fields.....	26
208	Medical electrical equipment — Part 1-8: General requirements for safety — Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	26
208.6.1.2.101	* Additional requirements for ALARM CONDITION priority.....	26
208.6.5.4.101	* Additional requirements for DEFAULT ALARM PRESET	27
208.6.8.5.101	Additional requirements for ALARM SIGNAL inactivation states, indication and access	27
Annex C (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	28
Annex D (informative)	Symbols on marking.....	32
Annex AA (informative)	Particular guidance and rationale.....	34

This is a preview of "ISO 80601-2-61:2011". [Click here to purchase the full version from the ANSI store.](#)

Annex BB (informative) Skin temperature at the PULSE OXIMETER PROBE	41
Annex CC (informative) Determination of ACCURACY	45
Annex DD (informative) Calibration standards	53
Annex EE (informative) Guideline for evaluating and documenting S_{pO_2} ACCURACY in human subjects	54
Annex FF (informative) Simulators, calibrators and FUNCTIONAL TESTERS for PULSE OXIMETER EQUIPMENT	61
Annex GG (informative) Concepts of ME EQUIPMENT response time	70
Annex HH (informative) Reference to the essential principles of safety and performance of medical devices in accordance with ISO/TR 16142	74
Bibliography	76
Alphabetized index of defined terms used in this standard	81

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

ISO 80601-2-61 was prepared by a Joint Working Group of Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment* and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*.

This first edition of ISO 80601-2-61 cancels and replaces the second edition of ISO 9919:2005, which has been revised to harmonize it with the third edition of IEC 60601-1:2005.

In this standard, the following print types are used.

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS TYPE.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

This is a preview of "ISO 80601-2-61:2011". [Click here to purchase the full version from the ANSI store.](#)

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.

Introduction

The approximation of arterial haemoglobin saturation and pulse rate using pulse oximetry is common practice in many areas of medicine. This standard covers BASIC SAFETY and ESSENTIAL PERFORMANCE requirements achievable within the limits of existing technology.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committee that led to a requirement and identifying the HAZARDS that the requirement addresses.

Annex BB is a literature survey relevant to the determination of the maximum safe temperature of the interface between a PULSE OXIMETER PROBE and a PATIENT'S tissue.

Annex CC discusses both the formulae used to evaluate the S_pO_2 ACCURACY of PULSE OXIMETER EQUIPMENT measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on when *in vitro* blood calibration of PULSE OXIMETER EQUIPMENT is needed.

Annex EE presents a guideline for a CONTROLLED DESATURATION STUDY for the calibration of PULSE OXIMETER EQUIPMENT.

Annex FF is a tutorial introduction to several kinds of testers used in pulse oximetry.

Annex GG describes concepts of PULSE OXIMETER EQUIPMENT response time.