

**ISO 80601-2-67****Medical electrical equipment —  
Part 2-67:  
Particular requirements for basic  
safety and essential performance of  
oxygen-conserving equipment**

*Appareils électromédicaux —*

*Partie 2-67: Exigences particulières pour la sécurité de base et les performances essentielles des économiseurs d'oxygène*

**Third edition  
2026-04**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2026

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

| <b>Contents</b>   | Page |
|---|------|
| Foreword .....  | iv   |
| Introduction.....   | vi   |
| 201.1 Scope, object and related standards.....  | 1    |
| 201.2 Normative references .....  | 3    |
| 201.3 Terms and definitions.....  | 4    |
| 201.4 General requirements.....   | 13   |
| 201.5 General requirements for testing of <i>ME equipment</i> .....   | 15   |
| 201.6 Classification of <i>ME equipment</i> and <i>ME systems</i> .....   | 16   |
| 201.7 <i>ME equipment</i> identification, marking and documents .....   | 17   |
| 201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i> .....   | 23   |
| 201.9 Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i> .....                           | 23   |
| 201.10 Protection against unwanted and excessive radiation <i>hazards</i> .....   | 23   |
| 201.11 Protection against excessive temperatures and other <i>hazards</i> .....   | 24   |
| 201.12 Accuracy of controls and instruments and protection against hazardous outputs .....                                      | 26   |
| 201.13 <i>Hazardous situations</i> and fault conditions .....   | 29   |
| 201.14 <i>Programmable electrical medical systems (PEMS)</i> .....  | 29   |
| 201.15 Construction of <i>ME equipment</i> .....  | 30   |
| 201.16 <i>ME systems</i> .....  | 30   |
| 201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i> .....   | 31   |
| 201.101 Gas connections .....   | 31   |
| 201.102 Requirements for parts and <i>accessories</i> .....   | 32   |
| 201.103 Oxygen pressure regulators .....  | 33   |
| 202 Electromagnetic disturbances – Requirements and tests .....   | 34   |
| 206 Usability .....   | 34   |
| Annex C (informative) Guide to <i>marking</i> and labelling requirements for <i>ME equipment</i> and<br><i>ME systems</i> ..... | 36   |
| Annex D (informative) <i>Symbols on marking</i> .....   | 41   |
| Annex AA (informative) Particular guidance and rationale .....  | 42   |
| Annex BB (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances ...                           | 51   |
| Annex CC (informative) Terminology — Alphabetized index of defined terms .....  | 55   |
| Bibliography .....  | 58   |

This is a preview of ISO 80601-2-67:2026. Click here to purchase the full version from the ANSI store.

## Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives) or [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs)).

ISO and IEC draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO and IEC take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO and IEC had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents) and <https://patents.iec.ch>. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html). In the IEC, see [www.iec.ch/understanding-standards](http://www.iec.ch/understanding-standards).

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Medical equipment, software, and systems*, Subcommittee SC D, *Particular medical equipment, software, and systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 80601-2-67:2020), which has been technically revised.

The main changes are as follows:

- updated references, where appropriate;
- harmonization with ISO 20417, where appropriate;
- updated uncertainty of measurement requirements;
- added *marking* requirements for *gas intake port*, external gas sources and MR compatibility;
- requirements for *processing* of the *enclosure*;
- added *cybersecurity* recommendations; and

This is a preview of ISO 80601-2-67:2026. Click here to purchase the full version from the ANSI store.

— updated *connector* requirements.

A list of all parts in the ISO 80601 and IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html) and [www.iec.ch/national-committees](http://www.iec.ch/national-committees).

## Introduction

Long-term oxygen therapy has been demonstrated in randomized, controlled clinical trials to prolong survival in *patients* with chronic respiratory disease and documented hypoxemia. Typical sources of therapeutic long-term oxygen therapy include gaseous oxygen from cylinders or from liquid oxygen and oxygen from an oxygen concentrator.

Most clinicians prescribe low flow oxygen therapy as continuous flow oxygen (CFO) delivery in l/min. CFO systems deliver the flow of oxygen without regard for the *patient's* breathing rate or pattern. Outside of the institutional care setting, the provision of CFO therapy is often a significant expense and can limit the mobility of a *patient* to the immediate vicinity of a stationary or fixed oxygen delivery system. To support mobility, *patients* use CFO from portable liquid or compressed oxygen systems with a limited storage capacity that can limit a *patient's* time and activities while away from a stationary oxygen supply.

*Conserving equipment* that delivers supplemental oxygen as a bolus conserves usage while allowing satisfactory *patient* arterial oxygen saturation ( $\text{SaO}_2$ ) to be maintained during daily activities. *Conserving equipment* delivers supplemental oxygen unlike CFO in that the therapy gas flow is delivered only during the inspiratory phase of the breathing cycle, when it is most likely to reach the alveoli. During both the expiratory and pause phase of the breathing cycle, the flow of supplemental oxygen is stopped, minimizing waste. Because flow over time produces a volume, the bolus delivered by the *conserving equipment* is typically represented as a volume of gas. Therapy using *conserving equipment* versus CFO results in lower operating costs and longer ambulatory times for *patients* using the same CFO storage capacity.

Operation of *conserving equipment* from various *manufacturers* can differ in the dose delivery mechanism resulting in variations in oxygen therapy to the *patient*. The use of CFO numerical *markings* for dose settings on *conserving equipment* can not directly correlate with CFO settings and can lead to misinterpretation of gas delivery rates and volumes for a particular *patient*. This can result in incorrect *patient* setup and therapy delivery over all breathing rates and patterns versus CFO. Because of the differences in delivery, settings, and *markings* versus CFO therapy, *conserving equipment* use has requirements for *patient* titration to determine the proper setting(s) needed to provide adequate  $\text{SaO}_2$  levels for the *patient* breathing patterns.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *terms defined in Clause 3 of the general standard, in this particular document or as noted: italic type; and*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term.

- “clause” means one of the three numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

This is a preview of ISO 80601-2-67:2026. [Click here to purchase the full version from the ANSI store.](#)

For the purposes of this document, the auxiliary verb:

- "shall" indicates a requirement;
- "should" indicates a requirement or a test is recommendation;
- "may" indicates a permission;
- "can" is used to describe a possibility or capability; and
- "must" is used to express an external constraint.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Annex D contains a summary of the *symbols* referenced in this document.

Requirements in this document have been decomposed so that each requirement is uniquely delineated. This is done to support automated requirements tracking.

This is a preview of ISO 80601-2-67:2026. Click here to purchase the full version from the ANSI store.

## Medical electrical equipment —

Part 2-67:

# Particular requirements for basic safety and essential performance of oxygen conserving equipment

### 201.1 Scope, object and related standards

NOTE 1 There is guidance or rationale for this clause contained in Clause AA.2.1.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

NOTE 2 The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

#### 201.1.1 Scope

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.1 is replaced by:

This document is applicable to the *basic safety* and *essential performance* of *oxygen conserving equipment*, hereafter referred to as *ME equipment*, in combination with its *accessories* intended to conserve supplemental oxygen by delivering gas intermittently and synchronized with the *patient's* inspiratory cycle, when used in the *home healthcare environment*. *Oxygen conserving equipment* is typically used by a *lay operator*.

NOTE 1 *Conserving equipment* can also be used in professional health care facilities.

This document is also applicable to *conserving equipment* that is incorporated with other equipment.

EXAMPLE *Conserving equipment* combined with a pressure regulator<sup>[4]</sup>, an oxygen concentrator<sup>[12]</sup> or liquid oxygen equipment<sup>[7]</sup>.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to *conserving equipment*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *conserving equipment*.

This document is intended to clarify the difference in operation of various *conserving equipment* models, as well as between the operation of *conserving equipment* and continuous flow oxygen equipment, by requiring standardized performance testing and labelling.

This document is only applicable to active devices (e.g. pneumatically or electrically powered) and is not applicable to non-active devices (e.g. reservoir cannulas).

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.