

First edition  
2018-07

---

---

**Medical electrical equipment —**  
**Part 2-80:**  
**Particular requirements for basic**  
**safety and essential performance of**  
**ventilatory support equipment for**  
**ventilatory insufficiency**

*Appareils électromédicaux —*

*Partie 2-80: Exigences particulières pour la sécurité de base et les performances essentielles des équipements d'assistance ventilatoire en cas d'insuffisance ventilatoire*



Reference number  
ISO 80601-2-80:2018(E)

© ISO 2018

This is a preview of "ISO 80601-2-80:2018". [Click here to purchase the full version from the ANSI store.](#)



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

This is a preview of "ISO 80601-2-80:2018". Click here to purchase the full version from the ANSI store.

## Contents

Foreword.....	vi
Introduction.....	viii
<b>201.1 Scope, object and related standards.....</b>	<b>1</b>
<b>201.1.1 * Scope.....</b>	<b>1</b>
<b>201.1.2 Object.....</b>	<b>2</b>
<b>201.1.3 Collateral standards.....</b>	<b>2</b>
<b>201.1.4 Particular standards.....</b>	<b>3</b>
<b>201.2 Normative references.....</b>	<b>3</b>
<b>201.3 Terms and definitions.....</b>	<b>5</b>
<b>201.4 General requirements.....</b>	<b>7</b>
<b>201.4.3 Essential performance.....</b>	<b>7</b>
<b>201.4.3.101 * Additional requirements for ESSENTIAL PERFORMANCE.....</b>	<b>7</b>
<b>201.4.6 * ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT.....</b>	<b>7</b>
<b>201.4.11.101 * Additional requirements for pressurized gas input.....</b>	<b>8</b>
<b>201.5 General requirements for testing of ME EQUIPMENT.....</b>	<b>9</b>
<b>201.5.101 * Additional requirements for the general requirements for testing of     ME EQUIPMENT.....</b>	<b>9</b>
<b>201.5.101.1 Ventilatory support equipment test conditions.....</b>	<b>9</b>
<b>201.5.101.2 * Gas flowrate and leakage specifications.....</b>	<b>9</b>
<b>201.5.101.3 * VENTILATORY SUPPORT EQUIPMENT testing errors.....</b>	<b>9</b>
<b>201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....</b>	<b>10</b>
<b>201.6.101 * Additional requirements for classification of ME EQUIPMENT and ME SYSTEMS.....</b>	<b>10</b>
<b>201.7 ME EQUIPMENT identification, marking and documents.....</b>	<b>10</b>
<b>201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....</b>	<b>17</b>
<b>201.9 Protection against mechanical hazards of ME EQUIPMENT and ME SYSTEMS.....</b>	<b>17</b>
<b>201.10 Protection against unwanted and excessive radiation HAZARDS.....</b>	<b>19</b>
<b>201.11 Protection against excessive temperatures and other HAZARDS.....</b>	<b>19</b>
<b>201.11.7 Biocompatibility of ME EQUIPMENT and ME SYSTEMS.....</b>	<b>20</b>
<b>201.11.8 Interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT.....</b>	<b>21</b>
<b>201.11.8.101 Additional requirements for interruption of the power supply/SUPPLY MAINS to       ME EQUIPMENT ALARM CONDITION.....</b>	<b>21</b>
<b>201.12 Accuracy of controls and instruments and protection against hazardous outputs.....</b>	<b>23</b>
<b>201.12.1 Accuracy of controls and instruments.....</b>	<b>23</b>
<b>201.12.1.101 Volume-controlled breath type.....</b>	<b>23</b>
<b>201.12.1.102 Pressure-controlled breath type.....</b>	<b>26</b>
<b>201.12.1.103 Other breath types.....</b>	<b>28</b>
<b>201.12.2.101 Usability of me equipment.....</b>	<b>29</b>
<b>201.12.4 Protection against hazardous output.....</b>	<b>29</b>
<b>201.12.4.101 * Measurement of AIRWAY PRESSURE.....</b>	<b>29</b>
<b>201.12.4.102 Measurement of expired volume.....</b>	<b>31</b>
<b>201.12.4.103 * Maximum limited pressure protection device.....</b>	<b>31</b>
<b>201.12.4.104 Hypoventilation ALARM CONDITION.....</b>	<b>31</b>
<b>201.12.4.105 * High leakage ALARM CONDITION.....</b>	<b>31</b>
<b>201.12.4.106 * CO<sub>2</sub> rebreathing.....</b>	<b>32</b>
<b>201.12.101 * Protection against accidental adjustments.....</b>	<b>32</b>

This is a preview of "ISO 80601-2-80:2018". [Click here to purchase the full version from the ANSI store.](#)

<b>201.13 Hazardous situations and fault conditions for ME EQUIPMENT</b> .....	<b>33</b>
<b>201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)</b> .....	<b>34</b>
<b>201.15 Construction of ME EQUIPMENT</b> .....	<b>34</b>
<b>201.15.101 Mode of operation</b> .....	<b>34</b>
<b>201.15.102 Pre-use check</b> .....	<b>34</b>
<b>201.16 ME SYSTEMS</b> .....	<b>34</b>
<b>201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS</b> .....	<b>35</b>
<b>201.101 Gas connections</b> .....	<b>35</b>
<b>201.101.1 VBS connectors</b> .....	<b>35</b>
<b>201.101.1.1 General</b> .....	<b>35</b>
<b>201.101.1.2 Other named ports</b> .....	<b>35</b>
<b>201.102 Requirements for the VBS and ACCESSORIES</b> .....	<b>36</b>
<b>201.102.1 * General</b> .....	<b>36</b>
<b>201.102.2 Labelling</b> .....	<b>37</b>
<b>201.102.3 Breathing sets</b> .....	<b>37</b>
<b>201.102.4 * Humidification</b> .....	<b>37</b>
<b>201.102.4.1 HUMIDIFIER</b> .....	<b>37</b>
<b>201.102.4.2 HEAT AND MOISTURE EXCHANGER (HME)</b> .....	<b>37</b>
<b>201.102.5 BREATHING SYSTEM FILTERS (BSF)</b> .....	<b>37</b>
<b>201.103 * Spontaneous breathing during loss of power supply</b> .....	<b>37</b>
<b>201.104 * Training</b> .....	<b>38</b>
<b>201.105 * Indication of duration of operation</b> .....	<b>38</b>
<b>201.106 Functional connection</b> .....	<b>38</b>
<b>201.106.1 General</b> .....	<b>38</b>
<b>201.106.2 * Connection to an electronic health record</b> .....	<b>39</b>
<b>201.106.3 * Connection to a distributed alarm system</b> .....	<b>39</b>
<b>201.106.4 Connection for remote control</b> .....	<b>39</b>
<b>201.107 Display loops</b> .....	<b>39</b>
<b>201.107.1 Pressure-volume loops</b> .....	<b>39</b>
<b>201.107.2 Flow-volume loops</b> .....	<b>39</b>
<b>201.108 Power supply cords</b> .....	<b>40</b>
<b>201.109 Ventilatory support equipment security</b> .....	<b>40</b>
<b>202 Electromagnetic disturbances — Requirements and tests</b> .....	<b>40</b>
<b>206 Usability</b> .....	<b>41</b>
<b>208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems</b> .....	<b>43</b>
<b>211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment</b> .....	<b>44</b>
<b>Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS</b> .....	<b>45</b>
<b>Annex D (informative) Symbols on marking</b> .....	<b>52</b>
<b>Annex AA (informative) Particular guidance and rationale</b> .....	<b>54</b>
<b>Annex BB (informative) Data interface requirements</b> .....	<b>69</b>
<b>Annex CC (informative) Reference to the ESSENTIAL PRINCIPLES</b> .....	<b>76</b>

This is a preview of "ISO 80601-2-80:2018". [Click here to purchase the full version from the ANSI store.](#)

<b>Annex DD (informative) Terminology — Alphabetized index of defined terms.....</b>	<b>80</b>
<b>Bibliography .....</b>	<b>84</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This first edition of ISO 80601-2-80, in combination with ISO 80601-2-79<sup>[1]</sup>, cancels and replaces the second edition of ISO 10651-6:2004<sup>[2]</sup>. This edition of ISO 80601-2-80 constitutes a major technical revision of ISO 10651-6:2004 and includes an alignment with the third edition of IEC 60601-1, the fourth edition of IEC 60601-1-2, the third edition of IEC 60601-1-6, the second edition of IEC 60601-1-8 and the second edition of IEC 60601-1-11.

The most significant changes are the following modifications:

- splitting the scope of ISO 10651-6:2004<sup>[2]</sup> into two parts:
  - one for ventilatory impairment, also known as respiratory impairment (ISO 80601-2-79);
  - one for ventilatory insufficiency, also known as respiratory insufficiency (this document);
- extending the scope to include the VENTILATORY SUPPORT EQUIPMENT and its ACCESSORIES, where the characteristics of those ACCESSORIES can affect the BASIC SAFETY or ESSENTIAL PERFORMANCE of the VENTILATORY SUPPORT EQUIPMENT, and thus not only the VENTILATORY SUPPORT EQUIPMENT itself;

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

This is a preview of "ISO 80601-2-80:2018". [Click here to purchase the full version from the ANSI store.](#)

- identification of ESSENTIAL PERFORMANCE for VENTILATORY SUPPORT EQUIPMENT and its ACCESSORIES;
- and the following additions:
- tests for ventilation performance;
  - tests for mechanical strength (via IEC 60601-1-11);
  - requiring capable of TRANSIT-OPERABLE use;
  - new symbols;
  - requirements for VENTILATORY SUPPORT EQUIPMENT as a component of an ME SYSTEM;
  - tests for ENCLOSURE integrity (water ingress via IEC 60601-1-11);
  - tests for CLEANING and DISINFECTION PROCEDURES (via IEC 60601-1-11);
  - consideration of contamination of the breathing gas delivered to the PATIENT from the GAS PATHWAYS.

## Introduction

This document specifies requirements for VENTILATORY SUPPORT EQUIPMENT that is intended for use in the HOME HEALTHCARE ENVIRONMENT for PATIENTS who are not dependent for ventilation for their life support. VENTILATORY SUPPORT EQUIPMENT is frequently used in locations where SUPPLY MAINS is not reliable. VENTILATORY SUPPORT EQUIPMENT is often supervised by non-healthcare personnel (LAY OPERATORS) with varying levels of training. VENTILATORY SUPPORT EQUIPMENT complying with this document can be used elsewhere (i.e. in healthcare facilities).

Varying levels of ventilatory support are needed for PATIENTS who have stable ventilatory needs and in some cases, changing needs as their disease worsens. This document addresses PATIENTS who typically have severe enough respiratory function to prohibit certain activities that the PATIENT might normally pursue, and to interfere with daily living, occurring in association with measurements of respiratory mechanics or gas exchange that are markedly abnormal. This is best characterised by lung functions worse than<sup>[3]</sup>

- $FEV_1/FVC^2 < 70 \%$ , or
- $FEV_1 < 50 \%$  predicted

where

$FEV_1$  is the forced expiratory volume in 1 s, and

FVC is the forced vital capacity.

Examples of diseases that require ventilation support are severe Chronic Obstructive Pulmonary Disease (COPD), Amyotrophic Lateral Sclerosis (ALS)<sup>[4]</sup>, severe bronchopulmonary dysplasia and muscular dystrophy. VENTILATORY SUPPORT EQUIPMENT intended for this group of PATIENTS typically can require TECHNICAL ALARM CONDITIONS in the event that ESSENTIAL PERFORMANCE is absent. The most fragile of these PATIENTS would likely experience injury, but not serious injury or death, with the loss of this artificial ventilation. For these PATIENTS, it is likely that ventilatory support is needed during waking hours while PATIENTS are moving inside or outside the home in order to facilitate mobility and functional independence in the activities of daily living.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD<sup>3</sup>, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the five numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.);

---

<sup>2</sup> This is also known as the Tiffeneau-Pinelli index.

<sup>3</sup> The general standard is IEC 60601-1:2005+AMD1:2012, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*.



This is a preview of "ISO 80601-2-80:2018". [Click here to purchase the full version from the ANSI store.](#)

- "subclause" means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe permission (e.g. a permissible way to achieve compliance with a requirement or test);
- "can" is used to describe a possibility or capability;
- "must" is used express an external constraint.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

The ISO and IEC 80601 family of documents are also parts of the IEC 60601 family of documents.