

STANDARD

8615

First edition
1991-11-01

**Implants for surgery — Fixation devices for use
in the ends of the femur in adults**

*Implants chirurgicaux — Implants de fixation à utiliser dans les
extrémités du fémur chez les adultes*



Reference number
ISO 8615:1991(E)

This is a preview of ISO 8615:1991. [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
1 Scope	1
2 Normative references	1
3 Materials	2
4 Types of nails and fixation screws	2
5 Dimensions	2
6 Ancillary components for assembly of fixation devices of multi-piece construction	4
7 Surface finish	4
8 Packaging	4
9 Marking	4

© ISO 1991

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

This is a preview of ISO 8615:1991. [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

International Standard ISO 8615 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Sub-Committee SC 5, *Osteosynthesis*.

Introduction

Fixation devices for use in the upper and lower ends of the femur may be made in a variety of configurations, according to the design of the originator or the individual manufacturer. However, certain dimensions are critical and may not be exceeded without risk of damage to the bone when the device is inserted. Other than specifying this limited number of features, the main object of this International Standard is to ensure uniformity in designating and describing the size of the device in order that devices of different origin but of the same nominal size will not differ significantly in their critical dimensions. It is required that these dimensions be stated on the packaging and marked on the device; it is also required that the packaging carry other information useful to the surgeon, such as details of instruments and ancillary devices necessary at the implantation of the device.

This is a preview of ISO 8615:1991. [Click here to purchase the full version from the ANSI store.](#)

Implants for surgery — Fixation devices for use in the ends of the femur in adults

1 Scope

This International Standard specifies the materials, dimensions, surface finish, and packaging and marking of metal surgical implants used for the fixation of fractures and osteotomies of the upper and lower ends of the femur in adults. The types of devices covered are as follows:

- a) nails and screws intended to be used alone;
- b) nail plates and blade plates of one-piece construction;
- c) fixation devices of multi-piece construction having either a fixed or an adjustable angle;
- d) fixation devices of multi-piece construction with sliding nail or screw, with or without a compression device.

NOTE 1 Figures 1 to 7 are intended to illustrate nomenclature and designation of dimensions, but the representation of the components does not otherwise form part of the requirements specified in this International Standard.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 261:1973, *ISO general-purpose metric screw threads — General plan.*

ISO 272:1982, *Fasteners — Hexagon products — Widths across flats.*

ISO 965-2:1980, *ISO general purpose metric screw threads — Tolerances — Part 2: Limits of sizes for general purpose bolt and nut threads — Medium quality.*

ISO 5832-1:1987, *Implants for surgery — Metallic materials — Part 1: Wrought stainless steel.*

ISO 5832-2:1978, *Implants for surgery — Metallic materials — Part 2: Unalloyed titanium.*

ISO 5832-3:1990, *Implants for surgery — Metallic materials — Part 3: Wrought titanium 6-aluminium 4-vanadium alloy.*

ISO 5832-4:1978, *Implants for surgery — Metallic materials — Part 4: Cobalt-chromium-molybdenum casting alloy.*

ISO 5832-5:1978, *Implants for surgery — Metallic materials — Part 5: Wrought cobalt-chromium-tungsten-nickel alloy.*

ISO 5832-6:1980, *Implants for surgery — Metallic materials — Part 6: Wrought cobalt-nickel-chromium-molybdenum alloy.*

ISO 5832-7:1984, *Implants for surgery — Metallic materials — Part 7: Forgeable and cold-formed cobalt-chromium-nickel-molybdenum-iron alloy.*

ISO 5832-8:1987, *Implants for surgery — Metallic materials — Part 8: Wrought cobalt-nickel-chromium-molybdenum-tungsten-iron alloy.*

ISO 5835:1991, *Implants for surgery — Metal bone screws with hexagonal drive connection, spherical under-surface of head, asymmetrical thread — Dimensions.*

ISO 5836:1988, *Implants for surgery — Metal bone plates — Holes corresponding to screws with asymmetrical thread and spherical under-surface.*