

First edition
2022-04

Piston-operated volumetric apparatus —

Part 8: Photometric reference measurement procedure for the determination of volume

Appareils volumétriques à piston —

*Partie 8: Mode opératoire de mesure photométrique de référence
pour la détermination de volumes*



Reference number
ISO 8655-8:2022(E)

© ISO 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 8655-8:2022". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	2
5 Test equipment	2
5.1 General.....	2
5.2 Spectrophotometer.....	2
5.3 Cuvette and mixer.....	2
5.4 Measuring devices.....	3
5.5 Equipment used for solution preparation.....	3
5.6 Balances.....	3
5.7 Density meter.....	4
5.8 pH meter.....	4
6 Reagents	4
6.1 General requirements.....	4
6.2 Water.....	4
6.3 Buffer solution.....	4
6.4 Copper(II) chloride solution.....	4
6.5 Ponceau S solutions.....	4
6.6 Calibrator solutions.....	5
6.7 Stability of solutions.....	6
6.7.1 General.....	6
6.7.2 Preservatives.....	6
6.7.3 Light sensitivity.....	6
6.7.4 Storage temperature.....	6
7 Test conditions	6
7.1 General.....	6
7.2 Test room.....	6
7.3 Evaporation.....	7
8 Procedure	7
8.1 General.....	7
8.1.1 Summary.....	7
8.1.2 Test conditions.....	7
8.1.3 Test volume.....	7
8.1.4 Number of measurements per volume to be tested.....	8
8.2 System calibration.....	8
8.2.1 General.....	8
8.2.2 System calibration procedure.....	8
8.2.3 Previous calibration.....	8
8.3 Photometric procedure.....	8
8.3.1 Preparation of cuvettes.....	8
8.3.2 Zero of the spectrophotometer.....	9
8.3.3 Starting absorbances.....	9
8.3.4 Dispensing of test liquids.....	9
8.3.5 Absorbance of the chromophore mixture.....	9
8.3.6 Calculation of the delivered test volume.....	9
8.4 Preparation.....	9
8.5 Single-channel air displacement pipettes (in accordance with ISO 8655-2).....	10
8.5.1 General.....	10

This is a preview of "ISO 8655-8:2022". [Click here to purchase the full version from the ANSI store.](#)

8.5.2	Test cycle	10
8.6	Multi-channel pipettes (in accordance with ISO 8655-2)	11
8.7	Positive displacement pipettes (in accordance with ISO 8655-2)	11
8.8	Burettes (in accordance with ISO 8655-3)	11
8.9	Dispensers (in accordance with ISO 8655-5)	12
8.10	Syringes (in accordance with ISO 8655-9)	12
8.10.1	General	12
8.10.2	Test cycle	12
9	Evaluation	13
9.1	Calculation of volume	13
9.1.1	Calibration constant	13
9.1.2	Volume of test liquid	13
9.1.3	Temperature correction	14
9.1.4	Mean volume	14
9.2	Systematic error of measurement	14
9.3	Random error of measurement	15
9.4	Uncertainty of measurement	15
10	Reporting of results	15
Annex A (normative) Calculation of volumes from balance readings		17
Bibliography		18

This is a preview of "ISO 8655-8:2022". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 332, *Laboratory equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 8655 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The ISO 8655 series addresses the needs of:

- manufacturers, as a basis for quality control including, where appropriate, the issuance of manufacturer's declarations;
- calibration laboratories, test houses, users of the equipment and other bodies as a basis for independent calibration, testing, verification, and routine tests.

The tests specified in the ISO 8655 series are intended to be carried out by trained personnel.