



**Elastomeric parts for parenterals and
for devices for pharmaceutical use —**

**Part 5:
Functional requirements and testing**

*Éléments en élastomère pour administration parentérale et
dispositifs à usage pharmaceutique —*

Partie 5: Exigences fonctionnelles et essais

ISO 8871-5

**Third edition
2025-03**



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of ISO 8871-5:2025. [Click here to purchase the full version from the ANSI store.](#)

| | |
|--|-----------|
| Foreword | iv |
| Introduction | vi |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Requirements | 2 |
| 4.1 Penetrability..... | 2 |
| 4.2 Fragmentation..... | 2 |
| 4.3 Self-sealing..... | 2 |
| 4.4 Dye solution tightness..... | 2 |
| 5 Preparation of closures for testing | 2 |
| 5.1 Sampling..... | 2 |
| 5.2 Cleaning..... | 3 |
| 5.3 Sterilization..... | 3 |
| Annex A (normative) Test for penetrability | 4 |
| Annex B (normative) Test for fragmentation | 5 |
| Annex C (normative) Test for self-sealing | 7 |
| Annex D (normative) Test for dye solution tightness | 9 |
| Bibliography | 11 |

This is a preview of ISO 8871-5:2025. [Click here to purchase the full version from the ANSI store.](#)

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 8871-5:2016), which has been technically revised.

The main changes are as follows:

- "aqueous solution tightness" has been replaced with "dye solution tightness";
- definition for fragmentation has been clarified;
- definitions for "filling volume" and "nominal volume" have been added;
- information on the fragmentation test and new information on the fragment size have been added;
- "pore size of 0,5 µm" has been replaced with "pore size of maximum 5,0 µm" to be aligned with ISO 11608-3;
- clarification on the fragments counting measurements has been added;
- "dye solution tightness" from the self-sealing test has been removed;
- "solution of methylene blue" has been replaced by "appropriate dye solution" and new information on the dye solution;
- "methylene blue" has been replaced by "dye solution";
- references to USP <788> and USP <1207> have been added.

A list of all parts in the ISO 8871 series can be found on the ISO website.

This is a preview of ISO 8871-5:2025. [Click here to purchase the full version from the ANSI store.](#)

This is a preview of ISO 8871-5:2025. [Click here to purchase the full version from the ANSI store.](#)

The elastomeric parts specified in the ISO 8871 series are produced from rubber and thermoplastic elastomers (TPE). These pharmaceutical use closures are used in combination with vials and many times in conjunction with piercing devices. There are three functional parameters which are important to the piercing process. These are penetrability, fragmentation and self-sealing. The three functional tests described in this document can be used as a reference method for testing closures that are pierced using injection needles made from metal. In addition, the dye solution tightness test can be used to verify the effectiveness of the sealing of a specific closure/vial combination.