

This is a preview of "ISO 9394:2012". [Click here to purchase the full version from the ANSI store.](#)

Third edition  
2012-10-01

---

---

## **Ophthalmic optics — Contact lenses and contact lens care products — Determination of biocompatibility by ocular study with rabbit eyes**

*Optique ophtalmique — Lentilles de contact et produits d'entretien  
pour lentilles de contact — Détermination de la biocompatibilité par  
évaluation de la tolérance oculaire chez le lapin*



Reference number  
ISO 9394:2012(E)

© ISO 2012

This is a preview of "ISO 9394:2012". [Click here to purchase the full version from the ANSI store.](#)



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

This is a preview of "ISO 9394:2012". [Click here to purchase the full version from the ANSI store.](#)

## Contents

Page

Foreword .....	iv
Introduction .....	v
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions .....	1
4 General requirements .....	1
5 Animals and husbandry .....	1
6 Reagents/materials .....	2
7 Apparatus .....	2
8 Test specimens .....	3
8.1 Lens parameters .....	3
8.2 Preparation and storage .....	3
9 Test procedure .....	3
9.1 Preliminary examination of animals .....	3
9.2 Insertion and removal of test lens .....	4
9.3 Examination of the rabbit's eye .....	4
9.4 Weighing of animals .....	4
9.5 Histological examination .....	5
9.6 Corneal metabolism .....	5
10 Assessment of results .....	5
11 Test report .....	5
Annex A (normative) McDonald-Shadduck score system — Slit lamp .....	7
Annex B (normative) Draize scale for scoring ocular lesions .....	11
Bibliography .....	13

This is a preview of "ISO 9394:2012". [Click here to purchase the full version from the ANSI store.](#)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9394 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This third edition cancels and replaces the second edition (ISO 9394:1998), which has been technically revised.

This is a preview of "ISO 9394:2012". [Click here to purchase the full version from the ANSI store.](#)

## Introduction

The ocular tissue of the rabbit is traditionally used to evaluate the irritant properties of materials which come in contact with ocular tissue.

The use of the device under evaluation is governed by the nature, degree, duration, frequency and conditions of exposure of humans to the device in normal intended use.

It is incumbent upon the investigator to conduct such evaluations using good scientific laboratory practices, complying with regulations related to animal welfare and the general principles set forth in the normative references.

ISO 10993-1 is the basic horizontal International Standard for biological evaluation of medical devices, and serves as a framework for planning biological evaluation tests.

ISO 10993-10 assesses possible contact hazards from device-released chemicals that may produce skin and mucosal irritation, eye irritation and delayed contact sensitization.

Usage tests for specific devices are defined in vertical standards. This International Standard describes one of several specific usage tests for contact lenses and contact lens care products.

The existence of this International Standard does not imply that rabbit-eye testing is a requirement in the determination of biocompatibility of contact lenses and contact lens care products, nor that this test is sufficient by itself to determine the biocompatibility of contact lenses and contact lens care products. Taking into consideration animal welfare requirements (ISO 10993-2), it is recommended that this *in vivo* test be carried out after obtaining data of *in vitro* toxicological testing such as that described in ISO 10993-5.

Care should be taken when extrapolating the test results to the human eye.