

This is a preview of "ISO 9847:2023". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2023-01

Solar energy — Calibration of pyranometers by comparison to a reference pyranometer

*Énergie solaire — Étalonnage des pyranomètres par comparaison à
un pyranomètre de référence*



Reference number
ISO 9847:2023(E)

© ISO 2023

This is a preview of "ISO 9847:2023". Click here to purchase the full version from the ANSI store.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO 9847:2023". [Click here to purchase the full version from the ANSI store.](#)

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Pyranometer calibration	4
4.1 General.....	4
4.2 Pyranometer sensitivity, measurement equation, measurand.....	4
4.3 Indoor and outdoor calibration compared.....	6
4.4 Method validation.....	6
4.5 Calibration uncertainty.....	6
5 Measuring equipment	7
5.1 Data acquisition and recording.....	7
5.2 Instrument platforms.....	8
5.3 Pyranometers.....	8
6 Indoor calibration (Type A)	8
6.1 Introductory remarks on indoor calibration.....	8
6.2 Reference pyranometers for indoor calibration.....	8
6.3 Indoor calibration systems.....	9
6.3.1 System with a direct beam source (type A1).....	9
6.3.2 Systems with an integrating sphere source (type A2).....	9
6.4 Indoor calibration procedures.....	9
6.4.1 Calibration procedure requirements (types A1 and A2).....	9
6.4.2 Indoor calibration procedures (types A1 and A2).....	9
6.4.3 Calculation of the sensitivity.....	10
6.4.4 Calibration conditions and optional correction of reference operating conditions.....	11
6.4.5 Uncertainty evaluation.....	11
7 Outdoor calibration (Type B)	12
7.1 Introductory remarks on outdoor calibration.....	12
7.2 Reference pyranometers for outdoor calibration.....	12
7.3 Outdoor calibration systems.....	12
7.3.1 Site selection for outdoor calibration.....	12
7.3.2 Tracking for normal incidence calibration (type B3).....	13
7.4 Outdoor calibration procedures.....	13
7.4.1 Calibration procedure requirements (B1, B2, B3).....	13
7.4.2 Outdoor horizontal calibration procedure (type B1).....	13
7.4.3 Outdoor tilted calibration procedure (type B2).....	14
7.4.4 Outdoor normal incidence calibration procedure (type B3).....	15
7.4.5 Calculation of the sensitivity.....	15
7.4.6 Calibration conditions and optional correction of reference operating conditions.....	16
7.4.7 Uncertainty evaluation.....	16
8 Calibration certificate	17
Annex A (informative) Examples of calibration systems using artificial sources	18
Annex B (informative) Calculation of daily average zenith angle	22
Annex C (informative) Introduction of a new pyranometer sensitivity	24
Annex D (informative) Data quality review for outdoor calibration	26
Annex E (informative) Uncertainty evaluation for outdoor calibration	29

This is a preview of "ISO 9847:2023". [Click here to purchase the full version from the ANSI store.](#)

Annex F (informative) Uncertainty evaluation for indoor calibration	30
Bibliography	31

This is a preview of "ISO 9847:2023". Click here to purchase the full version from the ANSI store.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 180, *Solar energy*, Subcommittee SC 1, *Climate – Measurement and data*.

This second edition cancels and replaces the first edition (ISO 9847:1992) which has been technically revised.

The main changes are as follows:

- focus on current calibration practices;
- adapted recommendations for mathematical treatment of data;
- adaptation of the terminology to the revised ISO 9060:2018 and ISO Guide 99^[1];
- added comments on uncertainty evaluation of the calibration with reference to ASTM G213^[2] and ISO/IEC Guide 98-3;
- inclusion of reference to non-spectrally-flat pyranometers, that are now also included in ISO 9060.

[Annexes A, B, C, D, E](#) and [F](#) are given for information only.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Pyranometers are instruments used to measure the irradiance (power per unit area) received from the sun for many purposes.

In recent years the application of hemispherical solar radiation measurement, using pyranometers, has risen sharply. The main application of pyranometers now is no longer scientific research, but assessment of the performance of solar power plants.

Accurate measurements of the hemispherical solar radiation are required for

- a) the determination of the energy input to solar energy systems such as photovoltaic (PV) -, and solar thermal systems, as a basis for performance assessment,
- b) the testing and assessment of solar technologies,
- c) the geographic mapping of solar energy resources, and
- d) other applications such as agriculture, building efficiency, material degradation and reliability, climate, weather, health, etc.

Today's growing solar energy performance assessment markets demand the lowest possible measurement uncertainties. To meet this demand, a measurement requires an uncertainty evaluation and an accurate time stamp^[3].

Calibration of measuring instruments is an essential part of the uncertainty evaluation and part of any quality management system. Regular instrument re-calibration according to this standard helps attaining the required low measurement uncertainties. Calibration usually will show the instrument is stable and then serves as:

- confirmation that the measurement data collected over the time interval from the previous to the present calibration are reliable
- the instrument is expected to remain stable, future measurement data are expected to be reliable.

Uncertainties mentioned in this document are expanded uncertainties with a coverage factor $k = 2$.