

**INTERNATIONAL
STANDARD**

**ISO/IEEE
11073-10201**

First edition
2004-12-15

**Health informatics — Point-of-care
medical device communication —
Part 10201:
Domain information model**

*Informatique de santé — Communication entre dispositifs médicaux sur le
site des soins —
Partie 10201: Modèle d'information du domaine*



Reference number
ISO/IEEE 11073-10201:2004(E)

© ISO/IEEE 2004

ISO/IEEE 11073-10201:2004(E)

**Health informatics — Point-of-care
medical device communication —
Part 10201:
Domain information model**

Sponsor

IEEE 1073™ Standard Committee

of the

IEEE Engineering in Medicine and Biology Society

Approved 24 June 2004

IEEE-SA Standards Board



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor the IEEE accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IEEE members. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat or the IEEE at the address given below.

Abstract: Within the context of the ISO/IEEE 11073 family of standards for point-of-care (POC) medical device communication (MDC), this standard provides an abstract object-oriented domain information model that specifies the structure of exchanged information, as well as the events and services that are supported by each object. All elements are specified using abstract syntax (ASN.1) and may be applied to many different implementation technologies, transfer syntaxes, and application service models. Core subjects include medical, alert, system, patient, control, archival, communication, and extended services. Model extensibility is supported, and a conformance model and statement template is provided.

Keywords: abstract syntax, alarm, alert, ASN.1, information model, medical device communications, medical information bus, MIB, point-of-care, POC, object-oriented, patient, remote control

This ISO/IEEE document is an International Standard and is copyright-protected by ISO and the IEEE. Except as permitted under the applicable laws of the user's country, neither this ISO/IEEE standard nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO or the IEEE at the addresses below.

ISO copyright office
Case postale 56 · CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Institute of Electrical and Electronics Engineers
Standards Association
Manager, Standards Intellectual Property
445 Hoes Lane
Piscataway, NJ 08854
E-mail: stds.ipr@ieee.org
Web: www.ieee.org

Copyright © 2004 ISO/IEEE. All rights reserved.
Published 15 December 2004. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by the Institute of Electrical and Electronics Engineers, Incorporated.

Print: ISBN 0-7381-4089-9 SH95256
PDF: ISBN 0-7381-4090-2 SS95256

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Use of an IEEE Standard is wholly voluntary. The IEEE disclaims liability for any personal injury, property or other damage, of any nature whatsoever, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, or reliance upon this, or any other IEEE Standard document.

The IEEE does not warrant or represent the accuracy or content of the material contained herein, and expressly disclaims any express or implied warranty, including any implied warranty of merchantability or fitness for a specific purpose, or that the use of the material contained herein is free from patent infringement. IEEE Standards documents are supplied **“AS IS.”**

The existence of an IEEE Standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE Standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard. Every IEEE Standard is subjected to review at least every five years for revision or reaffirmation. When a document is more than five years old and has not been reaffirmed, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE Standard.

In publishing and making this document available, the IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity. Nor is the IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing this, and any other IEEE Standards document, should rely upon the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

Interpretations: Occasionally questions may arise regarding the meaning of portions of standards as they relate to specific applications. When the need for interpretations is brought to the attention of IEEE, the Institute will initiate action to prepare appropriate responses. Since IEEE Standards represent a consensus of concerned interests, it is important to ensure that any interpretation has also received the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to interpretation requests except in those cases where the matter has previously received formal consideration. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position, explanation, or interpretation of the IEEE.

Comments for revision of IEEE Standards are welcome from any interested party, regardless of membership affiliation with IEEE. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Comments on standards and requests for interpretations should be addressed to:

Secretary, IEEE-SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

NOTE — Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken with respect to the existence or validity of any patent rights in connection therewith. The IEEE shall not be responsible for identifying patents for which a license may be required by an IEEE standard or for conducting inquiries into the legal validity or scope of those patents that are brought to its attention.

Authorization to photocopy portions of any individual standard for internal or personal use is granted by the Institute of Electrical and Electronics Engineers, Inc., provided that the appropriate fee is paid to Copyright Clearance Center. To arrange for payment of licensing fee, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

ISO Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

A pilot project between ISO and the IEEE has been formed to develop and maintain a group of ISO/IEEE standards in the field of medical devices as approved by Council resolution 43/2000. Under this pilot project, IEEE is responsible for the development and maintenance of these standards with participation and input from ISO member bodies.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. Neither ISO nor the IEEE shall be held responsible for identifying any or all such patent rights.

ISO/IEEE 11073-10201:2004(E) was prepared by IEEE 1073 Committee of the IEEE Engineering in Medicine and Biology Society.

IEEE Introduction

This introduction is not part of ISO/IEEE 11073-10201:2004(E), Health informatics — Point-of-care medical device communication — Part 10201: Domain information model.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. They provide automatic and detailed electronic data capture of patient vital signs information and device operational data. The primary goals are to:

- Provide real-time plug-and-play interoperability for patient-connected medical devices
- Facilitate the efficient exchange of vital signs and medical device data, acquired at the point-of-care, in all health care environments

“Real-time” means that data from multiple devices can be retrieved, time correlated, and displayed or processed in fractions of a second. “Plug-and-play” means that all the clinician has to do is make the connection — the systems automatically detect, configure, and communicate without any other human interaction.

“Efficient exchange of medical device data” means that information that is captured at the point-of-care (e.g., patient vital signs data) can be archived, retrieved, and processed by many different types of applications without extensive software and equipment support, and without needless loss of information. The standards are especially targeted at acute and continuing care devices, such as patient monitors, ventilators, infusion pumps, ECG devices, etc. They comprise a family of standards that can be layered together to provide connectivity optimized for the specific devices being interfaced.

Notice to users

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken with respect to the existence or validity of any patent rights in connection therewith. The IEEE shall not be responsible for identifying patents or patent applications for which a license may be required by to implement an IEEE standard or for conducting inquiries into the legal validity or scope of those patents that are brought to its attention.

Errata

Errata, if any, for this and all other standards can be accessed at the following URL: <http://standards.ieee.org/reading/ieee/updates/errata/index.html>. Users are encouraged to check this URL for errata periodically.

Interpretations

Current interpretations can be accessed at the following URL: <http://standards.ieee.org/reading/ieee/interp/index.html>.

Participants

At the time this standard was completed, the working group of the IEEE 1073 Standard Committee had the following membership:

Todd H. Cooper, *Chair*

Wolfgang Bleicher
Francis Cantraine
Thomas Canup
Mats Cardell
Michael Chilbert
Michael Flötotto
Ken Fuchs
Kai Hassing
Gunther Hellmann

Jörg Kampmann
Ron Kirkham
Michael Krämer
Alberto Macerata
Simon Meij
Angelo Rossi Mori
Thomas Norgall
Daniel Nowicki
Thomas Penzel
Francesco Pincioli

Melvin Reynolds
Paul Rubel
Lief Rystrom
Paul Schluter
Michael Spicer
Alpo Värri
Jan Wittenber
Paul Woolman
Christoph Zywietz

The following members of the individual balloting committee voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Thomas Canup
Michael Chilbert
Keith Chow
Todd H. Cooper
Grace Esche
Kenneth Fuchs

John Grider
Kai Hassing
Tom Kannally
Robert Kennelly
Randall Krohn
Yeou-Song Lee
Daniel Nowicki

Melvin Reynolds
Michael Spicer
Richard Schrenker
M. Michael Shabot
Lars Steubesand
Gin-shu Young

When the IEEE-SA Standards Board approved this standard on 24 June 2004, it had the following membership:

Don Wright, *Chair*

Steve M. Mills, *Vice Chair*

Judith Gorman, *Secretary*

Chuck Adams
H. Stephen Berger
Mark D. Bowman
Joseph A. Bruder
Bob Davis
Roberto de Marca Boisson
Julian Forster*
Arnold M. Greenspan

Mark S. Halpin
Raymond Hapeman
Richard J. Holleman
Richard H. Hulett
Lowell G. Johnson
Joseph L. Koepfinger*
Hermann Koch
Thomas J. McGean
Daleep C. Mohla

Paul Nikolic
T. W. Olsen
Ronald C. Petersen
Gary S. Robinson
Frank Stone
Malcolm V. Thaden
Doug Topping
Joe D. Watson

*Member Emeritus

Also included are the following nonvoting IEEE-SA Standards Board liaisons:

Satish K. Aggarwal, *NRC Representative*
Richard DeBlasio, *DOE Representative*
Alan Cookson, *NIST Representative*

Don Messina
IEEE Standards Project Editor

Contents

1.	Scope.....	1
2.	Normative references	1
3.	Definitions	4
4.	Abbreviations and acronyms	8
5.	General requirements	9
6.	DIM.....	9
6.1	General.....	9
6.1.1	Modeling concept	9
6.1.2	Scope of the DIM.....	11
6.1.3	Approach.....	11
6.1.4	Extension of the model	12
6.2	Package diagram—overview	12
6.3	Model for the Medical Package	12
6.3.1	VMO (i.e., virtual medical object).....	14
6.3.2	VMD (i.e., virtual medical device) object	14
6.3.3	Channel object	15
6.3.4	Metric object	15
6.3.5	Numeric object.....	15
6.3.6	Sample Array object	15
6.3.7	Real Time Sample Array object.....	15
6.3.8	Time Sample Array object.....	15
6.3.9	Distribution Sample Array object	16
6.3.10	Enumeration object	16
6.3.11	Complex Metric object	16
6.3.12	PM-Store (i.e., persistent metric) object	16
6.3.13	PM-Segment object.....	16
6.4	Model for the Alert Package	17
6.4.1	Alert object	17
6.4.2	Alert Status object.....	18
6.4.3	Alert Monitor object	18
6.5	Model for the System Package	19
6.5.1	VMS (i.e., virtual medical system) object	19
6.5.2	MDS object	19
6.5.3	Simple MDS object.....	20
6.5.4	Hydra MDS object	20
6.5.5	Composite Single Bed MDS object.....	20
6.5.6	Composite Multiple Bed MDS object	20
6.5.7	Log object	20
6.5.8	Event Log object.....	20
6.5.9	Battery object.....	20
6.5.10	Clock object	20
6.6	Model for the Control Package	21
6.6.1	SCO.....	22
6.6.2	Operation object.....	22
6.6.3	Select Item Operation object.....	22

6.6.4	Set Value Operation object	23
6.6.5	Set String Operation object.....	23
6.6.6	Toggle Flag Operation object	23
6.6.7	Activate Operation object	23
6.6.8	Limit Alert Operation object.....	23
6.6.9	Set Range Operation object	23
6.7	Model for the Extended Services Package	24
6.7.1	Scanner object.....	24
6.7.2	CfgScanner (i.e., configurable scanner) object.....	25
6.7.3	EpiCfgScanner (i.e., episodic configurable scanner) object.....	25
6.7.4	PeriCfgScanner (i.e., periodic configurable scanner) object	25
6.7.5	FastPeriCfgScanner (i.e., fast periodic configurable scanner) object.....	26
6.7.6	UcfgScanner (i.e., unconfigurable scanner) object.....	26
6.7.7	Context Scanner object	26
6.7.8	Alert Scanner object.....	26
6.7.9	Operating Scanner object.....	26
6.8	Model for the Communication Package	26
6.8.1	Communication Controller object	27
6.8.2	DCC (i.e., device communication controller) object	27
6.8.3	BCC (i.e., bedside communication controller) object	28
6.8.4	Device Interface object	28
6.8.5	MibElement object.....	28
6.8.6	Specialized MibElement object	28
6.9	Model for the Archival Package	28
6.9.1	Multipatient Archive object.....	29
6.9.2	Patient Archive object.....	29
6.9.3	Session Archive object.....	30
6.9.4	Physician object	30
6.9.5	Session Test object.....	30
6.9.6	Session Notes object	30
6.9.7	Ancillary object.....	30
6.10	Model for the Patient Package	30
6.10.1	Patient Demographics object	31
6.11	DIM—dynamic model.....	31
6.11.1	General.....	31
6.11.2	MDS communication finite state machine (FSM).....	31
6.11.3	Communicating systems—startup object interaction diagram	33
6.11.4	Communication Package—MibElement data access	34
6.11.5	Dynamic object relations	34
7.	DIM object definitions	36
7.1	General.....	36
7.1.1	Notation	36
7.1.2	Common data types	37
7.2	Top object	45
7.2.1	Attributes	45
7.2.2	Behavior.....	45
7.2.3	Notifications.....	45
7.3	Objects in the Medical Package.....	46
7.3.1	VMO	46
7.3.2	VMD object	47
7.3.3	Channel object	49
7.3.4	Metric object.....	50

	7.3.5	Numeric object.....	56
	7.3.6	Sample Array object	58
	7.3.7	Real Time Sample Array object.....	63
	7.3.8	Time Sample Array object.....	65
	7.3.9	Distribution Sample Array object	67
	7.3.10	Enumeration object	69
	7.3.11	Complex Metric object	72
	7.3.12	PM-Store object	75
	7.3.13	PM-Segment object.....	78
7.4		Objects in the Alert Package.....	80
	7.4.1	Alert object	80
	7.4.2	Alert Status object.....	82
	7.4.3	Alert Monitor object	84
7.5		Objects in the System Package	86
	7.5.1	VMS object	86
	7.5.2	MDS object	89
	7.5.3	Simple MDS object.....	93
	7.5.4	Hydra MDS object.....	93
	7.5.5	Composite Single Bed MDS object	93
	7.5.6	Composite Multiple Bed MDS object	93
	7.5.7	Log object	93
	7.5.8	Event Log object.....	95
	7.5.9	Battery object.....	97
	7.5.10	Clock object	99
7.6		Objects in the Control Package.....	105
	7.6.1	SCO.....	105
	7.6.2	Operation object.....	108
	7.6.3	Select Item Operation object.....	110
	7.6.4	Set Value Operation object	112
	7.6.5	Set String Operation object.....	113
	7.6.6	Toggle Flag Operation object	114
	7.6.7	Activate Operation object	116
	7.6.8	Limit Alert Operation object.....	116
	7.6.9	Set Range Operation object	118
7.7		Objects in the Extended Services Package	120
	7.7.1	Scanner object.....	120
	7.7.2	CfgScanner object.....	121
	7.7.3	EpiCfgScanner object	123
	7.7.4	PeriCfgScanner object	124
	7.7.5	FastPeriCfgScanner object.....	125
	7.7.6	UcfgScanner object.....	126
	7.7.7	Context Scanner object	127
	7.7.8	Alert Scanner object.....	129
	7.7.9	Operating Scanner object.....	130
7.8		Objects in the Communication Package	132
	7.8.1	Communication Controller object	132
	7.8.2	DCC object	135
	7.8.3	BCC object.....	135
	7.8.4	Device Interface object	135
	7.8.5	MibElement object.....	136
	7.8.6	Device Interface MibElement object	137
	7.8.7	General Communication Statistics MibElement object.....	138
7.9		Objects in the Archival Package.....	140
	7.9.1	Multipatient Archive object.....	140

7.9.2	Patient Archive object.....	141
7.9.3	Session Archive object.....	142
7.9.4	Physician object	143
7.9.5	Session Test object.....	145
7.9.6	Session Notes object	146
7.10	Objects in the Patient Package.....	147
7.10.1	Patient Demographics object	147
8.	Service model for communicating systems	151
8.1	General.....	151
8.2	Communicating systems	151
8.3	General service model overview.....	152
8.3.1	Conceptual architecture of communicating systems	153
8.4	General object management services definition	154
8.4.1	EVENT REPORT service.....	155
8.4.2	GET service	156
8.4.3	SET service	156
8.4.4	ACTION service	157
8.4.5	CREATE service.....	158
8.4.6	DELETE service	159
9.	MDIB nomenclature	160
10.	Conformance model.....	161
10.1	Applicability	161
10.2	Conformance specification	161
10.3	ICSs.....	162
10.3.1	General format	162
10.3.2	General ICS.....	162
10.3.3	Service Support ICS.....	164
10.3.4	DIM managed object class (MOC) ICS.....	165
10.3.5	MOC Attribute ICS.....	165
10.3.6	MOC Behavior ICS	166
10.3.7	MOC Notification ICS.....	166