

First edition  
2019-03

---

---

## Health informatics — Personal health device communication —

Part 10207:

### Domain information and service model for service-oriented point-of- care medical device communication

*Informatique de santé — Communication entre dispositifs de santé  
personnels —*

*Partie 10207: Informations de domaine et modèle de services pour la  
communication orientée services entre dispositifs médicaux sur le site  
des soins*



Reference number  
ISO/IEEE 11073-10207:2019(E)

© IEEE 2018



**COPYRIGHT PROTECTED DOCUMENT**

© IEEE 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from IEEE at the address below.

Institute of Electrical and Electronics Engineers, Inc  
3 Park Avenue, New York  
NY 10016-5997, USA

Email: [stds.ipr@ieee.org](mailto:stds.ipr@ieee.org)  
Website: [www.ieee.org](http://www.ieee.org)

Published in Switzerland

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see [www.iso.org/directives](http://www.iso.org/directives)).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

ISO/IEEE 11073-10207 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-10207-2017) and drafted in accordance with its editorial rules. It was adopted, under the "fast-track procedure" defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

A list of all parts in the ISO 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

**Abstract:** Within the context of the ISO/IEEE 11073™ family of standards for point-of-care medical device communication, a Participant Model derived from the ISO/IEEE11073-10201 Domain Information Model is provided in this standard. The Participant Model specifies the structure of medical information objects. This standard also defines an abstract Communication Model to support the exchange of medical information objects. All elements of the Participant Model and Communication Model are specified using XML Schema. Core subjects of the Participant Model comprise modelling of medical device-related data, e.g., measurements and settings, alert systems, contextual information (e.g., patient demographics and location information), remote control, and archival information. Model extensibility is provided inherently through the use of XML Schema.

**Keywords:** alert systems, IEEE 11073-10207™, medical device communication, patient, point-of-care, remote control, service-oriented architecture, XML Schema

---

The Institute of Electrical and Electronics Engineers, Inc.  
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2018 by The Institute of Electrical and Electronics Engineers, Inc.  
All rights reserved. Published 21 February 2018. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-4627-3      STD22961  
Print: ISBN 978-1-5044-4628-0      STDPD22961

*IEEE prohibits discrimination, harassment, and bullying.*

*For more information, visit <http://www.ieee.org/web/aboutus/whatis/policies/p9-26.html>.*

*No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.*

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

## Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading "Important Notices and Disclaimers Concerning IEEE Standards Documents." They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/IPR/disclaimers.html>.

## Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association ("IEEE-SA") Standards Board. IEEE ("the Institute") develops its standards through a consensus development process, approved by the American National Standards Institute ("ANSI"), which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; noninfringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied "AS IS" and "WITH ALL FAULTS."

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

## Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

This is a preview of "ISO/IEEE 11073-10207...". [Click here to purchase the full version from the ANSI store.](#)

## Official statements

A statement, written or oral, that is not processed in accordance with the IEEE-SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

## Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE-SA Standards Board  
445 Hoes Lane  
Piscataway, NJ 08854 USA

## Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so

## Copyrights

IEEE draft and approved standards are copyrighted by IEEE under U.S. and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

## Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, nonexclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, noncommercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

## Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every ten years. When a document is more than ten years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit the IEEE-SA Website at <http://ieeexplore.ieee.org/xpl/standards.jsp> or contact IEEE at the address listed previously. For more information about the IEEE-SA or IEEE's standards development process, visit the IEEE-SA Website at <http://standards.ieee.org>.

## Errata

Errata, if any, for all IEEE standards can be accessed on the IEEE-SA Website at the following URL: <http://standards.ieee.org/findstds/errata/index.html>. Users are encouraged to check this URL for errata periodically.

## Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE-SA Website at <http://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or nondiscriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

## Participants

At the time this standard was submitted to the IEEE-SA Standards Board for approval, the Point of Care Devices Working Group had the following membership:

**Jan Wittenber**, *Chair*  
**Stefan Schlichting**, *Subgroup Chair*

Bjoern Anderson  
 Malcolm Clarke  
 Todd Cooper  
 Chris Courville  
 Michael Faughn  
 Kenneth Fuchs  
 John Garguilo

Frank Golatowski  
 David Gregorczyk  
 Kai Hassing  
 John Hatcliff  
 Stefan Karl  
 Martin Kasparick  
 Koichiro Matsumoto  
 Joerg-Uwe Meyer

Stephan Poehlsen  
 Tracy Rausch  
 John Rhoads  
 Paul Schluter  
 Masato Tanaka  
 Eugene Vasserman  
 Stan Wiley

The following members of the individual balloting committee voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Bjoern Andersen  
 Susan Burgess  
 Craig Carlson  
 Keith Chow  
 Malcolm Clarke  
 Todd Cooper  
 Sourav Dutta  
 Kenneth Fuchs  
 Joel Goergen  
 Frank Golatowski

David Gregorczyk  
 Randall Groves  
 Kai Hassing  
 Robert Heile  
 Werner Hoelzl  
 Noriyuki Ikeuchi  
 Atsushi Ito  
 Stefan Karl  
 Piotr Karocki

Martin Kasparick  
 Joerg-Uwe Meyer  
 Stephan Poehlsen  
 Beth Pumo  
 Stefan Schlichting  
 Eugene Stoudenmire  
 Walter Struppler  
 J. Wiley  
 Jan Wittenber  
 Oren Yuen

When the IEEE-SA Standards Board approved this standard on 6 December 2017, it had the following membership:

**Jean-Philippe Faure**, *Chair*  
**Gary Hoffman**, *Vice Chair*  
**John D. Kulick**, *Past Chair*  
**Konstantinos Karachalios**, *Secretary*

Chuck Adams  
 Masayuki Ariyoshi  
 Ted Burse  
 Stephen Dukes  
 Doug Edwards  
 J. Travis Griffith  
 Michael Janezic

Thomas Kochy  
 Joseph L. Koepfinger\*  
 Kevin Lu  
 Daleep Mohla  
 Damir Novosel  
 Ronald C. Petersen  
 Annette D. Reilly

Robby Robson  
 Dorothy Stanley  
 Adrian Stephens  
 Mehmet Ulema  
 Phil Wennblom  
 Howard Wolfman  
 Yu Yuan

\*Member Emeritus



This is a preview of "ISO/IEEE 11073-10207...". [Click here to purchase the full version from the ANSI store.](#)

## Introduction

This introduction is not part of IEEE Std 11073-10207-2017, Health informatics—Point-of-care medical device communication—Part 10207: Domain Information and Service Model for Service-Oriented Point-of-Care Medical Device Communication.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. They provide automatic and detailed electronic data capture of patient vital signs information and device operational data. The primary goals of these standards are to

- Provide real-time, plug-and-play interoperability for patient-connected medical devices.
- Facilitate the efficient exchange of vital signs and medical device data, acquired at the point of care, in all health care environments.

“Real-time” means that data from multiple devices can be retrieved, time correlated, and displayed or processed in fractions of a second. “Plug-and-play” means that all the clinician has to do is make the connection—the systems automatically detect, configure, and communicate without any other human interaction.

“Efficient exchange of medical device data” means that information that is captured at the point of care (e.g., patient vital signs data) can be archived, retrieved, and processed by many different types of applications without extensive software and equipment support and without needless loss of information. The standards are especially targeted at acute and continuing care devices, such as patient monitors, ventilators, infusion pumps, and electrocardiogram (EKG) devices. This family of standards can be layered together to provide connectivity optimized for the specific devices being interfaced.

Note that normative statements of requirements are presented in this standard in the following manner:

Rnnnn: Statement text here.

where "nnnn" is replaced by a number that is unique among the requirements in this standard and thereby forms a unique requirement identifier, for example,

R0007: All HANDLEs SHALL be unique within one MDIB sequence of a SERVICE PROVIDER.

## Contents

1. Overview .....	21
1.1 Scope .....	21
1.2 Purpose .....	21
2. Normative references.....	22
3. Definitions and notational conventions .....	23
3.1 Definitions .....	23
3.2 Notational conventions .....	25
4. Introduction to BICEPS.....	28
5. Participant Model .....	30
5.1 Overview .....	30
5.2 Background.....	30
5.3 Descriptive part.....	34
5.4 State part.....	41
6. Alert signal delegation.....	49
6.1 General .....	49
6.2 Delegable alert signals.....	49
6.3 Fallback alert signals .....	52
7. Communication Model.....	53
7.1 General .....	53
7.2 Message exchange patterns.....	53
7.3 Service Model.....	54
7.4 Message Model.....	59
8. Extension Model.....	62
9. Discovery Model .....	63
9.1 General .....	63
9.2 Implicit discovery .....	63
9.3 Explicit discovery.....	63
10. Nonfunctional requirements .....	64
10.1 General .....	64
10.2 Patient safety considerations.....	64
10.3 Cybersecurity considerations.....	64
10.4 Clinical effectiveness.....	65
10.5 Regulatory considerations .....	65
11. Conformance .....	66
11.1 Overview .....	66
11.2 General format.....	66
11.3 ICS tables.....	67
Annex A (normative) Extension Model .....	70
A.1 Extension .....	70
A.2 ExtensionType .....	71

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

A.3 MustUnderstand.....	71
Annex B (normative) Participant Model .....	72
B.1 AbstractAlertDescriptor .....	74
B.2 AbstractAlertState.....	74
B.3 AbstractAlertState/@ActivationState .....	75
B.4 AbstractComplexDeviceComponentDescriptor .....	75
B.5 AbstractComplexDeviceComponentDescriptor/AlertSystem .....	76
B.6 AbstractComplexDeviceComponentDescriptor/Sco.....	77
B.7 AbstractComplexDeviceComponentState.....	78
B.8 AbstractContextDescriptor.....	79
B.9 AbstractContextState .....	80
B.10 AbstractContextState/@ContextAssociation .....	81
B.11 AbstractContextState/@BindingMdbVersion.....	81
B.12 AbstractContextState/@UnbindingMdbVersion.....	81
B.13 AbstractContextState/@BindingStartTime .....	82
B.14 AbstractContextState/@BindingEndTime .....	82
B.15 AbstractContextState/Validator .....	82
B.16 AbstractContextState/Identification.....	83
B.17 AbstractDescriptor .....	83
B.18 AbstractDescriptor/@Handle.....	84
B.19 AbstractDescriptor/@DescriptorVersion .....	84
B.20 AbstractDescriptor/@SafetyClassification .....	84
B.21 AbstractDescriptor/Type.....	85
B.22 AbstractDeviceComponentDescriptor .....	86
B.23 AbstractDeviceComponentDescriptor/ProductionSpecification .....	86
B.24 AbstractDeviceComponentDescriptor/ProductionSpecification/SpecType .....	87
B.25 AbstractDeviceComponentDescriptor/ProductionSpecification/ ProductionSpec .....	88
B.26 AbstractDeviceComponentDescriptor/ProductionSpecification/ComponentId .....	88
B.27 AbstractDeviceComponentState .....	89
B.28 AbstractDeviceComponentState/@ActivationState.....	90
B.29 AbstractDeviceComponentState/@OperatingHours.....	90
B.30 AbstractDeviceComponentState/@OperatingCycles.....	90
B.31 AbstractDeviceComponentState/CalibrationInfo.....	90
B.32 AbstractDeviceComponentState/NextCalibration.....	91
B.33 AbstractDeviceComponentState/PhysicalConnector .....	92
B.34 AbstractMetricDescriptor.....	93
B.35 AbstractMetricDescriptor/@MetricCategory.....	94
B.36 AbstractMetricDescriptor/@DerivationMethod.....	94
B.37 AbstractMetricDescriptor/@MetricAvailability .....	95
B.38 AbstractMetricDescriptor/@MaxMeasurementTime.....	95
B.39 AbstractMetricDescriptor/@MaxDelayTime.....	95
B.40 AbstractMetricDescriptor/@DeterminationPeriod.....	96
B.41 AbstractMetricDescriptor/@LifeTimePeriod .....	96
B.42 AbstractMetricDescriptor/@ActivationDuration .....	96
B.43 AbstractMetricDescriptor/Unit .....	97
B.44 AbstractMetricDescriptor/BodySite.....	98
B.45 AbstractMetricDescriptor/Relation .....	99
B.46 AbstractMetricDescriptor/Relation/@Kind .....	100
B.47 AbstractMetricDescriptor/Relation/@Entries .....	100
B.48 AbstractMetricDescriptor/Relation/Code.....	101
B.49 AbstractMetricDescriptor/Relation/Identification.....	102
B.50 AbstractMetricState .....	103
B.51 AbstractMetricState/@ActivationState.....	104
B.52 AbstractMetricState/@ActiveDeterminationPeriod.....	104
B.53 AbstractMetricState/@LifeTimePeriod .....	104

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.54	AbstractMetricState/BodySite.....	105
B.55	AbstractMetricState/PhysicalConnector .....	106
B.56	AbstractMetricValue .....	106
B.57	AbstractMetricValue/@StartTime .....	107
B.58	AbstractMetricValue/@StopTime .....	107
B.59	AbstractMetricValue/@DeterminationTime .....	107
B.60	AbstractMetricValue/MetricQuality .....	107
B.61	AbstractMetricValue/MetricQuality/@Validity.....	108
B.62	AbstractMetricValue/MetricQuality/@Mode .....	108
B.63	AbstractMetricValue/MetricQuality/@Qi.....	108
B.64	AbstractMetricValue/Annotation .....	109
B.65	AbstractMetricValue/Annotation/Type.....	109
B.66	AbstractMultiState .....	110
B.67	AbstractMultiState/@Handle .....	111
B.68	AbstractMultiState/Category .....	111
B.69	AbstractOperationDescriptor .....	112
B.70	AbstractOperationDescriptor/@OperationTarget .....	113
B.71	AbstractOperationDescriptor/@MaxTimeToFinish.....	113
B.72	AbstractOperationDescriptor/@InvocationEffectiveTimeout.....	113
B.73	AbstractOperationDescriptor/@Retriggerable .....	113
B.74	AbstractOperationDescriptor/@AccessLevel .....	114
B.75	AbstractOperationState .....	114
B.76	AbstractOperationState/@OperatingMode .....	115
B.77	AbstractSetStateOperationDescriptor .....	115
B.78	AbstractSetStateOperationDescriptor/ModifiableData .....	116
B.79	AbstractState .....	116
B.80	AbstractState/@StateVersion.....	117
B.81	AbstractState/@DescriptorHandle .....	117
B.82	AbstractState/@DescriptorVersion .....	117
B.83	ActivateOperationDescriptor .....	118
B.84	ActivateOperationDescriptor/Argument .....	119
B.85	ActivateOperationDescriptor/Argument/ArgName .....	119
B.86	ActivateOperationDescriptor/Argument/Arg.....	120
B.87	ActivateOperationState .....	120
B.88	AlertConditionDescriptor.....	121
B.89	AlertConditionDescriptor/@Kind .....	122
B.90	AlertConditionDescriptor/@Priority.....	122
B.91	AlertConditionDescriptor/@DefaultConditionGenerationDelay .....	122
B.92	AlertConditionDescriptor/@CanEscalate .....	122
B.93	AlertConditionDescriptor/@CanDeescalate .....	123
B.94	AlertConditionDescriptor/Source .....	123
B.95	AlertConditionDescriptor/CauseInfo .....	123
B.96	AlertConditionState .....	124
B.97	AlertConditionState/@ActualConditionGenerationDelay .....	125
B.98	AlertConditionState/@ActualPriority .....	125
B.99	AlertConditionState/@Rank .....	125
B.100	AlertConditionState/@Presence.....	125
B.101	AlertConditionState/@DeterminationTime .....	125
B.102	AlertSignalDescriptor .....	126
B.103	AlertSignalDescriptor/@ConditionSignaled .....	127
B.104	AlertSignalDescriptor/@Manifestation.....	127
B.105	AlertSignalDescriptor/@Latching .....	128
B.106	AlertSignalDescriptor/@DefaultSignalGenerationDelay .....	128
B.107	AlertSignalDescriptor/@MinSignalGenerationDelay.....	128
B.108	AlertSignalDescriptor/@MaxSignalGenerationDelay .....	128
B.109	AlertSignalDescriptor/@SignalDelegationSupported.....	128

This is a preview of "ISO/IEEE 11073-10207...". [Click here to purchase the full version from the ANSI store.](#)

B.110	AlertSignalDescriptor/@AcknowledgementSupported	128
B.111	AlertSignalDescriptor/@AcknowledgeTimeout	129
B.112	AlertSignalState	129
B.113	AlertSignalState/@ActualSignalGenerationDelay	130
B.114	AlertSignalState/@Presence	130
B.115	AlertSignalState/@Location	130
B.116	AlertSignalState/@Slot	130
B.117	AlertSystemDescriptor	131
B.118	AlertSystemDescriptor/@MaxPhysiologicalParallelAlarms	132
B.119	AlertSystemDescriptor/@MaxTechnicalParallelAlarms	132
B.120	AlertSystemDescriptor/@SelfCheckPeriod	132
B.121	AlertSystemDescriptor/AlertCondition	132
B.122	AlertSystemDescriptor/AlertSignal	133
B.123	AlertSystemState	134
B.124	AlertSystemState/@LastSelfCheck	135
B.125	AlertSystemState/@SelfCheckCount	135
B.126	AlertSystemState/@PresentPhysiologicalAlarmConditions	135
B.127	AlertSystemState/@PresentTechnicalAlarmConditions	135
B.128	AlertSystemState/SystemSignalActivation	136
B.129	ApprovedJurisdictions	136
B.130	ApprovedJurisdictions/ApprovedJurisdiction	137
B.131	BaseDemographics	137
B.132	BaseDemographics/Givenname	138
B.133	BaseDemographics/Middlename	138
B.134	BaseDemographics/Familyname	138
B.135	BaseDemographics/Birthname	139
B.136	BaseDemographics/Title	139
B.137	BatteryDescriptor	139
B.138	BatteryDescriptor/CapacityFullCharge	140
B.139	BatteryDescriptor/CapacitySpecified	141
B.140	BatteryDescriptor/VoltageSpecified	141
B.141	BatteryState	142
B.142	BatteryState/@ChargeStatus	143
B.143	BatteryState/@ChargeCycles	143
B.144	BatteryState/CapacityRemaining	144
B.145	BatteryState/Voltage	144
B.146	BatteryState/Current	145
B.147	BatteryState/Temperature	145
B.148	BatteryState/RemainingBatteryTime	146
B.149	CalibrationInfo	146
B.150	CalibrationInfo/@ComponentCalibrationState	147
B.151	CalibrationInfo/@Type	147
B.152	CalibrationInfo/@Time	147
B.153	CalibrationInfo/CalibrationDocumentation	147
B.154	CalibrationInfo/CalibrationDocumentation/Documentation	148
B.155	CalibrationInfo/CalibrationDocumentation/CalibrationResult	148
B.156	CalibrationInfo/CalibrationDocumentation/CalibrationResult/Code	149
B.157	CalibrationInfo/CalibrationDocumentation/CalibrationResult/Value	150
B.158	CauseInfo	150
B.159	CauseInfo/RemedyInfo	150
B.160	CauseInfo/Description	151
B.161	ChannelDescriptor	152
B.162	ChannelDescriptor/Metric	153
B.163	ChannelState	154
B.164	ClinicalInfo	155
B.165	ClinicalInfo/Type	156

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.166 ClinicalInfo/Code.....	157
B.167 ClinicalInfo/Criticality .....	157
B.168 ClinicalInfo/Description .....	158
B.169 ClinicalInfo/RelatedMeasurement .....	159
B.170 ClinicalInfo/RelatedMeasurement/@Validity.....	159
B.171 ClinicalInfo/RelatedMeasurement/Value.....	160
B.172 ClinicalInfo/RelatedMeasurement/ReferenceRange .....	160
B.173 ClinicalInfo/RelatedMeasurement/ReferenceRange/Range .....	161
B.174 ClinicalInfo/RelatedMeasurement/ReferenceRange/Meaning .....	162
B.175 ClockDescriptor .....	163
B.176 ClockDescriptor/@Resolution .....	163
B.177 ClockDescriptor/TimeProtocol .....	164
B.178 ClockState .....	165
B.179 ClockState/@DateAndTime .....	166
B.180 ClockState/@RemoteSync .....	166
B.181 ClockState/@Accuracy .....	166
B.182 ClockState/@LastSet .....	167
B.183 ClockState/@TimeZone .....	167
B.184 ClockState/@CriticalUse .....	167
B.185 ClockState/ActiveSyncProtocol.....	167
B.186 ClockState/ReferenceSource.....	168
B.187 CodedValue.....	169
B.188 CodedValue/@Code .....	170
B.189 CodedValue/@CodingSystem .....	170
B.190 CodedValue/@CodingSystemVersion.....	171
B.191 CodedValue/@SymbolicCodeName.....	171
B.192 CodedValue/CodingSystemName.....	171
B.193 CodedValue/ConceptDescription.....	172
B.194 CodedValue/Translation .....	172
B.195 CodedValue/Translation/@Code .....	173
B.196 CodedValue/Translation/@CodingSystem .....	173
B.197 CodedValue/Translation/@CodingSystemVersion.....	173
B.198 ContainmentTree.....	173
B.199 ContainmentTree/Entry.....	174
B.200 ContainmentTreeEntry.....	175
B.201 ContainmentTreeEntry/Type.....	176
B.202 DistributionSampleArrayMetricDescriptor.....	177
B.203 DistributionSampleArrayMetricDescriptor/@Resolution.....	178
B.204 DistributionSampleArrayMetricDescriptor/TechnicalRange.....	179
B.205 DistributionSampleArrayMetricDescriptor/DomainUnit.....	180
B.206 DistributionSampleArrayMetricDescriptor/DistributionRange .....	181
B.207 DistributionSampleArrayMetricState.....	182
B.208 DistributionSampleArrayMetricState/MetricValue .....	183
B.209 DistributionSampleArrayMetricState/PhysiologicalRange.....	184
B.210 EnsembleContextDescriptor .....	185
B.211 EnsembleContextState .....	186
B.212 EnumStringMetricDescriptor.....	187
B.213 EnumStringMetricDescriptor/AllowedValue.....	188
B.214 EnumStringMetricDescriptor/AllowedValue/Value .....	188
B.215 EnumStringMetricDescriptor/AllowedValue/Type .....	189
B.216 EnumStringMetricDescriptor/AllowedValue/Identification .....	190
B.217 EnumStringMetricDescriptor/AllowedValue/Characteristic .....	190
B.218 EnumStringMetricState.....	191
B.219 ImagingProcedure .....	192
B.220 ImagingProcedure/AccessionIdentifier.....	193
B.221 ImagingProcedure/RequestedProcedureId.....	193

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.222 ImagingProcedure/StudyInstanceUid.....	194
B.223 ImagingProcedure/ScheduledProcedureStepId.....	195
B.224 ImagingProcedure/Modality.....	196
B.225 ImagingProcedure/ProtocolCode.....	197
B.226 InstanceIdentifier.....	198
B.227 InstanceIdentifier/@Root.....	199
B.228 InstanceIdentifier/@Extension.....	199
B.229 InstanceIdentifier/Type.....	200
B.230 InstanceIdentifier/IdentifierName.....	201
B.231 LimitAlertConditionDescriptor.....	202
B.232 LimitAlertConditionDescriptor/@AutoLimitSupported.....	203
B.233 LimitAlertConditionDescriptor/MaxLimits.....	203
B.234 LimitAlertConditionState.....	204
B.235 LimitAlertConditionState/@MonitoredAlertLimits.....	205
B.236 LimitAlertConditionState/@AutoLimitActivationState.....	205
B.237 LimitAlertConditionState/Limits.....	206
B.238 LocalizedText.....	206
B.239 LocalizedText/@Ref.....	207
B.240 LocalizedText/@Lang.....	207
B.241 LocalizedText/@Version.....	208
B.242 LocalizedText/@TextWidth.....	208
B.243 LocationContextDescriptor.....	208
B.244 LocationContextState.....	209
B.245 LocationContextState/LocationDetail.....	210
B.246 LocationDetail.....	211
B.247 LocationDetail/@PoC.....	211
B.248 LocationDetail/@Room.....	211
B.249 LocationDetail/@Bed.....	211
B.250 LocationDetail/@Facility.....	212
B.251 LocationDetail/@Building.....	212
B.252 LocationDetail/@Floor.....	212
B.253 LocationReference.....	212
B.254 LocationReference/Identification.....	213
B.255 LocationReference/LocationDetail.....	214
B.256 MdDescription.....	214
B.257 MdDescription/@DescriptionVersion.....	215
B.258 MdDescription/Mds.....	216
B.259 Mdib.....	217
B.260 Mdib/MdDescription.....	218
B.261 Mdib/MdState.....	218
B.262 MdsDescriptor.....	219
B.263 MdsDescriptor/MetaData.....	220
B.264 MdsDescriptor/MetaData/Udi.....	221
B.265 MdsDescriptor/MetaData/Udi/DeviceIdentifier.....	221
B.266 MdsDescriptor/MetaData/Udi/HumanReadableForm.....	221
B.267 MdsDescriptor/MetaData/Udi/Issuer.....	222
B.268 MdsDescriptor/MetaData/Udi/Jurisdiction.....	222
B.269 MdsDescriptor/MetaData/LotNumber.....	223
B.270 MdsDescriptor/MetaData/Manufacturer.....	223
B.271 MdsDescriptor/MetaData/ManufactureDate.....	224
B.272 MdsDescriptor/MetaData/ExpirationDate.....	224
B.273 MdsDescriptor/MetaData/ModelName.....	224
B.274 MdsDescriptor/MetaData/ModelNumber.....	225
B.275 MdsDescriptor/MetaData/SerialNumber.....	225
B.276 MdsDescriptor/SystemContext.....	226
B.277 MdsDescriptor/Clock.....	227



This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.278 MdsDescriptor/Battery	228
B.279 MdsDescriptor/ApprovedJurisdictions	229
B.280 MdsDescriptor/Vmd	229
B.281 MdsState	230
B.282 MdsState/@Lang	231
B.283 MdsState/@OperatingMode	231
B.284 MdsState/OperatingJurisdiction	232
B.285 MdState	232
B.286 MdState/@StateVersion	233
B.287 MdState/State	233
B.288 MeansContextDescriptor	234
B.289 MeansContextState	235
B.290 Measurement	236
B.291 Measurement/@MeasuredValue	236
B.292 Measurement/MeasurementUnit	237
B.293 NeonatalPatientDemographicsCoreData	238
B.294 NeonatalPatientDemographicsCoreData/GestationalAge	239
B.295 NeonatalPatientDemographicsCoreData/BirthLength	239
B.296 NeonatalPatientDemographicsCoreData/BirthWeight	240
B.297 NeonatalPatientDemographicsCoreData/HeadCircumference	240
B.298 NeonatalPatientDemographicsCoreData/Mother	241
B.299 NumericMetricDescriptor	242
B.300 NumericMetricDescriptor/@Resolution	243
B.301 NumericMetricDescriptor/@AveragingPeriod	243
B.302 NumericMetricDescriptor/TechnicalRange	244
B.303 NumericMetricState	245
B.304 NumericMetricState/@ActiveAveragingPeriod	246
B.305 NumericMetricState/MetricValue	246
B.306 NumericMetricState/PhysiologicalRange	247
B.307 NumericMetricValue	248
B.308 NumericMetricValue/@Value	248
B.309 OperatingJurisdiction	249
B.310 OperatorContextDescriptor	249
B.311 OperatorContextState	250
B.312 OperatorContextState/OperatorDetails	251
B.313 OrderDetail	252
B.314 OrderDetail/Start	252
B.315 OrderDetail/End	252
B.316 OrderDetail/Performer	253
B.317 OrderDetail/Service	254
B.318 OrderDetail/ImagingProcedure	255
B.319 PatientContextDescriptor	255
B.320 PatientContextState	256
B.321 PatientContextState/CoreData	257
B.322 PatientDemographicsCoreData	258
B.323 PatientDemographicsCoreData/Sex	259
B.324 PatientDemographicsCoreData/PatientType	259
B.325 PatientDemographicsCoreData/DateOfBirth	260
B.326 PatientDemographicsCoreData/Height	260
B.327 PatientDemographicsCoreData/Weight	261
B.328 PatientDemographicsCoreData/Race	261
B.329 PersonParticipation	262
B.330 PersonParticipation/Role	263
B.331 PersonReference	264
B.332 PersonReference/Identification	264
B.333 PersonReference/Name	265



This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.334 PhysicalConnectorInfo .....	265
B.335 PhysicalConnectorInfo/@Number .....	266
B.336 PhysicalConnectorInfo/Label .....	266
B.337 Range .....	267
B.338 Range/@Lower .....	267
B.339 Range/@Upper .....	268
B.340 Range/@StepWidth .....	268
B.341 Range/@RelativeAccuracy .....	268
B.342 Range/@AbsoluteAccuracy .....	268
B.343 RealTimeSampleArrayMetricDescriptor .....	269
B.344 RealTimeSampleArrayMetricDescriptor/@Resolution .....	270
B.345 RealTimeSampleArrayMetricDescriptor/@SamplePeriod .....	270
B.346 RealTimeSampleArrayMetricDescriptor/TechnicalRange .....	271
B.347 RealTimeSampleArrayMetricState .....	272
B.348 RealTimeSampleArrayMetricState/MetricValue .....	273
B.349 RealTimeSampleArrayMetricState/PhysiologicalRange .....	274
B.350 RemedyInfo .....	274
B.351 RemedyInfo/Description .....	275
B.352 SampleArrayValue .....	276
B.353 SampleArrayValue/@Samples .....	277
B.354 SampleArrayValue/ApplyAnnotation .....	277
B.355 SampleArrayValue/ApplyAnnotation/@AnnotationIndex .....	277
B.356 SampleArrayValue/ApplyAnnotation/@SampleIndex .....	277
B.357 ScoDescriptor .....	278
B.358 ScoDescriptor/Operation .....	279
B.359 ScoState .....	280
B.360 ScoState/@InvocationRequested .....	281
B.361 ScoState/@InvocationRequired .....	281
B.362 ScoState/OperationGroup .....	281
B.363 ScoState/OperationGroup/@OperatingMode .....	281
B.364 ScoState/OperationGroup/@Operations .....	282
B.365 ScoState/OperationGroup/Type .....	282
B.366 SetAlertStateOperationDescriptor .....	283
B.367 SetAlertStateOperationState .....	284
B.368 SetComponentStateOperationDescriptor .....	285
B.369 SetComponentStateOperationState .....	286
B.370 SetContextStateOperationDescriptor .....	287
B.371 SetContextStateOperationState .....	288
B.372 SetMetricStateOperationDescriptor .....	289
B.373 SetMetricStateOperationState .....	290
B.374 SetStringOperationDescriptor .....	291
B.375 SetStringOperationDescriptor/@MaxLength .....	292
B.376 SetStringOperationState .....	292
B.377 SetStringOperationState/AllowedValues .....	292
B.378 SetStringOperationState/AllowedValues/Value .....	293
B.379 SetValueOperationDescriptor .....	293
B.380 SetValueOperationState .....	294
B.381 SetValueOperationState/AllowedRange .....	295
B.382 StringMetricDescriptor .....	296
B.383 StringMetricState .....	297
B.384 StringMetricState/MetricValue .....	298
B.385 StringMetricValue .....	299
B.386 StringMetricValue/@Value .....	299
B.387 SystemContextDescriptor .....	300
B.388 SystemContextDescriptor/PatientContext .....	301
B.389 SystemContextDescriptor/LocationContext .....	302

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.390 SystemContextDescriptor/EnsembleContext .....	302
B.391 SystemContextDescriptor/OperatorContext.....	303
B.392 SystemContextDescriptor/WorkflowContext .....	304
B.393 SystemContextDescriptor/MeansContext .....	304
B.394 SystemContextState .....	305
B.395 SystemSignalActivation.....	306
B.396 SystemSignalActivation/@Manifestation .....	306
B.397 SystemSignalActivation/@State .....	307
B.398 VmdDescriptor.....	307
B.399 VmdDescriptor/ApprovedJurisdictions.....	308
B.400 VmdDescriptor/Channel .....	308
B.401 VmdState .....	309
B.402 VmdState/OperatingJurisdiction .....	310
B.403 WorkflowContextDescriptor.....	311
B.404 WorkflowContextState.....	312
B.405 WorkflowContextState/WorkflowDetail .....	313
B.406 WorkflowContextState/WorkflowDetail/Patient .....	314
B.407 WorkflowContextState/WorkflowDetail/AssignedLocation .....	314
B.408 WorkflowContextState/WorkflowDetail/VisitNumber.....	315
B.409 WorkflowContextState/WorkflowDetail/DangerCode .....	316
B.410 WorkflowContextState/WorkflowDetail/RelevantClinicalInfo .....	317
B.411 WorkflowContextState/WorkflowDetail/RequestedOrderDetail.....	318
B.412 WorkflowContextState/WorkflowDetail/RequestedOrderDetail/ ReferringPhysician.....	319
B.413 WorkflowContextState/WorkflowDetail/RequestedOrderDetail/ RequestingPhysician.....	319
B.414 WorkflowContextState/WorkflowDetail/RequestedOrderDetail/ PlacerOrderNumber .....	320
B.415 WorkflowContextState/WorkflowDetail/PerformedOrderDetail.....	321
B.416 WorkflowContextState/WorkflowDetail/PerformedOrderDetail/ FillerOrderNumber .....	322
B.417 WorkflowContextState/WorkflowDetail/PerformedOrderDetail/ ResultingClinicalInfo.....	323
B.418 AlertActivation .....	323
B.419 AlertConditionKind .....	324
B.420 AlertConditionMonitoredLimits .....	324
B.421 AlertConditionPriority .....	324
B.422 AlertConditionReference .....	325
B.423 AlertSignalManifestation .....	325
B.424 AlertSignalPresence.....	325
B.425 AlertSignalPrimaryLocation .....	326
B.426 CalibrationState.....	326
B.427 CalibrationType .....	326
B.428 CodeIdentifier .....	326
B.429 ComponentActivation .....	327
B.430 ContextAssociation .....	327
B.431 DerivationMethod .....	327
B.432 EntryRef.....	328
B.433 GenerationMode .....	328
B.434 Handle .....	328
B.435 HandleRef.....	328
B.436 LocalizedTextContent .....	329
B.437 LocalizedTextRef.....	329
B.438 LocalizedTextWidth.....	329
B.439 MdsOperatingMode .....	330
B.440 MeasurementValidity.....	330
B.441 MetricAvailability.....	331
B.442 MetricCategory .....	331
B.443 OperatingMode .....	331
B.444 OperationRef.....	332
B.445 PatientType .....	332

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

B.446 QualityIndicator .....	332
B.447 RealTimeValueType .....	333
B.448 ReferencedVersion .....	333
B.449 SafetyClassification .....	333
B.450 Sex .....	334
B.451 SymbolicCodeName .....	334
B.452 Timestamp .....	334
B.453 TimeZone .....	335
B.454 VersionCounter .....	335
B.455 ContainmentTreeInfo .....	335
B.456 ContainmentTreeInfo/@HandleRef .....	335
B.457 ContainmentTreeInfo/@ParentHandleRef .....	336
B.458 ContainmentTreeInfo/@EntryType .....	336
B.459 ContainmentTreeInfo/@ChildrenCount .....	336
B.460 MdibVersionGroup .....	336
B.461 MdibVersionGroup/@MdibVersion .....	336
B.462 MdibVersionGroup/@SequenceId .....	337
B.463 MdibVersionGroup/@InstanceId .....	337
Annex C (normative) Message Model .....	338
C.1 Activate .....	339
C.2 Activate/Argument .....	339
C.3 Activate/Argument/ArgValue .....	339
C.4 ActivateResponse .....	340
C.5 DescriptionModificationReport .....	340
C.6 DescriptionModificationReport/ReportPart .....	341
C.7 DescriptionModificationReport/ReportPart/@ParentDescriptor .....	342
C.8 DescriptionModificationReport/ReportPart/@ModificationType .....	342
C.9 DescriptionModificationReport/ReportPart/Descriptor .....	342
C.10 DescriptionModificationReport/ReportPart/State .....	343
C.11 EpisodicAlertReport .....	344
C.12 EpisodicComponentReport .....	344
C.13 EpisodicContextReport .....	345
C.14 EpisodicMetricReport .....	346
C.15 EpisodicOperationalStateReport .....	346
C.16 GetContainmentTree .....	347
C.17 GetContainmentTree/HandleRef .....	347
C.18 GetContainmentTreeResponse .....	348
C.19 GetContainmentTreeResponse/ContainmentTree .....	348
C.20 GetContextStates .....	349
C.21 GetContextStates/HandleRef .....	349
C.22 GetContextStatesByFilter .....	350
C.23 GetContextStatesByFilter/@ContextType .....	350
C.24 GetContextStatesByFilter/Filter .....	350
C.25 GetContextStatesByFilterResponse .....	351
C.26 GetContextStatesByFilterResponse/ContextState .....	352
C.27 GetContextStatesByIdentification .....	353
C.28 GetContextStatesByIdentification/@ContextType .....	353
C.29 GetContextStatesByIdentification/Identification .....	354
C.30 GetContextStatesByIdentificationResponse .....	355
C.31 GetContextStatesByIdentificationResponse/ContextState .....	356
C.32 GetContextStatesResponse .....	357
C.33 GetContextStatesResponse/ContextState .....	358
C.34 GetDescriptor .....	359
C.35 GetDescriptor/HandleRef .....	359
C.36 GetDescriptorResponse .....	360

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

C.37 GetDescriptorResponse/Descriptor .....	360
C.38 GetDescriptorsFromArchive .....	361
C.39 GetDescriptorsFromArchive/DescriptorRevisions .....	362
C.40 GetDescriptorsFromArchive/TimeFrame .....	362
C.41 GetDescriptorsFromArchive/Handle .....	363
C.42 GetDescriptorsFromArchiveResponse .....	363
C.43 GetDescriptorsFromArchiveResponse/Descriptor .....	364
C.44 GetLocalizedText .....	365
C.45 GetLocalizedText/Ref .....	365
C.46 GetLocalizedText/Version .....	366
C.47 GetLocalizedText/Lang .....	366
C.48 GetLocalizedText/TextWidth .....	366
C.49 GetLocalizedText/NumberOfLines .....	367
C.50 GetLocalizedTextResponse .....	367
C.51 GetLocalizedTextResponse/Text .....	368
C.52 GetMdDescription .....	369
C.53 GetMdDescription/HandleRef .....	369
C.54 GetMdDescriptionResponse .....	369
C.55 GetMdDescriptionResponse/MdDescription .....	370
C.56 GetMdib .....	370
C.57 GetMdibResponse .....	371
C.58 GetMdibResponse/Mdib .....	371
C.59 GetMdState .....	372
C.60 GetMdState/HandleRef .....	372
C.61 GetMdStateResponse .....	373
C.62 GetMdStateResponse/MdState .....	373
C.63 GetStatesFromArchive .....	374
C.64 GetStatesFromArchive/StateRevisions .....	374
C.65 GetStatesFromArchive/TimeFrame .....	375
C.66 GetStatesFromArchive/Handle .....	375
C.67 GetStatesFromArchiveResponse .....	376
C.68 GetStatesFromArchiveResponse/State .....	377
C.69 GetSupportedLanguages .....	377
C.70 GetSupportedLanguagesResponse .....	378
C.71 GetSupportedLanguagesResponse/Lang .....	378
C.72 ObservedValueStream .....	379
C.73 ObservedValueStream/Value .....	379
C.74 ObservedValueStream/Value/@Metric .....	380
C.75 ObservedValueStream/Value/@StateVersion .....	380
C.76 ObservedValueStream/Value/Value .....	380
C.77 OperationInvokedReport .....	381
C.78 OperationInvokedReport/ReportPart .....	382
C.79 OperationInvokedReport/ReportPart/@OperationHandleRef .....	382
C.80 OperationInvokedReport/ReportPart/@OperationTarget .....	382
C.81 OperationInvokedReport/ReportPart/InvocationInfo .....	383
C.82 OperationInvokedReport/ReportPart/InvocationSource .....	383
C.83 PeriodicAlertReport .....	384
C.84 PeriodicComponentReport .....	385
C.85 PeriodicContextReport .....	385
C.86 PeriodicMetricReport .....	386
C.87 PeriodicOperationalStateReport .....	387
C.88 Retrievability .....	387
C.89 Retrievability/By .....	388
C.90 SetAlertState .....	388
C.91 SetAlertState/ProposedAlertState .....	389
C.92 SetAlertStateResponse .....	389

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

C.93 SetComponentState.....	390
C.94 SetComponentState/ProposedComponentState.....	391
C.95 SetComponentStateResponse.....	392
C.96 SetContextState.....	392
C.97 SetContextState/ProposedContextState.....	393
C.98 SetContextStateResponse.....	394
C.99 SetMetricState.....	394
C.100 SetMetricState/ProposedMetricState.....	395
C.101 SetMetricStateResponse.....	396
C.102 SetString.....	396
C.103 SetString/RequestedStringValue.....	397
C.104 SetStringResponse.....	397
C.105 SetValue.....	397
C.106 SetValue/RequestedNumericValue.....	398
C.107 SetValueResponse.....	398
C.108 SystemErrorReport.....	399
C.109 SystemErrorReport/ReportPart.....	399
C.110 SystemErrorReport/ReportPart/ErrorCode.....	400
C.111 SystemErrorReport/ReportPart/ErrorInfo.....	401
C.112 WaveformStream.....	401
C.113 WaveformStream/State.....	402
C.114 AbstractAlertReport.....	403
C.115 AbstractAlertReport/ReportPart.....	404
C.116 AbstractAlertReport/ReportPart/AlertState.....	404
C.117 AbstractComponentReport.....	405
C.118 AbstractComponentReport/ReportPart.....	405
C.119 AbstractComponentReport/ReportPart/ComponentState.....	406
C.120 AbstractContextReport.....	407
C.121 AbstractContextReport/ReportPart.....	407
C.122 AbstractContextReport/ReportPart/ContextState.....	408
C.123 AbstractGet.....	409
C.124 AbstractGetResponse.....	409
C.125 AbstractMetricReport.....	410
C.126 AbstractMetricReport/ReportPart.....	411
C.127 AbstractMetricReport/ReportPart/MetricState.....	411
C.128 AbstractOperationalStateReport.....	412
C.129 AbstractOperationalStateReport/ReportPart.....	413
C.130 AbstractOperationalStateReport/ReportPart/OperationState.....	413
C.131 AbstractReport.....	414
C.132 AbstractReportPart.....	414
C.133 AbstractReportPart/SourceMds.....	415
C.134 AbstractSet.....	415
C.135 AbstractSet/OperationHandleRef.....	415
C.136 AbstractSetResponse.....	416
C.137 AbstractSetResponse/InvocationInfo.....	416
C.138 InvocationInfo.....	417
C.139 InvocationInfo/TransactionId.....	417
C.140 InvocationInfo/InvocationState.....	418
C.141 InvocationInfo/InvocationError.....	418
C.142 InvocationInfo/InvocationErrorMessage.....	419
C.143 RetrievabilityInfo.....	419
C.144 RetrievabilityInfo/@Method.....	420
C.145 RetrievabilityInfo/@UpdatePeriod.....	420
C.146 TimeFrame.....	420
C.147 TimeFrame/@Start.....	420
C.148 TimeFrame/@End.....	421

This is a preview of "ISO/IEEE 11073-10207...". Click here to purchase the full version from the ANSI store.

C.149 VersionFrame.....	421
C.150 VersionFrame/@Start .....	421
C.151 VersionFrame/@End.....	421
C.152 DescriptionModificationType.....	422
C.153 InvocationError.....	422
C.154 InvocationState .....	423
C.155 RetrievabilityMethod .....	423
C.156 TransactionId .....	423
Annex D (informative) Examples.....	424
D.1 Containment tree.....	424
D.2 MDIB versioning.....	424
D.3 MDIB sequencing.....	429
D.4 BICEPS discovery .....	430
D.5 Remote control.....	432
D.6 Alert signal generation.....	433
Annex E (informative) Bibliography.....	435