First edition 2004-10-15

# Piston-operated volumetric instruments — Determination of uncertainty for volume measurements made using the photometric method

Instruments volumétriques actionnés par piston — Détermination de l'incertitude de mesure pour les mesurages volumétriques au moyen de la méthode photométrique



#### **PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org Published in Switzerland

## Contents

Fore	word	iv
1	Scope	1
2	Uncertainty analysis for the replaceable cell photometric method as described in ISO 8655-7:—, Annex A	4
2.1	Uncertainties of each measurand	
2.1	Mathematical model of method.	
2.2	Sensitivity coefficients	
2.3 2.4	Uncertainty budget	
3	Uncertainty analysis for the flow cell photometric method as given in ISO 8655-7:—,	
	Annex B	
3.1	Uncertainties of each measurand	16
3.2	Mathematical model of method	
3.3	Sensitivity coefficients	16
3.4	Combined standard uncertainty	25
4	Traceability of the method	25
5	Validation of photometric method	
5.1	Principle of validation	
5.2	Equipment for validation	
5.3	Density of solutions	
5.4	Experimental results	
5.5	Additional validation studies of the photometric method	
Biblie	ography	28

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In exceptional circumstances, when a technical committee has collected data of a different kind from that which is normally published as an International Standard ("state of the art", for example), it may decide by a simple majority vote of its participating members to publish a Technical Report. A Technical Report is entirely informative in nature and does not have to be reviewed until the data it provides are considered to be no longer valid or useful.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TR 16153 was prepared by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*.