

This is a preview of "ISO/TR 16153:2023". [Click here to purchase the full version from the ANSI store.](#)

Second edition
2023-02

Determination of uncertainty for volume measurements of a piston- operated volumetric apparatus using a photometric method

*Détermination de l'incertitude de mesure pour les mesurages
volumétriques des appareils volumétriques à piston au moyen de la
méthode photométrique*



Reference number
ISO/TR 16153:2023(E)

© ISO 2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

This is a preview of "ISO/TR 16153:2023". [Click here to purchase the full version from the ANSI store.](#)

Contents

| | Page |
|--|-----------|
| Foreword..... | iv |
| Introduction..... | v |
| 1 Scope..... | 1 |
| 2 Normative references..... | 1 |
| 3 Terms and definitions..... | 1 |
| 4 Modelling the measurement..... | 1 |
| 5 General procedure for the uncertainty calculation..... | 4 |
| 6 Standard uncertainty components associated with the measuring system (photometric measurement procedure)..... | 4 |
| 6.1 General information on the estimation of standard uncertainty components..... | 4 |
| 6.2 Standard uncertainty of the copper(II) chloride solution volume..... | 5 |
| 6.3 Standard uncertainty of the cuvette mixture absorbance at 520 nm..... | 5 |
| 6.4 Standard uncertainty of the cuvette starting absorbance at 730 nm..... | 6 |
| 6.5 Standard uncertainty of the cuvette starting absorbance at 520 nm..... | 6 |
| 6.6 Standard uncertainty of the volume of Ponceau S and copper(II) chloride solutions used in calibrators..... | 7 |
| 6.7 Standard uncertainty of the absorbances of the calibrator solutions..... | 7 |
| 7 Standard uncertainty components associated with the POVA..... | 9 |
| 7.1 Standard uncertainty of the resolution..... | 9 |
| 7.2 Standard uncertainty of the setting..... | 9 |
| 7.3 Standard uncertainty related to air cushion effects..... | 9 |
| 7.4 Standard uncertainty of the cubic expansion coefficient..... | 9 |
| 8 Standard uncertainty components associated with the liquid delivery process..... | 10 |
| 8.1 Repeatability (experimental standard deviation)..... | 10 |
| 8.2 Reproducibility..... | 10 |
| 9 Combined standard uncertainty of measurement associated with the systematic error of mean volume..... | 11 |
| 10 Sensitivity coefficients..... | 12 |
| 11 Coverage factor k..... | 13 |
| 12 Expanded uncertainty of measurement associated with the volume \bar{V}..... | 14 |
| 13 Example for determining the uncertainty of the volume measurement of POVA..... | 14 |
| 13.1 Measurement conditions..... | 14 |
| 13.2 Results..... | 16 |
| 13.2.1 Standard uncertainty of the POVA mean volume..... | 16 |
| 13.2.2 Expanded uncertainty of the measurement..... | 17 |
| 13.2.3 Result of measurement..... | 17 |
| 13.2.4 Uncertainty in use and corrections for pressure changes..... | 17 |
| 13.2.5 General remarks..... | 17 |
| 13.2.6 Note on the conformity with ISO/IEC Guide 98-3..... | 17 |
| Annex A (informative) Approaches for the estimation of uncertainty in use of a single delivered volume..... | 19 |
| Annex B (informative) Volume correction due to pressure changes..... | 22 |
| Bibliography..... | 24 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*.

This second edition cancels and replaces the first edition (ISO/TR 16153:2004), which has been technically revised.

The main changes are as follows:

- the term “standard deviation of the mean delivered volume” has been replaced in this document by “repeatability” according to ISO/IEC Guide 99 (VIM);
- a new uncertainty calculation example has been supplied;
- new uncertainty components have been added, namely, reproducibility, air cushion and resolution;
- new [Annex A](#) concerning the uncertainty in use of a single delivered volume has been added;
- new [Annex B](#) concerning volume correction due to pressure changes has been added.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This is a preview of "ISO/TR 16153:2023". [Click here to purchase the full version from the ANSI store.](#)

Introduction

The example given in this document is informative and supports the requirements found in ISO 8655-8:2022, 9.4 and ISO 8655-7:2022, 4.2, to perform an estimation of measurement uncertainty when calibrating POVA according to the measurement procedures described in these documents and the principles of ISO/IEC Guide 98-3.

The revision of this document coincides with a major revision of the ISO 8655 series in 2022, reflecting the state-of-the-art measurement procedures and approaches for the estimation of measurement uncertainty.