

First edition
2005-01-15

**Accuracy (trueness and precision) of
measurement methods and results —
Practical guidance for the use of
ISO 5725-2:1994 in designing,
implementing and statistically analysing
interlaboratory repeatability and
reproducibility results**

*Exactitude (justesse et fidélité) des résultats et méthodes de mesure —
Lignes directrices pratiques pour l'utilisation de l'ISO 5725-2:1994 pour
la conception, la mise en œuvre et l'analyse statistique des résultats de
répétabilité et de reproductibilité interlaboratoires*



Reference number
ISO/TR 22971:2005(E)

© ISO 2005

This is a preview of "ISO/TR 22971:2005". [Click here to purchase the full version from the ANSI store.](#)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

This is a preview of "ISO/TR 22971:2005". [Click here to purchase the full version from the ANSI store.](#)

Contents

Page

Foreword	v
Introduction	vii
1 Scope.....	1
2 Organization of an inter-laboratory programme	1
2.1 Requirements for a precision experiment	1
2.2 The responsibilities of the personnel involved in a precision experiment.....	1
3 Critical examination of the data.....	3
3.1 Description of the data	3
3.2 Tests for outliers	7
3.3 Conclusions.....	9
4 Estimation of repeatability and reproducibility standard deviations	11
4.1 Analysis of variance	11
4.2 Description of the model	11
4.3 Examples.....	13
4.4 Use of repeatability and reproducibility limits	15
5 Worked examples using statistical software	16
5.1 General	16
5.2 Determination of sulfur content in coal	16
5.3 Thermometric titration of creosote oil	23
Annex A (normative) Symbols and abbreviations	30
Bibliography	33

This is a preview of "ISO/TR 22971:2005". [Click here to purchase the full version from the ANSI store.](#)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In exceptional circumstances, when a technical committee has collected data of a different kind from that which is normally published as an International Standard ("state of the art", for example), it may decide by a simple majority vote of its participating members to publish a Technical Report. A Technical Report is entirely informative in nature and does not have to be reviewed until the data it provides are considered to be no longer valid or useful.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TR 22971 was prepared by Technical Committee ISO/TC 69, *Applications of statistical methods*, Subcommittee SC 6, *Measurement methods and results*.

This is a preview of "ISO/TR 22971:2005". [Click here to purchase the full version from the ANSI store.](#)

Introduction

ISO 5725 consists of six parts, the general structure of which is shown in Figure 1.

ISO 5725-2 was developed as a guidance document for ISO Technical Committees and other organizations responsible for undertaking inter-laboratory studies for characterizing the variability of standard measurement methods. Two measures of variability, repeatability and reproducibility, are accepted in many disciplines as representative of data encountered in measurement processes.

Repeatability refers to the variability among measurements made on nominally identical samples or materials under identical circumstances. It is recognized that, because of unknown or uncontrollable factors which influence the measurement process, repeated measurements will usually not agree. The extent of this variability can be expressed by a standard deviation, called the repeatability standard deviation, of the results of within-laboratory comparisons.

Reproducibility refers to the variability among measurements made on identical samples or materials under differing conditions by different laboratories following the same standard measurement method. Reproducibility includes effects caused by differences among instruments, reagents, operators, laboratories, and environmental conditions. The variability of results under these conditions may be described by a standard deviation called the reproducibility standard deviation.

This guidance document is divided into four clauses in addition to the Scope (Clause 1):

- Clause 2, Organization of an inter-laboratory programme, deals with the organization of the inter-laboratory test and covers the roles of the executive officer, laboratory personnel, and statistician in preparing for and administering the test; the choice of materials and levels of interest for the test; and the selection of laboratories. It also describes how the number of replicate measurements (to be made on each sample) is to be statistically treated and the manner in which the resulting data are to be reported.
- Clause 3, Critical examination of the data, deals with data using graphical and numerical procedures. Guidance is given as to when data are anomalous, i.e. if they are inconsistent with other data from the study, and for outlier tests that are used to identify the presence or absence of anomalous data.
- Clause 4, Estimation of repeatability and reproducibility standard deviations, deals with the estimation and interpretation of repeatability and reproducibility standard deviations. Also included is a comparison of the relative contributions of the repeatability and reproducibility standard deviations to the total variability of the test method.
- Clause 5, Worked examples using statistical software, deals with worked examples that highlight various techniques that can be used.

It is recommended that this guidance document be read in conjunction with ISO 5725-2 and should not be used as a replacement for ISO 5725-2.

This is a preview of "ISO/TR 22971:2005". [Click here to purchase the full version from the ANSI store.](#)




Figure 1 — Structure of ISO 5725 — Application of a standardized test method to the analyses of a sample or product in different laboratories