ANSI Z80.11-2007

AMERICAN NATIONAL STANDARI

for Ophthalmics – Laser Systems for Corneal Reshaping



This is a preview of "ANSI Z80.11-2007". Click here to purchase the full version from the ANSI store.					

This is a preview of "ANSI Z80.11-2007". Click here to purchase the full version from the ANSI store.

ANSI[®] Z80.11-2007

American National Standard for Ophthalmics –

Laser Systems for Corneal Reshaping

Secretariat

Information Technology Industry Council

Approved July 31, 2007

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -

Optical Laboratories Association Z80 Secretariat 11096 Lee Highway A101 Fairfax, VA 22030-5039

Published by

Optical Laboratories Association 11096 Lee Highway A101 Fairfax, VA 22030-5039

Copyright © 2007 by Optical Laboratories Association All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Pa	ige
Fore	word	.iv
1	Scope and purpose	. 1
2	Normative references	. 1
3	Definitions	. 3
4	Mechanical, thermal, and environmental requirements	. 5
4.1	Combination of different devices	. 5
4.2	Materials	. 5
4.3	Resistance to transport and storage conditions	. 5
5	Safety requirements	. 6
5.1	Protection against contaminants	. 6
5.2	Protection against toxins and allergens	. 6
5.3	Photobiological hazards	. 6
5.4	Thermal hazards	. 6
5.5	Mechanical hazards	. 6
5.6	Electrical safety	. 7
5.7 5.7.1	Radiation safetyLight hazards	
5.8	Gas safety (for gas lasers)	. 7
5.9	Safety in use	. 7
5.10	System hazard analysis	. 8
6	Optical requirements	. 8
6.1	Alignment system	. 8
6.2	Fail safe monitoring	. 8
6.3	System calibration	. 8
7	System control and performance	. 9
7.1	Software	. 9
8	Clinical evaluation	. 9
8.1	Clinical investigation plan	. 9
8.2	Surgical procedure	. 9
8.3	Reporting periods and evaluations	. 9
8.4	Adverse events	10
9	Test methods	10
9.1 9.1.1 9.1.2	Verification of alignment system	10

i

	F	Page			
9.2 9.2.1	Verification of the cylinder axis alignment				
-	Procedure				
10	Accompanying documents	11			
11	Marking	. 11			
Annexes					
Α	Spectral weighting function for ultraviolet radiation hazard analysis	. 12			
В	Methods for system calibration	. 14			
B.1	Plastic plate ablation and measurement	. 14			
B.1.1	Materials	. 14			
B.1.2	Procedure	. 14			
B.2	Laminated calibration plate method	. 14			
B.2.1	Materials	. 14			
B.2.2	Procedure	. 15			
С	Characterization of laser ablation beams and treatment patterns	. 16			
C.1	Ablation characteristics of the beam	. 16			
C.2	Mathematical models and simulations	. 17			
C.3	Validation of ablation algorithm software	. 17			
D	Guidance on clinical study design of refractive procedures that use laser systems for corneal reshaping	18			
D.1	General	. 18			
D.2	Study objectives	. 18			
D.3	Design of the clinical study	. 19			
D.4	Study duration	. 19			
D.5	Enrollement of subjects	. 19			
D.6	Inclusion and exclusion criteria for subject selection	. 20			
D.6.1	Inclusion criteria	. 20			
D.6.2	2 Exclusion criteria	. 20			
D.7	Examination schedule	. 21			
D.8	Evaluations and methodology	. 23			
D.8.1	Visual acuity and manifest refraction	. 23			
D.8.2	Measurement of intraocular pressure	. 24			
D.8.3	Subject questionnaire	. 25			
D.8.4	Mesopic pupil size	. 25			
	Contrast sensitivity				

		Page
D.8.6	6 Low contrast letter acuity testing	26
D.8.7	7 Specular microscopy	27
D.9	Adverse device effects/adverse events	29
E	Statistical sample size considerations	30
E.1	Statistical symbols and definitions	30
E.2	Calculation of necessary sample sizes	31
E.2.1	1 Sample size based on safety estimates	31
E.2.2	2 Sample size based on effectiveness estimates using noninferiority hypothesis testing	32
E.3	Clinical substudies	34
E.3.1	1 Sample size for a contract sensitivity study	34
E.3.2	2 Sample size for endothelial cell density study	34
F	Presentation of results of clinical studies	36
F.1	General	36
F.2	Accountability of subjects	36
F.3	Refractive stability	38
F.4	Safety	39
F.5	Effectiveness	39
F.6	Retreatment	40
G	Bibliography	41
Tabl	les	
D.1	Recommended postoperative examination schedule	22
E.1	Statistical symbols and definitions	30
E.2	Normal quantiles to use in equations	31
F.1	Accountabilty by postoperative visit	37

Foreword (This foreword is not part of American National Standard ANSI Z80.11-2007.)

This American National Standard was developed to address the expressed needs of those members of the ophthalmic community who correct the refractive errors of the human eye using laser refractive correction procedures, those who manufacture the lasers systems for corneal reshaping used to perform these procedures, and those who insure the public interest by ensuring that these systems are made in such a way so that they may perform their function in a safe and effective way when used correctly by those skilled in their use.

It must be realized that correcting the refractive error of the human eye with laser corrective surgery is a medical procedure involving not only a laser system for corneal reshaping but also (1) other devices used during surgery, (2) the assessment of the refractive state of the eye prior to surgery, (3) decisions on the best approach to take for treatment that involve not only the judgment of the physician but the desires of the patient, (4) the postsurgical care, and (5) the effects of healing, known and unknown. This standard only addresses the laser system for corneal reshaping and makes no attempt to standardize the procedure itself.

However, in response to a perceived need, informative annexes have been included in the standard to give guidance on types of clinical testing deemed to be adequate to ensure that the entire procedure is safe and effective. It was felt that a service would be performed for those in the field if this information were to be placed in a public document where it would be readily available to all.

While it is true that the outcome of a laser refractive procedure will not prove acceptable if the laser system for corneal reshaping used to perform it is not adequate for the task, it cannot be assumed that a laser system is inadequate if outcomes are not acceptable, as this may be the result of deficiencies in other important parts of the medical procedure. Thus, no claim is made that, if a laser system for corneal reshaping complies with this standard for the tasks it is designed to perform, that surgical procedures performed with the laser will have acceptable outcomes.

This standard was created by a special working group created by the Z80 Subcommittee on Medical Ophthalmic Devices and included experts in the field of laser refractive correction from the clinical, manufacturing and academic areas of the ophthalmic community and by experts from the regulatory agency given oversight in this field in the United States of America.

This standard contains seven annexes. Annexes A and B are normative and are considered part of the standard. Annexes C through G are informative and are not considered part of this standard.

Suggestions for improvement of this standard will be welcome. They should be sent to the Optical Laboratories Association, P.O. Box 2000, Merrifield, VA 22116-2000.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas C. White, M.D., Chairman Quido Cappelli, Vice-Chairman Robert Rosenberg, O.D., Secretary Daniel Torgersen, Secretariat

Organization Represented	Name of Representative
Advance Medical Technologies Association	Douglas J. Fortunato Glenn Davies (Alt.)
	Stanley J. Rogaski (Alt.)
American Academy of Ophthalmology	Bernie Liebler (Alt.)
	Gerhard Cibis (Alt.)
	Norman Lanphear (Alt.) Paul F.Vinger (Alt.)
American Academy of Optometry	David S. Loshin
American Ceramic Society	Lyle Rubin
American Glaucoma Society	Herbert Hoover (Alt.)
	Douglas J. Rhee (Alt.)
American Optometric Association	Donald Pitts
	William J. Benjamin (Alt.)
	William L. Brown (Alt.)
	Robert Rosenberg (Alt.) Jeffrey Weaver (Alt.)
American Society of Cataract and Refractive Surgery	Stephen Klyce
	Jack T. Holladay (Alt.)
	Stephen H. Johnson (Alt.)
Contact Lens Institute	Ed Schilling
Contact Lens Manufacturers Association	Quido Cappelli Jan Svochak (Alt.)
Department of Veterans Affairs	John Townsend
	Sharon R. Atkin (Alt.)
Federated Cornea Societies	Michael W. Belin Ć
	David Glasser (Alt.)
Food & Drug Administration	Donald Calogero Robert Landry (Alt.)
National Association of Optometrists & Opticians	Franklin D. Rozak
National Academy of Opticianry	James Elkin
Optical Laboratories Association	Daniel Torgersen
•	Gregory S. Jacobs (Alt.)
	Virginia "Susie" Lesher (Alt.)
Optical Society of America	Jonathan Schwartz (Alt.)
Opticians Association of America	Catherine I angley
	Tina M. Schott (Alt.)
Prevent Blindness	Christine Bradley
	Jeff Todd (Alt.)
Sunglass Association of America	
	Scott Macguffie (Alt.) Rick Van Arnam (Alt.)
US Leader to ISO TC 172/SC7	Charles F. Campbell
Vision Council of America	
	Steve Drake (Alt.)
	Neil Roche (Alt.)
	Dick Whitney (Alt.)
	Kenneth O. Wood (Alt.)

The Working Group for Laser Systems for Corneal Reshaping, which falls under the Medical Ophthalmic Devices Subcommittee, had the following members who worked on the writing of this standard:

Charles E. Campbell, WG Chair Gerhard Cibis, Subcommittee Chair

John J. Alpar MD
Michael W. Belin MD
Don Calagero
lan Cox PhD
Bruce Drum PhD
Daniel Durrie MD
Malvina Eydelman MD
Robert Faaland
Barbara Fant PharmD
Emily Graham
Gene Hilmantel OD
Jack T. Holladay MD

Guy M. Kezirian MD
Michael Kliewer PhD
Stephen Klyce PhD
Kitty Lombardo
Scott MacCrea MD
Shirley K. McGarvey
Mohinder Merchea OD PhD
Audrey Munnerlyn
Charles Munnerlyn PhD
Terrence O'Brien MD
Mark Odrich MD
Jan S. Peterson

Alan Russell PhD
Alex Sacharoff PhD
Stephen C. Schallhorn MD
David Sliney PhD
Doyle Stulting MD
Yoshitaka Suzuki
Larry Thibos PhD
Mark Von Tress PhD
Thomas C. White MD
Bonita Wong OD

AMERICAN NATIONAL STANDARD

ANSI Z80.11-2007

American National Standard for Ophthalmics –

Laser Systems for Corneal Reshaping

1 Scope and purpose

This standard applies to any laser system whose primary intended use is to alter the shape of the cornea through the removal of corneal tissue, resulting in the improvement of visual performance.

This standard addresses the vocabulary, performance requirements, labeling, and clinical investigations necessary for this type of device.

2 Normative references

The following standards contain provisions that, through reference in this text, constitute provisions of this American National Standard. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of the IEC and ISO maintain registers of currently valid International Standards.

ANSI Z136.1-2000, Safe Use of Lasers

ANSI Z136.3-2005, Safe Use of Lasers in Health Care Facilities

IEC 60601-1:2001, Medical Electrical Equipment – Part 1: General Requirements for Safety

IEC 60601-1-1:2000, Medical Electrical Equipment – Part 1-1: General Requirements for Safety – Collateral Standard: Safety Requirements for Medical Electrical Systems

IEC 60601-1-2:1993, Medical Electrical Equipment – Part 1: General Requirements for Safety – 2. Collateral Standard: Electromagnetic Compatibility – Requirements and Tests

IEC 60601-1-4:2000, Medical Electrical Equipment – Part 1-4: General Requirements for Safety – Collateral Standard: Programmable Electrical Medical Systems

IEC 60601-2-22:1995, Medical Electrical Equipment – Part 2: Particular Requirements for the Safety of Diagnostic and Therapeutic Laser Equipment

IEC 60825-1:2001, Safety of Laser Products – Part 1: Equipment Classification, Requirements and User's Guide

ISO 8598:1996, Optics and optical instruments – Focimeters