

ANSI Z80.17-2013

# American National Standard

*for Ophthalmics –  
Focimeters*



This is a preview of "ANSI Z80.17-2013". [Click here to purchase the full version from the ANSI store.](#)

**ANSI®**  
**Z80.17-2013**  
(Revision of  
ANSI Z80.17-2008)

American National Standard  
for Ophthalmics –

**Focimeters**

Secretariat  
**The Vision Council**

Approved March 14, 2013  
**American National Standards Institute, Inc.**

## American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

**CAUTION NOTICE:** This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

### *Developed by*

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council  
Z80 Secretariat  
225 Reinekers Lane, Suite 700  
Alexandria, VA 22314

### *Published by*

The Vision Council  
225 Reinekers Lane, Suite 700  
Alexandria, VA 22314

Copyright © 2013 by The Vision Council  
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

## Contents

	Page
Foreword .....	ii
1 Scope .....	1
2 Normative References.....	1
3 Definitions.....	1
4 Design requirements and recommendations for general purpose focimeters. ....	4
5 Accuracy Requirements .....	6
6 Testing.....	9
7 Design requirements and recommendations for test lenses .....	13

## Tables

1 Tolerances of measured vertex power for continuously indicating instruments.....	7
2 Tolerances of measured prismatic power for continuously indicating instruments.....	7
3 Permissible deviations of measured vertex power reading from the nominal value of the test lenses for digitally rounding instruments .....	8
4 Permissible deviations of measured prismatic power reading from the nominal value of the test lenses for digitally rounding instruments .....	8
5 Cylinder power and cylinder axis requirements for focimeters that measure sphere, cylinder and axis simultaneously .....	9
6 Tolerance on repeatability of cylinder axis measurements for 0.25 D cylinder (degrees) .....	9
7 Design range for the standard test lenses.....	14
8 Tolerances for spherical test lenses .....	14
9 Tolerances for prismatic test lenses .....	15

## Figures

1 Permissible movement of the adjusting rail.....	5
2 Example of a lens support for spectacle lenses in cross section .....	6
3 Test lens for verifying cylinder power and cylinder axis requirements for focimeters that measure sphere, cylinder and axis simultaneously.....	16

## Annex

A Manufacture of test lenses for focimeters.....	17
--	----

## Foreword (This foreword is not part of American National Standard ANSI Z80.17-2013.)

This standard was developed by a group of experts under the direction of the ANSI Instrument subcommittee chair, William L. Brown, O.D., Ph.D. This standard defines the requirement for devices that measure the sphere and cylindrical vertex power, prismatic power and axis for spectacle and contact lenses. Accuracy requirements for tolerances or deviation of readings are also defined.

In this revised standard, a method is added for checking the cylinder power and cylinder axis requirements for focimeters that measure sphere, cylinder and axis simultaneously.

Suggestions for improvement of this standard will be welcome. Comments should be sent to The Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas White, M.D., Chairman  
 Quido Cappelli, Vice-Chairman  
 Robert Rosenberg, O.D., Secretary  
 Daniel Torgersen, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies Association .....	Paul Ludington Richard Courtney (Alt.) Glenn Davies (Alt.) Bernie Liebler (Alt.)
American Academy of Ophthalmology .....	Tom White Carl Tubbs (Alt.) Pradeep Ramulu (Alt.) Shannon Curtis (Alt.)
American Academy of Optometry.....	David S. Loshin
American Ceramic Society .....	Lyle Rubin Herbert Hoover (Alt.)
American Glaucoma Society .....	Steven Gedde Douglass Rhee (Alt.)
American Optometric Association .....	Karl Citek Robert Rosenberg (Alt.) William Benjamin (Alt.) William Brown (Alt.)
American Society of Cataract and Refractive Surgery .....	Stephen Klyce Jack Holladay (Alt.)
Association for Research in Vision & Ophthalmology (ARVO)...	Jan Bergmanson Gerald McGwin (Alt.) Elmer T. Yu (Alt.)
Contact Lens Institute.....	Stan Rogaski Peter Mathers (Alt.)
Contact Lens Manufacturers Association .....	Quido Cappelli Troy Miller (Alt.)
Department of Veterans Affairs .....	John Townsend Michael White (Alt.)
Federated Cornea Societies/ASCRS .....	Michael Belin David Glasser (Alt.)
Food & Drug Administration CDRH/ Division of Federal Agencies.....	Donald Calogero Robert James (Alt.) Dexiu Shi (Alt.)
National Association of Optometrists & Opticians .....	Nick Mileti Doug Pelkey (Alt.) Franklin Rozak (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
Optical Laboratories Association.....	Richard Tinson
	Jonathan Schwartz (Alt.)
Opticians Association of America.....	Tom Hicks
	Chris Allen (Alt.)
Sunglass Association of America.....	Kenneth Frederick
	Scott MacGuffie (Alt.)
	Rick Van Arnam (Alt.)
U.S. Assistant TAG Leader to ISO TC 172/SC7 .....	Jeff Endres
The Vision Council .....	Jeff Endres
	Neil Roche (Alt.)
	Dick Whitney (Alt.)
	Mike Vitale (Alt.)

The subcommittee on Ophthalmic Instruments, which developed this standard, had the following participating members:

William L. Brown, O. D., Ph.D, Chairman	Charles E. Campbell
	Bruce Drum
	Robert Rosenberg
	Tom White
	Dick Whitney





## American National Standard for Ophthalmics –

# Focimeters

### 1. Scope

This standard specifies requirements for continuously indicating and digitally rounding focimeters with which the vertex powers and prismatic powers of spherical and astigmatic lenses, including lenses mounted in frames, can be measured and with which lenses can be oriented and marked.

### 2. Normative references

The following standard contains provisions which, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

ANSI Z80.1-2005, *Ophthalmics – Prescription Ophthalmic Lenses - Recommendations*

ISO 8429:1986, *Optics and optical instruments – Ophthalmology – Graduated Dial Scale*

### 3. Definitions

For the purposes of this standard, the following definitions apply.

**3.1. adjusting rail:** Movable rail or bar used as the reference axis for spectacles during measurement, which is aligned to be perpendicular to the optical axis of the focimeter and parallel to the axis direction of 0° to 180°. This is also called the lens table or frame rest.

**3.2. astigmatic power lens:** Lens bringing a paraxial pencil of parallel rays to two separate line foci mutually at right angles and hence, unlike a spherical lens, having two principal powers.

One of these powers may be zero, with the corresponding focal line at infinity. Lenses referred to as toric lenses, sphero-cylindrical lenses and cylinder lenses are all astigmatic lenses.