

AMERICAN NATIONAL STANDARD



*for Ophthalmic Instruments –
Fundamental Requirements
and Test Methods*

This is a preview of "ANSI Z80.25-1996 (R2...)". [Click here to purchase the full version from the ANSI store.](#)

ANSI®
Z80.25-1996 (R2002)

American National Standard
for Ophthalmic Instruments –
**Fundamental Requirements
and Test Methods**

Secretariat

Optical Laboratories Association

Approved December 11, 2001
Reaffirmed February 28, 2002

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires verification by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Published by
Optical Laboratories Association
P. O. Box 2000
Merrifield, VA 22116-2000
U.S.A.

Copyright © 1996 by Optical Laboratories Association.
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

CPII.96.500

Contents

	Page
Foreword	ii
1 Scope	1
2 Normative references	1
3 Definitions	2
4 Fundamental requirements	3
5 Environmental conditions	3
6 Special requirements for active ophthalmic instruments	3
7 Test methods	5
8 Information supplied by the manufacturer	6
Tables	
1 Environmental conditions in use	8
2 Storage conditions	8
3 Transport conditions	8
4 Environmental tests	9
Annexes	
A Spectral weighting functions for retinal hazard analysis	10
B Radiometric measures	13
C Example of Information given by manufacturer on the avoidance of over-exposure to a potentially optical radiation hazard	15
D Product related standards	16

Foreword (This foreword is not part of American National Standard Z80.25-1996)

This is a standard for ophthalmic instruments which specifies characteristics which are common to many such instruments, such as: safety issues, environmental use conditions, electrical requirements, storage and transport conditions, radiation hazards, marking requirements, accompanying documents and others. Since these characteristics are common to many ophthalmic instruments it is economical to standardize them in a single horizontal standard rather than include these characteristics in each of the individual instrument standards.

In 1982, the Optical Laboratories Association (OLA) assumed the responsibility of the Secretariat; and in 1985 the Z80 Committee became an accredited standards committee. The scope of the Z80 committee is for the establishment of standards that shall apply to ophthalmic lenses and to equipment, instruments and to processes used in the final fabrication level which affect their performance; to ophthalmic frames, sunglasses, and fashion eyewear; to contact lenses and accessories for their use; to intraocular implant lenses; to low vision aids and ophthalmic contact devices in addition to contact lenses; and to optical instrumentation used in ophthalmic procedures and vision evaluation.

The current ophthalmic standards are drafted by subcommittees of the Z80 committee. These subcommittees may, in turn, establish working groups, as needed, to address detailed areas in the assigned project.

Suggestions for improvement of this standard will be welcome. They should be sent to the Optical Laboratories Association, P.O. Box 2000, Merrifield, VA 22116-2000, U.S.A.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

David E. Eifrig, Chairman
F. Dow Smith, Vice Chairman
Robeert Rosenberg, Secretary

<i>Organization Represented</i>	<i>Name of Representative</i>
American Academy of Ophthalmology	David E. Eifrig Edmund Thall (Alt.) Thomas C. White (Alt.)
American Academy of Optometry	David S. Loshin
American Ceramic Society	John R. Hansen Jackson S. Stroud (Alt.) Herbert Hoover (Alt.) Clifton Wheeler (Alt.)
American Optometric Association	Robert Rosenberg William Benjamin (Alt.) Donald Pitts (Alt.) Gregory L. Stephens (Alt.)
American Society of Cataract & Refractive Surgery	John J. Alpar Jack T. Holladay (Alt.) Stephen H. Johnson (Alt.)
Contact Lens Institute	Paul C. Nicolson Richard Baron (Alt.)
Contact Lens Manufacturers Association	Quido Cappelli
Food and Drug Administration	Morris Waxler Robert Landry (Alt.) David Whipple (Alt.) Leroy Schroeder (Alt.) Everette T. Beers (Alt.)
Health Industry Manufacturers Association	Douglas J. Fortunato Dennis Hahn (Alt.) Janet Trunzo (Alt.) Stanley J. Rogaski (Alt.)
Industrial Safety Equipment Association	Arthur J. Salce Janice C. Bradley (Alt.) Dale B. Pfriem (Alt.)
National Academy of Opticianry	Jeffrey C. Snodgrass Floyd H. Holmgrain (Alt.)
National Association of Optometrists and Opticians	Arthur Newman John Piana (Alt.)
National Institute of Standards and Technology	Thomas Lettieri
National Society to Prevent Blindness	Todd Turriff
Opticians Association of America	C. Richard Sanders Paul Houghland (Alt.)
Optical Laboratories Association	Daniel Torgersen Joseph S. Gill (Alt.) Jeffrey S. Kosh (Alt.)
OMA The Optical Industry Association	Eugene A. Keeney John Young (Alt.) James E. Sheedy (Alt.) Richard Waido (Alt.) Richard Whitney (Alt.)
Optical Society of America	F. Dow Smith Richard A. Phillips (Alt.)
Sunglass Association of America	Thomas J. Loomis David Elliott (Alt.) James Pritts (Alt.) Thomas Fairbrother (Alt.)
United States Army	William Rimm David Hsieh (Alt.) Kenyon K. Kramer (Alt.) William Hately (Alt.) David Sliney (Alt.)
United States Air Force	Bob Buckingham
United States Navy	David Mazur Stanley Freed (Alt.)
Veterans Health Administration	Charles F. Mullen

The Z80 subcommittee on Ophthalmic Instruments, which developed this standard, had the following members:

James E. Sheedy, Chairman

David Eitrig
Charles Campbell
John J. Alpar
Rick Ambler
Bernard Grolman
David A. Hsieh
Gordon Keane
Jeffrey S. Kosh
Robert Landry
Scott Lewis
Dave A. Luce
James W. Maisel
Rush Manbert
David Marco
James C. Messier
Charles F. Mullen
Margaret Noel
Donald G. Pitts
Robert Rosenberg
Craig C. Ross
Edmond Thall
Richard E. VanArsdel
Morris Waxler
Tom White
Donald H. Winfield

American National Standard
for Ophthalmics –

Instruments – Fundamental Requirements and Test Methods

1 Scope and purpose

1.1 Scope

This standard is applicable to non-invasive, active and non-active ophthalmic instruments. This standard includes low vision aids and tonometers, but excludes other instruments which are used in contact with the globe of the eye. This standard is not applicable to operation microscopes, endoscopes and devices intended for laser treatment (surgery) of the eye. This standard takes precedent over the corresponding requirement of the other general standards quoted in clause 2, if differences exist. In addition to the requirements of this standard, the supplementary or modified requirements specified in the relevant product-related standards will apply. They are listed in annex D.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreement based on this standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 7944:1984 *Optics and optical instruments - Reference wavelength*¹⁾

ISO 9022-2:1994 *Optics and optical instruments - Environmental test methods, Part 2:*

*Cold, heat, humidity*¹⁾

ISO 9022-3:1994 *Optics and optical instruments - Environmental test methods, Part 3: Mechanical stresses*¹⁾

ISO 9022-5:1994 *Optics and optical instruments - Environmental test methods, Part 5: Combined cold, low air pressure*¹⁾

IEC 601-1:1988 *Safety of medical-electrical equipment, Part 1: general requirements*¹⁾

IEC 825:1993 *Radiation safety of laser products, equipment classification, requirements and user's guides*¹⁾

3 Definitions

For the purpose of this standard the following definition shall apply:

3.1 Corneal plane

The plane that is perpendicular to the optic axis of the instrument and that is tangential to the corneal surface closest to the instrument when in normal use.

3.2 Non-invasive ophthalmic instrument

Ophthalmic instrument which does not in whole or in part penetrate inside the body, either through a body orifice or through the surface of the body.

3.3 Active ophthalmic instrument

Any ophthalmic instrument connected to or equipped with a source of electrical energy.

3.4 Manufacturer

The natural or legal person, who places the instrument on the market.

1) Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036.