



BSI Standards Publication

Medical vehicles and their equipment — Road ambulances

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

National foreword

This British Standard is the UK implementation of EN 1789:2020+A1:2023. It supersedes BS EN 1789:2020, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to CEN text carry the number of the CEN amendment. For example, text altered by CEN amendment A1 is indicated by A1 A1.

The UK participation in its preparation was entrusted to Technical Committee CH/239, Rescue systems.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication has been prepared under a mandate given to the European Standards Organizations by the European Commission and the European Free Trade Association. It is intended to support requirements of the EU legislation detailed in the European Foreword. A European Annex, usually Annex ZA or ZZ, describes how this publication relates to that EU legislation.

For the Great Britain market (England, Scotland and Wales), if UK Government has designated this publication for conformity with UKCA marking (or similar) legislation, it may contain an additional National Annex. Where such a National Annex exists, it shows the correlation between this publication and the relevant UK legislation. If there is no National Annex of this kind, the relevant Annex ZA or ZZ in the body of the European text will indicate the relationship to UK regulation applicable in Great Britain. References to EU legislation may need to be read in accordance with the UK designation and the applicable UK law. Further information on designated standards can be found at www.bsigroup.com/standardsandregulation.

For the Northern Ireland market, UK law will continue to implement relevant EU law subject to periodic confirmation. Therefore Annex ZA/ZZ in the European text, and references to EU legislation, are still valid for this market.

UK Government is responsible for legislation. For information on legislation and policies relating to that legislation, consult the relevant pages of www.gov.uk.

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 24716 9

ICS 11.160; 43.160

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2020.

Amendments/corrigenda issued since publication

Date	Text affected
31 March 2024	Implementation of CEN amendment A1:2023

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

EUROPÄISCHE NORM

December 2023

ICS 11.160; 43.160

Supersedes EN 1789:2020

English Version

Medical vehicles and their equipment - Road ambulances

Véhicules de transport sanitaire et leurs équipements -
Ambulances routières

Rettungsdienstfahrzeuge und deren Ausrüstung -
Krankenkraftwagen

This European Standard was approved by CEN on 13 April 2020 and includes Amendment approved by CEN on 21 September 2023. This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 31 January 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	4
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	9
4 Requirements	10
4.1 General requirements	10
4.2 Electrical requirements	10
4.2.1 General	10
4.2.2 Electromagnetic compatibility (EMC)	11
4.2.3 Battery and alternator	11
4.2.4 Electrical installation	12
4.2.5 Visual warning system and audible warning system (siren)	13
4.2.6 Reversing systems	13
4.2.7 Exterior illumination lights	13
4.3 Vehicle body	14
4.3.1 Fire safety	14
4.3.2 Driver's seat configuration	14
4.3.3 Minimum passenger capacity	14
4.3.4 Bulkhead	15
4.3.5 Openings (doors, windows, emergency exits)	15
4.3.6 Loading area	16
4.4 Patient's compartment	18
4.4.1 General	18
4.4.2 Safety	18
4.4.3 Hygiene	18
4.4.4 Patient's compartment dimensions	19
4.4.5 Patient and crew seating	24
4.4.6 Ventilation and anaesthetic gas scavenging systems	25
4.4.7 Temperature control system	25
4.4.8 Interior lighting	26
4.4.9 Interior noise level	26
4.4.10 Holding system for infusion	26
4.4.11 Retention, fixation and restraint systems	27
4.4.12 Mass reserve	27
5 Testing	27
5.1 General	27
5.2 Testing of the interior noise level	27
5.2.1 Specific measurement conditions	27
5.2.2 Measurements	28
5.3 Testing of retention systems and fixation of the equipment in the patient's compartment	28
5.3.1 General	28
5.3.2 Testing of the stretcher fixation on the vehicle floor	30

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

5.3.3	Testing of the medical devices fixation.....	30
5.3.4	Testing of furniture.....	30
5.3.5	Test procedure	31
5.4	Testing of rounded edges and radius inside the patient's compartment.....	32
5.4.1	Testing of rounded edges.....	32
5.4.2	Testing of radius inside the patient's compartment.....	33
5.5	Procedure to verify the patient's compartment specifications	33
5.6	Procedure to verify the loading area specifications	33
5.6.1	General	33
5.6.2	Procedure to verify the loading angle of 16°	33
5.7	Procedure to verify the dimensions of the patient's compartment	35
5.7.1	Type A and B road ambulances.....	35
5.7.2	Type C road ambulances	35
5.8	Procedure to verify the seats dimensions of the patient's compartment.....	36
5.9	Testing of the ventilation system.....	36
5.10	Testing of the heating system.....	37
5.11	Testing of the cooling system	37
5.11.1	Test procedure	37
5.11.2	Testing of independent air conditioning system	38
5.12	Testing of interior lighting.....	38
5.13	Testing of infusion holding system	38
6	Equipment and medical devices.....	39
6.1	Provision of medical devices	39
6.2	Medical devices storage.....	39
6.3	Requirements for medical devices	39
6.3.1	General	39
6.3.2	Temperature	39
6.3.3	Humidity and ingress of liquids	39
6.3.4	Mechanical strength	39
6.3.5	Fixation of devices.....	40
6.3.6	Electrical safety	40
6.3.7	User interface.....	40
6.3.8	Gas installation	40
6.3.9	Marking and instructions	42
6.3.10	Maintenance	42
6.4	List of equipment.....	42
	Annex A (informative) Test summary.....	52
	Annex B (informative) Recognition.....	53
B.1	Recognition and visibility of ambulances.....	53
B.2	Recognition of crew	53
	Annex C (informative) Hygiene.....	54
	Annex D (informative) A-deviations	55
D.1	Deviation in Spain.....	55
☞	Annex E (normative) Electric propelled vehicles	56
	Annex ZA (informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [O] L 169] aimed to be covered.....	57
	Bibliography	58

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

European foreword

This document (EN 1789:2020+A1:2023) has been prepared by Technical Committee CEN/TC 239 “Rescue systems”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2024, and conflicting national standards shall be withdrawn at the latest by June 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

A1 This document supersedes EN 1789:2020 **A1**.

This document includes Amendment 1, approved by CEN on 2023-09-21.

The start and finish of text introduced or altered by amendment is indicated in the text by tags **A1** **A1**

This document has been prepared under a standardization request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA which is an integral part of this document.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

This is a preview of BS EN 1789:2020+A1:2023. [Click here to purchase the full version from the ANSI store.](#)

Introduction

Road ambulances are subject to a higher risk in use. The exact circumstances of operation cannot always be planned or anticipated in detail.

Vehicles are designed so as to be safe. Design requirements can be derived from European and national occupational safety and health legislation.

Under EU law, employers are responsible for carrying out a risk assessment (89/391/EEC, OSH framework directive) and for provision of safe work equipment (89/655/EEC, use of work equipment directive) that allows employees to work without their health being at risk.

The document was first developed in the late 1990s to define a common approach to requirements to enhance patient and crew safety. The document has evolved and matured through several amendments and revisions.

A1 This revision in 2020 had two key objectives: **A1**

- The first objective was to revise the technical side of the document with more manageable verification in mind, while maintaining the high quality and strict nature of the requirements.
- The second objective was to check all the references and regulations, paying special attention to EU regulations and updated standardization rules.

Testing of special purpose vehicle, such as an ambulance, is complex. Multiple functions (e.g. fixations, maintain systems, noise, illumination, heating, cooling etc.) may require numerous tests, which can be destructive. In this edition, carefully planned tests according to worst-case scenario strategies have reduced the number of destructive tests without sacrificing test qualities.

A1 The previous edition EN 1789:2007+A2:2014 **A1** contained a number of direct references to EU regulations. According to CEN Internal Regulations Part 3:2017 and to avoid duplication as well as outdated references and to enable use of this standard independently of the ECE rules, EU regulations and directives, these references have now been removed from the normative section of the standard.

This document is a reference document which can be used in support of regulations.

For the purpose of verification of an ambulance according to EU vehicle approval process, a section of **A1** EN 1789:2020 **A1** (i.e. patient's compartment) has been referenced directly in Regulation (EU) 2018/858.

A1 The energy sources for motor vehicles are in turmoil due to environmental fight against global climate warming. Alternative energies are becoming more regular in motor vehicles and electric vehicles are already common by most vehicle manufacturers.

In all standardization criteria, combustion engine characteristics have guided the requirements. Therefore, the most obvious ambulance standard requirements (EN 1789) need adjustments introduced by an Amendment to allow verification of electric engine ambulances as compliant to this document. **A1**

1 Scope

This document specifies requirements for the design, testing, performance and equipping of road ambulances used for the transport, monitoring, treatment and care of patients. It contains requirements for the patient's compartment in terms of the working environment, ergonomic design and the safety of the crew and patients. This document does not cover the training of the crew, which is the responsibility of the authority/authorities in the country where the ambulance is to be registered.

This document is applicable to road ambulances capable of transporting at least one patient on a stretcher and excludes the transportation of hospital beds.

This document also specifies requirements for ambulances intended to carry transport incubator systems.

This document covers the specific requirements of each type of road ambulance, which are designated according to the patient condition.

This document gives general requirements for medical devices carried in road ambulances and used therein and outside hospitals and clinics in situations where the ambient conditions can differ from normal indoor conditions.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

☞ EN 16165:2021, *Determination of slip resistance of pedestrian surfaces - Methods of evaluation* ☞

EN 3-7:2004+A1:2007, *Portable fire extinguishers - Part 7: Characteristics, performance requirements and test methods*

EN 443:2008, *Helmets for fire fighting in buildings and other structures*

☞ EN 455-1:2020+A1:2022, *Medical gloves for single use - Part 1: Requirements and testing for freedom from holes* ☞

EN 455-2:2015, *Medical gloves for single use - Part 2: Requirements and testing for physical properties*

EN 794-3:1998+A2:2009, *Lung ventilators - Part 3: Particular requirements for emergency and transport ventilators*

☞ EN ISO 20417:2021, *Medical devices - Information to be supplied by the manufacturer (ISO 20417:2021, Corrected version 2021-12)* ☞

EN 1865-1:2010+A1:2015, *Patient handling equipment used in ambulances - Part 1: General stretcher systems and patient handling equipment*

☞ prEN 1865-2:2022¹, *Patient handling equipment used in ambulances - Part 2: Power assisted stretcher* ☞

EN 1865-4:2012, *Patient handling equipment used in ambulances - Part 4: Foldable patient transfer chair*

EN 1865-5:2012, *Patient handling equipment used in ambulances - Part 5: Stretcher support*